

INVITATION FOR QUOTATION

HSCC/PUR/Low Value/AIIMS Rae Bareli/ Temp OPD Equipment /05/2018 Dated: 24.07.18

To,

All Bidders

Subject: Invitation for Quotations for supply of Low Value Items for setting up of Temporary OPD for AIIMS, Rae Bareli under PMSSY.

Dear Sirs,

1. HSCC (India) Ltd. for
and on behalf of Ministry of Health & Family Welfare, Government of India invites your most competitive quotation for the following goods:

S. No.	Name of Equipment/ Instruments	Qty.
1	AUTOMATED ESR ANALYSER <ul style="list-style-type: none">· Based on Westergren method· Reading method infrared sensor technology with inbuilt temperature correction upto 18 degree C to ensure accurate reading.· Tubes prefilled with sodium citrate.· In built quality control programme, in-built mixer and printer.· Ensure continuous supply of tubes and strips· Warranty & CMC as per Bid document The whole system should be CE/USFDA /BIS certified (bidder to attach the technical literature for the same)	1

2	<p>CENTRIFUGE (TABLE TOP)</p> <p>SPECIFICATIONS:</p> <ol style="list-style-type: none"> 1. Compact size of centrifuge. light weight and table top . 2. Have the brushless motor to provide the maintenance free drive. 3. With high max rcf of 21, 000 X g for shorter centrifugation period. 4. Have ECO shut-off function turns compressor off after eight hours idle automatically to reduce energy consumption. 5. The technologically advanced intuitive design provides active cooling to eliminate heat build-up in the rotor chamber 6. The refrigerated centrifuge temperature range: - 100 C to + 400 C. 7. Should have automatic timer & digital display. 8. Brake to protect delicate samples; fully adjustable end –of run-alarm. 9. Rotors: Aerosol tight and PTEF coated rotor (24 X 1.5/2.0ml tubes, 0.5ml with adaptors), 4 X 8 tube PCR strip rotor, Rotor for 15ml tube, 18- place rotor for spin columns and all rotor should be autoclavable. 10. System should include the user/technical/maintenance manual. 11. Suitable UPS for one hour backup. <p>The whole system should be CE/USFDA /BIS certified (bidder to attach the technical literature for the same)</p>	1
3	<p><u>ENT OPD ITEMS</u></p> <p>NASAL SPECULUM</p> <ul style="list-style-type: none"> • Stainless Steel Rust Free, Standard Size <p>OTOSCOPE Rechargeable otoscope (Speculam 2.5,3,4,5mmm,1.5v-2 batteries, bulb 2.5 V.0.45 A, X2 magnification,Metal battery handle)</p> <p>TUNNING FORK</p> <ul style="list-style-type: none"> • Tunning fork(256,512,1024 Hz) <p>HEAD LAMP</p> <p>Fiber-optic head light with 3.5mm cable, light adjustable, light weight, head band supplied with</p>	02

	<p>suitable xenon light source. 4 port turret.</p> <p>LARYNGEAL MIRROR MEDIUM & LARGE</p> <ul style="list-style-type: none"> • Stainless Steel Rust Free, Standard Size <p>NASOPHARYNEAL MIRROR</p> <ul style="list-style-type: none"> • Stainless Steel Rust Free, Standard Size <p>BULLS LAMP</p> <ul style="list-style-type: none"> • Stainless Steel Rust Free, Standard Size 	
04	<p>AUTOCLAVE (VERTICAL)</p> <ol style="list-style-type: none"> 1. Automatic adjustable working pressure system. 2. Double walled. 3. Inside boiler made of stainless steel & outside mild steel finished in cream enamel. 4. Radial locking system lid. 5. The panel is provided with on/off switch, pressure gauge, steam release valve & indicators to show the working of mains & pressure control system. 6. Electrically operated on 220V A/C with stainless steel basket. 7. Digital Temperature controller with inbuilt timer that will cut off the heaters automatically after lapse of pre set time at the pre set temperature. 8. Microprocessor based PID controller. 9. Height: 600mm. 10. Diameter inside: 450mm(24 x18) 11. Should be FDA or CE or BIS approved product <p>The whole system should be CE/USFDA /BIS certified (bidder to attach the technical literature for the same)</p>	02

2. Quotation,

2.1 The contract shall be for the full quantity as described above. The rate quoted should be valid till 31.08.18

2.2 Corrections, if any, shall be made by crossing out, initialling, dating and re writing.

2.3 The prices quoted by the bidder shall be fixed for the duration of the contract

and shall not be subject to adjustment on any account. Purchaser reserve the right to increase the quantity at the time of order upto 100% of the original quantity tendered.

2.4 The unit price/ rate of the item should be clearly indicated in the tender. Rates /Prices quoted Inclusive of all taxes, duties, forwarding, insurance & transportation upto destination (AIIMS, Rae Bareli).

3. Each bidder shall submit only one quotation.

4. Evaluation of Quotations,

The Purchaser will evaluate and compare the quotations determined to be substantially responsive i.e. which are properly signed along with Make & Model must be quoted along with literature duly marked with specification asked in quotation; and confirm to the terms and conditions, and specifications.

5. The Quotations would be evaluated item wise.

6. Award of contract:

The Purchaser will award the contract to the bidder whose quotation has been determined to be substantially responsive and who has offered the lowest rate for the item.

6.1 Notwithstanding the above, the Purchaser reserves the right to accept or reject any quotations and cancel the bidding process and reject all quotations at any time prior to the issue of Purchase order without assigning any reason.

7. Delivery period – Material should be available Ex stock and delivery to be done within four days from the date of placement of order at AIIMS Rae Bareli

8. Payment shall be made in Indian Rupees as follows:

Satisfactory Acceptance - 100% of total cost

9. All supplied items shall under warranty of **12** months from the date of successful acceptance of items.

10. You are requested to submit your offer at HSCC I Ltd. , E-6A, Sector -1 Noida as per the following schedule:

Due date of submission 26.07.18 at 1500 hrs.

Due date of opening 26.07.18 at 1530 hrs.

11. Information brochures/ Product catalogue, if any must be accompanied with the quotation clearly indicating the model quoted for.

12. Sealed quotation to be submitted/ delivered at the address mentioned below,

General Manager (Procurement)
HSCC (India) Ltd
E- 6 (A), Sector -1
NOIDA – 201 301

THE COVER SHOULD BE SUPERSCRIBED WITH THE FOLLOWING:

Reference to letter of enquiry.
Due date of submission 26.07.18 at 1500 hrs.
Due date of opening 26.07.18 at 1530 hrs.

13. We look forward to receiving your quotation and thank you for your interest in this project.

14. Quoted amount should be in Indian Rupees only and should be valid till 31.08.18

15. For all items the Technical Evaluation Committee, may opt for Demonstration of the items. The Committee may also ask for Demonstration / Inspection before supply / delivery of the items for quality assurance.

16. Insurance of 110% value upto the delivery of site Shall be arranged by supplier and cost of the same should be included in the price quoted.

Note: Please indicate the quotation reference no. (given at the top of page 1 of this letter) on the quotation envelope.

The details of various medical equipment are also made available at www.hsccltd.co.in, modification/amendments etc, if any, shall only be notified on website.

General Manager (Procurement)

Price bid

SI No.	Name of Item	Qty(Nos)	Unit Price (Rs.) (Incl. Of forwarding, packing, insurance, transportation up to destination)	Tax (if any)	Unit Price (Incl. Tax)	Total Price(Incl. Tax) Rs.

We accept all terms and conditions of the above Invitation of Quotation.

Signature

Authorized signatory *Name & Designation*

-