

DOMESTIC TENDER ENQUIRY DOCUMENT

**FOR PURCHASE OF
MEDICAL EQUIPMENT
FOR AIIMS LIKE INSTITUTION**

**FOR
MINISTRY OF HEALTH & FAMILY WELFARE,
GOVT. OF INDIA**



HSCC (INDIA) LTD
(A GOVERNMENT OF INDIA ENTERPRISE)

**Plot No. 6-A, Block-E,
Sector-1,
NOIDA (U.P.) – 201 301
PHONE: 0120-2540153
FAX: 0120-2542447
URL: www.hsccltd.com**

HSCC/PUR/AIIMS/RAEBARELI/TEMP OPD/2014

INDEX

Section	Topic	Page No.
Section I	-- Notice inviting Tender (NIT) -----	3
Section II	-- General Instructions to Tenderers (GIT) -----	08
Section III	-- Special Instructions to Tenderers (SIT) -----	24
Section IV	-- General Conditions of Contract (GCC) -----	26
Section V	-- Special Conditions of Contract (SCC) -----	39
Section VI	-- List of Requirements -----	40
Section VII	-- Technical Specifications -----	41
Section VIII	-- Quality Control Requirements -----	92
Section IX	-- Qualification Criteria -----	77
Section X	-- Tender Form -----	79
Section XI	-- Price Schedules -----	80
Section XII	-- Questionnaire -----	83
Section XIII	-- Bank Guarantee Form for EMD -----	84
Section XIV	-- Manufacturer's Authorisation Form -----	85
Section XV	-- Bank Guarantee Form for Performance Security /CMC Security -----	86
Section XVI	-- Contract Form (A & B) -----	87
Section XVII	-- Proforma of Consignee Receipt Certificate -----	91
Section XVIII	-- Proforma of Final Acceptance Certificate by the Consignee -----	92
Section XIX	-- Instructions from Ministry of Shipping/Surface Transport (Annexure 1) ----	94
Section XX	-- Check List for the Tenderers -----	95
Section XXI	-- Consignee-----	98

SECTION I
NOTICE INVITING TENDERS (NIT)

For Domestic Tender from
Ministry of Health & Family Welfare
Through

HSCC (INDIA) LTD

(A GOVERNMENT OF INDIA ENTERPRISE)

Plot No. 6-A, Block-E, Sector-1, NOIDA (U.P.) – 201 301

PHONE: 0120-2540153

FAX: 0120-2542447

URL: www.hsccltd.co.in

Tender Enquiry No.: HSCC/PUR/AIIMS/RAEBARELI/TEMP OPD/2014 Dated 24.01.2014

NOTICE INVITING TENDERS (NIT)

Bids are invited on behalf of Ministry of Health & Family Welfare through HSCC (India) Ltd from eligible and qualified tenderers for supply of following Medical Equipments for Temporary OPD at AIIMS – RAEBARELI:-

Tender No.: HSCC/PUR/AIIMS/RAEBARELI/TEMP OPD/2014

S. No.	DESCRIPTION	Qty.	EMD in Rs.
	Radiology (Imaging)		
1	Whole body Color Ultrasound	1	50,000
2	300-mA X-ray Machine, Auto Film Processor, Accessories (Cassettes, Aprons Etc.)	1	83,000
	LAB		EMD in Rs.
3	Haematology analyzer automatic 3-part differential	1	12,000
4	Biochemistry Auto Analyzer	1	50,000
5	Autoclave vertical type	1	3,000
	Gyn. OPD - 2 Nos.	Qty.	EMD in Rs.
6	Spot light for gynae exam.	2	6,000
	Dental OPD - 02 No.	Qty.	EMD in Rs.
7	Progmmable Dental Chair Unit Complete Set with Accessories, Dental X-Ray, Compressor, Hand Piece Cleaning System, Piezo-electric Scalar, Dental Front Loading Autoclave, Intra-oral Camera with Monitor etc.)	2	48,000

HSCC (India) Limited

	Minor OT (1no.)	Qty.	EMD in Rs.
8	Aneasthesia Machine	1	30,000
9	Diathermy	1	10,000
10	OT Table Hydraulic	1	16,000
11	Mobile Light	1	6,000
12	Instruments for Suture Set (4nos.), Stich Removal Set (4nos.), Minor Surgery Set (4nos.)	1	12,000
	Ophthalmology - 2 No.	Qty.	EMD in Rs.
13	Slit Lamp	2	16,000
14	Lensometer	2	8,000

Sl. No.	Description	Schedule
i.	Dates of sale of tender enquiry documents	24.01.2014 to 07.02.2014, 12.00 hrs IST
ii.	Place of sale of Tender Enquiry Documents	HSCC (India) Ltd, Plot No. 6-A, Block-E, Sector-1, Noida (U.P)-201301
iii.	Cost of the Tender Enquiry Document	INR 2, 000/-
iv.	Pre Tender Meeting Date & Time	on 30.01.2014, 11.00 hrs IST
v.	Pre Tender Meeting Venue	Same as 2 (ii)
vi.	Closing date & time for receipt of Tender	on 07.02.2014, 14.30 hrs IST
vii.	Time and date of opening of Techno – Commercial tenders	on 07.02.2014, 15.00 hrs IST
Viii	Venue of Opening of Techno Commercial Tender	Same as 2 (ii)

3. Interested tenderers may obtain further information about this requirement from the above office selling the documents. Tender Enquiry Documents may be purchased on payment of non-refundable fee of INR 2,000/- per set in the form of account payee Demand Draft/Pay Order/Cashier's Cheque/Banker's Cheque, drawn on a scheduled Bank in India, in favour of "**HSCC (India) Ltd**" payable at New Delhi/ Noida.
4. Tenderer may also download the tender enquiry documents from the web site <http://eprocure.gov.in/cppp> or www.hscltd.com submit its tender by utilizing the downloaded document, along with the required non-refundable fee as mentioned in Para 3 above.
5. All prospective tenderers may attend the Pre Tender meeting. The venue, date and time indicated in the Para 2 above.
6. Tenderers shall ensure that their tenders, complete in all respects, are dropped in the Tender Box located at **HSCC (India) Ltd, Plot No. 6-A, Block-E, Sector-1, NOIDA (U.P.) – 201 301** on or before the closing date and time indicated in the Para 2 above, failing which the tenders will be treated as late and rejected.
8. The Tender Enquiry Documents are not transferable.

Dy. Secretary (PMSSY), MOHF&W

SECTION - II**GENERAL INSTRUCTIONS TO TENDERERS (GIT)
CONTENTS**

Sl. No.	Topic	Page No.
A	PREAMBLE	
1	Definitions and Abbreviations	8
2	Introduction	9
3	Availability of Funds	10
4	Language of Tender	10
5	Eligible Tenderers	10
6	Eligible Goods and Services	10
7	Tendering Expense	10
B	TENDER ENQUIRY DOCUMENTS	
8	Contents of Tender Enquiry Documents	10
9	Amendments to Tender Enquiry Documents	11
10	Clarification of Tender Enquiry Documents	11
C	PREPARATION OF TENDERS	
11	Documents Comprising the Tender	11
12	Tender Currencies	12
13	Tender Prices	12
14	Indian Agent	14
15	Firm Price / Variable Price	14
16	Alternative Tenders	14
17	Documents Establishing Tenderer's Eligibility and Qualifications	15
18	Documents Establishing Good's Conformity to Tender Enquiry Document	15
19	Earnest Money Deposit (EMD)	15
20	Tender Validity	16
21	Signing and Sealing of Tender	16
D		
22	Submission of Tenders	17
23	Late Tender	17
24	Alteration and Withdrawal of Tender	17
E	TENDER OPENING	

25	Opening of Tenders	17
F	SCRUTINY AND EVALUATION OF TENDERS	
26	Basic Principle	18
27	Preliminary Scrutiny of Tenders	18
28	Minor Infirmary/Irregularity/Non-Conformity	18
29	Discrepancy in Prices	19
30	Discrepancy between original and copies of Tender	19
31	Qualification Criteria	19
32	Conversion of Tender Currencies to Indian Rupees	19
33	Schedule-wise Evaluation	19
34	Comparison of Tenders	19
35	Additional Factors and Parameters for Evaluation and Ranking of Responsive Tenders	20
36	Tenderer's capability to perform the contract	20
37	Contacting the Purchaser	20
G	AWARD OF CONTRACT	
38	Purchaser's Right to Accept any Tender and to Reject any or All Tenders	21
39	Award Criteria	21
40	Variation of Quantities at the Time of Award	21
41	Notification of Award	21
42	Issue of Contract	21
43	Non-receipt of Performance Security and Contract by the Purchaser/Consignee	21
44	Return of EMD	22
45	Publication of Tender Result	22
46	Corrupt or Fraudulent Practices	22

GENERAL INSTRUCTIONS TO TENDERERS (GIT)

A. PREAMBLE

1. Definitions and Abbreviations

1.1 The following definitions and abbreviations which have been used in these documents shall have the meanings as indicated below:

1.2. Definitions:

- (i) "Purchaser" means MoH&FW or their nominee/assignee
- (ii) "Tender" means Bids / Quotation / Tender received from a Firm / Tenderer / Bidder.
- (iii) "Tenderer" means Bidder/ the Individual or Firm submitting Bids / Quotation / Tender
- (iv) "Supplier" means the individual or the firm supplying the goods and services as incorporated in the contract.
- (v) "Goods" means the articles, material, commodities, livestock, furniture, fixtures, raw material, spares, instruments, machinery, equipment, medical equipment, industrial plant etc. which the supplier is required to supply to the purchaser under the contract.
- (vi) "Services" means services allied and incidental to the supply of goods, such as transportation, installation, commissioning, provision of technical assistance, training, after sales service, maintenance service and other such obligations of the supplier covered under the contract.
- (vii) "Earnest Money Deposit" (EMD) means Bid Security/ monetary or financial guarantee to be furnished by a tenderer along with its tender.
- (viii) "Contract" means the written agreement entered into between the purchaser and/or consignee and the supplier, together with all the documents mentioned therein and including all attachments, annexure etc. therein.
- (ix) "Performance Security" means monetary or financial guarantee to be furnished by the successful tenderer for due performance of the contract placed on it. Performance Security is also known as Security Deposit.
- (x) "Consignee" means the Hospital (AIIMS)/Institute/Medical College/ Person to whom the goods are required to be delivered as specified in the Contract. If the goods are required to be delivered to a person as an interim consignee for the purpose of despatch to another person as provided in the Contract then that "another" person is the consignee, also known as ultimate consignee.
- (xi) "Specification" means the document/standard that prescribes the requirement with which goods or service has to conform.
- (xii) "Inspection" means activities such as measuring, examining, testing, gauging one or more characteristics of the product or service and comparing the same with the specified requirement to determine conformity.
- (xiii) "Day" means calendar day.

1.3 Abbreviations:

- (i) "TE Document" means Tender Enquiry Document
- (ii) "NIT" means Notice Inviting Tenders.
- (iii) "GIT" means General Instructions to Tenderers
- (iv) "SIT" means Special Instructions to Tenderers
- (v) "GCC" means General Conditions of Contract

- (vi) "SCC" means Special Conditions of Contract
- (vii) "DGS&D" means Directorate General of Supplies and Disposals
- (viii) "NSIC" means National Small Industries Corporation
- (ix) "PSU" means Public Sector Undertaking
- (x) "CPSU" means Central Public Sector Undertaking
- (xi) "LSI" means Large Scale Industry
- (xii) "SSI" means Small Scale Industry
- (xiii) "LC" means Letter of Credit
- (xiv) "DP" means Delivery Period
- (xv) "BG" means Bank Guarantee
- (xvi) "ED" means Excise Duty
- (xvii) "CD" means Custom Duty
- (xviii) "VAT" means Value Added Tax
- (xix) "CENVAT" means Central Value Added Tax
- (xx) "CST" means Central Sales Tax
- (xxi) "RR" means Railway Receipt
- (xxii) "BL" means Bill of Lading
- (xxiii) "FOB" means Free on Board
- (xxiv) "FCA" means Free Carrier
- (xxv) "FOR" means Free On Rail
- (xxvi) "CIF" means Cost, Insurance and Freight
- (xxvii) "CIP (Destinations)" means Carriage and Insurance Paid up to Consignee Site. Additionally the Insurance (local transportation and storage) would be extended and borne by the Supplier from ware house to the consignee site for a period including 3 months beyond date of delivery.
- (xxviii) "DDP" means Delivery Duty Paid named place of destination (consignee site)
- (xxix) "INCOTERMS" means International Commercial Terms as on the date of Tender Opening
- (xxx) "MOH&FW" means Ministry of Health & Family Welfare, Government of India
- (xxxi) "Dte. GHS" means Directorate General and Health Services, MOH&FW.
- (xxxii) "CMC" means Comprehensive maintenance Contract (labour, spare and preventive maintenance)
- (xxxiii) "RT" means Re-Tender.

2. Introduction

- 2.1 The Purchaser has issued these TE documents for purchase of goods and related services as mentioned in Section – VI – "List of Requirements", which also indicates, *inter alia*, the required delivery schedule, terms and place of delivery.
- 2.2 This section (Section II - "General Instruction Tenderers") provides the relevant information as well as instructions to assist the prospective tenderers in preparation and submission of tenders. It also includes the mode and procedure to be adopted by the purchaser for receipt and opening as well as scrutiny and evaluation of tenders and subsequent placement of contract.
- 2.3 The tenderers shall also read the Special Instructions to Tenderers (SIT) related to this purchase, as contained in Section III of these documents and follow the same accordingly. Whenever there is a conflict between the GIT and the SIT, the provisions contained in the SIT shall prevail over those in the GIT.
- 2.4 Before formulating the tender and submitting the same to the purchaser, the tenderer should read and examine all the terms, conditions, instructions, checklist etc. contained in the TE documents. Failure to provide and/or comply with the required information, instructions etc. incorporated in these TE documents may result in rejection of its tender.

3. Availability of Funds

- 3.1 Expenditure to be incurred for the proposed purchase will be met from the funds available with the purchaser/consignee.

4. Language of Tender

- 4.1 The tender submitted by the tenderer and all subsequent correspondence and documents relating to the tender exchanged between the tenderer and the purchaser, shall be written in the English language only.

5. Eligible Tenderers

- 5.1 This invitation for tenders is open to all suppliers who fulfil the eligibility criteria specified in these documents.

6. Eligible Goods and Services

- 6.1 **This being domestic tender goods should be manufactured or sourced at India.**

7. Tendering Expense

- 7.1 The tenderer shall bear all costs and expenditure incurred and/or to be incurred by it in connection with its tender including preparation, mailing and submission of its tender and for subsequent processing the same. The purchaser will, in no case be responsible or liable for any such cost, expenditure etc regardless of the conduct or outcome of the tendering process.

B. TENDER ENQUIRY DOCUMENTS

8. Content of Tender Enquiry Documents

- 8.1 In addition to Section I – “Notice inviting Tender” (NIT), the TE documents include:

- Section II – General Instructions to Tenderers (GIT)
- Section III – Special Instructions to Tenderers (SIT)
- Section IV – General Conditions of Contract (GCC)
- Section V – Special Conditions of Contract (SCC)
- Section VI – List of Requirements
- Section VII – Technical Specifications
- Section VIII – Quality Control Requirements
- Section IX – Qualification Criteria
- Section X – Tender Form
- Section XI – Price Schedules
- Section XII – Questionnaire
- Section XIII – Bank Guarantee Form for EMD
- Section XIV – Manufacturer’s Authorisation Form
- Section XV – Bank Guarantee Form for Performance Security/CMC Security
- Section XVI – Contract Forms A & B
- Section XVII – Proforma of Consignee Receipt Certificate
- Section XVIII – Proforma of Final Acceptance Certificate by the consignee
- Section XIX – Instructions from Ministry of Shipping/ Surface Transport (Annexure 1 & 2)
- Section XX – Check List for the Tenderers
- Section XXI – Consignee List

- 8.2 The relevant details of the required goods and services, the terms, conditions and procedure for tendering, tender evaluation, placement of contract, the applicable contract terms and, also, the standard formats to be used for this purpose are incorporated in the above-

mentioned documents. The interested tenderers are expected to examine all such details etc to proceed further.

9. Amendments to TE documents

9.1 At any time prior to the deadline for submission of tenders, the purchaser may, for any reason deemed fit by it, modify the TE documents by issuing suitable amendment(s) to it.

9.2 Such an amendment will be notified and published in HSCC website and no separate correspondence will be made for this.

10. Clarification of TE documents

10.1 A tenderer requiring any clarification or elucidation on any issue of the TE documents may take up the same with the purchaser in writing during Pre- Bid meeting. The purchaser may respond to such if required provided the same is received by the purchaser.

C. PREPARATION OF TENDERS

11. Documents Comprising the Tender

11.1 The **Two Tender System**, i.e. “Techno – Commercial Tender” and “Price Tender” prepared by the tenderer shall comprise the following:

A) Techno – Commercial Tender (Un priced Tender)

- i) Earnest money furnished in accordance with GIT clause 19.1 alternatively, documentary evidence as per GIT clause 19.2 for claiming exemption from payment of earnest money.
- ii) Tender Form as per Section X (without indicating any prices).
- iii) Documentary evidence, as necessary in terms of clauses 5 and 17 establishing that the tenderer is eligible to submit the tender and, also, qualified to perform the contract if its tender is accepted.
- iv) Tenderer/Agent who quotes for goods manufactured by other manufacturer shall furnish Manufacturer’s Authorisation Form. **While giving authorization to agent , to quote on their behalf, manufacturer has to give the reasons for not quoting directly against this tender.**
- v) Power of Attorney in favour of signatory of TE documents and signatory of Manufacturer’s Authorisation Form.
- vi) Documents and relevant details to establish in accordance with GIT clause 18 that the goods and the allied services to be supplied by the tenderer conform to the requirement of the TE documents.
- vii) Performance Statement as per section IX along with relevant copies of orders and end users’ satisfaction certificate.
- viii) Price Schedule(s) as per Section XI filled up with all the details including Make, Model etc. of the goods offered with prices blank (without indicating any prices).
- ix) Certificate of Incorporation in the country of origin.
- x) Checklist as per Section XX.

B) Price Tender:

The information given at clause no. 11.1 A) ii) & viii) above should be reproduced with the prices indicated.

NOTE:

1. All pages of the Tender should be page numbered and indexed.
2. It is the responsibility of tenderer to go through the TE document to ensure furnishing all required documents in addition to above, if any.

- 11.2 The authorized signatory of the tenderer must sign the tender duly stamped at appropriate places and initial all the remaining pages of the tender. Individuals signing the tender or other documents connected with a contract must specify whether he signs as:
- i. A 'Sole Proprietor' of the firm or constituted attorney of such Sole Proprietor.
 - ii. A partner of the firm, if it be a partnership, in which case he must have authority to quote & to refer to arbitration dispute concerning the business of the partnership either by virtue of the partnership agreement or a power of attorney;
 - iii. Constituted attorney of the firm if it is a company.

NOTE:

1. In case of (ii) above, a copy of the partnership agreement or general power of attorney, in either, case, attested by a Notary Public should be furnished, or affidavit on stamped paper of all the partners admitting execution of the partnership agreement or the general power of attorney should be furnished.
 2. In case of the Partnership firms, where no authority to refer disputes concerning the business of the partnership has been conferred on any partner, the tender and all other related documents must be signed by every partner of the firm.
 3. A person signing the tender form or any documents forming part of the contract on behalf of another shall be deemed to warrantee that he has authority to bind such other persons and if, on enquiry, it appears that the persons so signing had no authority to do so, the purchaser may, without prejudice to other civil and criminal remedies, cancel the contract and hold the signatory liable for all cost and damages.
- 11.3 A tender, which does not fulfil any of the above requirements and/or gives evasive information/reply against any such requirement, shall be liable to be ignored and rejected.
- 11.4 Tender sent by fax/telex/cable/electronically shall be ignored.

12. Tender currencies

- 12.1 The tenderer supplying indigenous goods or already imported goods shall quote only in Indian Rupees.
- 12.2 Deleted
- 12.3 Tenders, where prices are quoted in any other way shall be treated as non -responsive and rejected.

13 Tender Prices

- 13.1 The Tenderer shall indicate on the Price Schedule provided under Section XI all the specified components of prices shown therein including the unit prices and total tender prices of the goods and services it proposes to supply against the requirement. All the columns shown in the price schedule should be filled up as required. If any column does not apply to a tenderer, same should be clarified as "NA" by the tenderer.
- 13.2 If there is more than one schedule in the List of Requirements, the tenderer has the option to submit its quotation for any one or more schedules and, also, to offer special discount for combined schedules. However, while quoting for a schedule, the tenderer shall quote for the complete requirement of goods and services as specified in that particular schedule.
- 13.3 Deleted
- 13.4 While filling up the columns of the Price Schedule, the following aspects should be noted for compliance:
- 13.4.1 For domestic goods or goods of foreign origin located within India, the prices in the corresponding price schedule shall be entered separately in the following manner:
- a) the price of the goods, quoted ex-factory/ ex-showroom/ ex-warehouse/ off-the-shelf, as applicable, including all taxes and duties like sales tax, CST VAT, CENVAT, Custom

- Duty, Excise Duty etc. already paid or payable on the components and raw material used in the manufacture or assembly of the goods quoted ex-factory etc. or on the previously imported goods of foreign origin quoted ex-showroom etc;
- b) any sales or other taxes and any duties including excise duty, which will be payable on the goods in India if the contract is awarded;
 - c) charges towards Packing & Forwarding, Inland Transportation, Insurance (local transportation and storage) would be borne by the Supplier from ware house to the consignee site for a period including 3 months beyond date of delivery, Loading/Unloading and other local costs incidental to delivery of the goods to their final destination as specified in the List of Requirements and Price Schedule;
 - d) the price of Incidental Services, as mentioned in List of Requirements and Price Schedule;
 - e) the prices of Turnkey (if any), as mentioned in List of Requirements, Technical Specification and Price Schedule; and
 - f) the price of annual CMC, as mentioned in List of Requirements, Technical Specification and Price Schedule.

13.4.2 Deleted.

13.5 Additional information and instruction on Duties and Taxes:

13.5.1 If the Tenderer desires to ask for excise duty, sales tax/ VAT, Service Tax, Works Contract Tax etc. to be paid extra, the same must be specifically stated. In the absence of any such stipulation the price will be taken inclusive of such duties and taxes and no claim for the same will be entertained later.

13.5.2 Excise Duty:

- a) If reimbursement of excise duty is intended as extra over the quoted prices, the supplier must specifically say so also indicating the rate, quantum and nature of the duty applicable. In the absence of any such stipulation it will be presumed that the prices quoted are firm and final and no claim on account of excise duty will be entertained after the opening of tenders.
- b) If a Tenderer chooses to quote a price inclusive of excise duty and also desires to be reimbursed for variation, if any, in the excise duty during the time of supply, the tenderer must clearly mention the same and also indicate the rate and quantum of excise duty included in its price. Failure to indicate all such details in clear terms may result in rejection of that tender.
- c) Subject to sub clauses 13.5.2 (a) & (b) above, any change in excise duty upward/downward as a result of any statutory variation in excise duty taking place within contract terms shall be allowed to the extent of actual quantum of excise duty paid by the supplier. In case of downward revision in excise duty, the actual quantum of reduction of excise duty shall be reimbursed to the purchaser by the supplier. All such adjustments shall include all reliefs, exemptions, rebates, concession etc. if any obtained by the supplier.

13.5.3 Sales Tax:

If a tenderer asks for sales tax/ VAT, Service Tax and Works Contract Tax to be paid extra, the rate and nature of sales tax applicable should be shown separately. The sales tax / VAT, Service Tax and Works Contract Tax will be paid as per the rate at which it is liable to be assessed or has actually been assessed provided the transaction of sale is legally liable to sales tax / VAT, Service Tax and Works Contract Tax and is payable as per the terms of the contract. If any refund of Tax is received at a later date, the Supplier must return the amount forth-with to the purchaser.

13.5.4 Octroi Duty and Local Duties & Taxes:

Normally, goods to be supplied to government departments against government contracts are exempted from levy of town duty, Octroi duty, terminal tax and other levies of local bodies. However, on some occasions, the local bodies (like town body, municipal body etc.) as per their regulations allow such exemptions only on production of certificate to this effect from the concerned government department. Keeping this in view, the supplier shall ensure that the stores to be supplied by the supplier against the contract placed by the purchaser are exempted from levy of any such duty or tax and, wherever necessary, obtain the exemption certificate from the purchaser.

However, if a local body still insists upon payment of such local duties and taxes, the same should be paid by the supplier to the local body to avoid delay in supplies and possible demurrage charges and obtain a receipt for the same. The supplier should forward the receipt obtained for such payment to the purchaser to enable the purchaser reimburse the supplier and take other necessary action in the matter.

13.5.5 Deleted

13.6 Deleted

13.7 For insurance of goods to be supplied, relevant instructions as provided under GCC Clause 11 shall be followed.

13.8 Deleted

13.9 The need for indication of all such price components by the tenderers, as required in this clause (viz., GIT clause 13) is for the purpose of comparison of the tenders by the purchaser and will no way restrict the purchaser's right to award the contract on the selected tenderer on any of the terms offered.

14. Indian Agent

14.1 Deleted

15. Firm Price

15.1 Unless otherwise specified in the SIT, prices quoted by the tenderer shall remain firm and fixed during the currency of the contract and not subject to variation on any account.

15.2 However, as regards taxes and duties, if any, chargeable on the goods and payable, the conditions stipulated in GIT clause 13 will apply.

16. Alternative Tenders

16.1 Alternative Tenders are not permitted.

16.2 However the Tenderers can quote alternate models meeting the tender specifications of same manufacturer with single EMD.

16.3 a). If a tenderer, either the Indian Agent on behalf of the Principal / OEM or Principal / OEM itself can bid but both cannot bid simultaneously for the same item/ product in the same tender

b). If an agent submits bid on behalf of the Principal / OEM, the same agent shall not submit a bid on behalf of another Principal / OEM in the same tender for the same item / product.

17 Documents Establishing Tenderer's Eligibility and Qualifications

- 17.1 Pursuant to GIT clause 11, the tenderer shall furnish, as part of its tender, relevant details and documents establishing its eligibility to quote and its qualifications to perform the contract if its tender is accepted.
- 17.2 The documentary evidence needed to establish the tenderer's qualifications shall fulfil the following requirements:
- a) in case the tenderer offers to supply goods, which are manufactured by some other firm, the tenderer has been duly authorised by the goods manufacturer to quote for and supply the goods to the purchaser. The tenderer shall submit the manufacturer's authorization letter to this effect as per the standard form provided under Section XIV in this document.
 - b) the tenderer has the required financial, technical and production capability necessary to perform the contract and, further, it meets the qualification criteria incorporated in the Section IX in these documents.
 - c) in case the tenderer is not doing business in India, it is duly represented by an agent stationed in India fully equipped and able to carry out the required contractual functions and duties of the supplier including after sale service, maintenance & repair etc. of the goods in question, stocking of spare parts and fast moving components and other obligations, if any, specified in the conditions of contract and/or technical specifications.

18. Documents establishing Good's Conformity to TE document.

- 18.1 The tenderer shall provide in its tender the required as well as the relevant documents like technical data, literature, drawings etc. to establish that the goods and services offered in the tender fully conform to the goods and services specified by the purchaser in the TE documents. For this purpose the tenderer shall also provide a clause-by-clause commentary on the technical specifications and other technical details incorporated by the purchaser in the TE documents to establish technical responsiveness of the goods and services offered in its tender.
- 18.2 In case there is any variation and/or deviation between the goods & services prescribed by the purchaser and that offered by the tenderer, the tenderer shall list out the same in a chart form without ambiguity and provide the same along with its tender.
- 18.3 If a tenderer furnishes wrong and/or misleading data, statement(s) etc. about technical acceptability of the goods and services offered by it, its tender will be liable to be ignored and rejected in addition to other remedies available to the purchaser in this regard.

19. Earnest Money Deposit (EMD)

- 19.1 The earnest money shall be denominated in Indian Rupees or equivalent currencies as per GIT clause 12.2. The earnest money shall be furnished in one of the following forms:
- i) Account Payee Demand Draft
 - ii) Banker's cheque and
 - iii) Bank Guarantee
- 19.2 The demand draft or banker's cheque shall be drawn on any commercial bank in India or country of the tenderer, in favour of the "**HSCC (India) Ltd**" payable at New Delhi/Noida. In case of bank guarantee, the same is to be provided from any commercial bank in India or country of the tenderer as per the format specified under Section XIII in these documents
- 19.3 The earnest money shall be valid for a period of forty-five (45) days beyond the validity period of the tender. As validity period of Tender as per Clause 20 of GIT is 120 days, the EMD shall be valid for 165 days from Techno – Commercial Tender opening date.
- 19.4 Unsuccessful tenderers' earnest money will be returned to them without any interest, after expiry of the tender validity period, but not later than thirty days after conclusion of the

resultant contract. Successful tenderer's earnest money will be returned without any interest, after receipt of performance security from that tenderer.

- 19.5 Earnest Money is required to protect the purchaser against the risk of the Tenderer's conduct, which would warrant the forfeiture of the EMD. Earnest money of a tenderer will be forfeited, if the tenderer withdraws or amends its tender or impairs or derogates from the tender in any respect within the period of validity of its tender or if it comes to notice that the information/documents furnished in its tender is incorrect, false, misleading or forged without prejudice to other rights of the purchaser. The successful tenderer's earnest money will be forfeited without prejudice to other rights of Purchaser if it fails to furnish the required performance security within the specified period.

- 19.6 **Deleted.**

20. Tender Validity

- 20.1 If not mentioned otherwise in the SIT, the tenders shall remain valid for acceptance for a period of 120 days (One hundred and twenty days) after the date of tender opening prescribed in the TE document. Any tender valid for a shorter period shall be treated as unresponsive and rejected.
- 20.2 In exceptional cases, the tenderers may be requested by the purchaser to extend the validity of their tenders up to a specified period. Such request(s) and responses thereto shall be conveyed by surface mail or by fax/ telex/cable followed by surface mail. The tenderers, who agree to extend the tender validity, are to extend the same without any change or modification of their original tender and they are also to extend the validity period of the EMD accordingly. A tenderer, however, may not agree to extend its tender validity without forfeiting its EMD.
- 20.3 In case the day up to which the tenders are to remain valid falls on/ subsequently declared a holiday or closed day for the purchaser, the tender validity shall automatically be extended up to the next working day.

21. Signing and Sealing of Tender

- 21.1 The tenderers shall submit their tenders as per the instructions contained in GIT Clause 11.
- 21.2 Unless otherwise mentioned in the SIT, a tenderer shall submit Two copies of its tender marking them as "Original" and "Duplicate". Duplicate tenders may contain all pages including Technical Literature/Catalogues as per in Original tenders.
- 21.3 The original and other copies of the tender shall either be typed or written in indelible ink and the same shall be signed by the tenderer or by a person(s) who has been duly authorized to bind the tenderer to the contract. The letter of authorization shall be by a written power of attorney, which shall also be furnished along with the tender.
- 21.4 Both the copies of the tender shall be duly signed at the appropriate places as indicated in the TE documents and all other pages of the tender including printed literature, if any shall be initialled by the same person(s) signing the tender. The tender shall not contain any erasure or overwriting, except as necessary to correct any error made by the tenderer and, if there is any such correction; the same shall be initialled by the person(s) signing the tender.
- 21.5 The tenderer is to seal the original and each copy of the tender in separate envelopes, duly marking the same as "Original", "Duplicate", "Triplicate" and so on and writing the address of the purchaser and the tender reference number on the envelopes. The sentence "NOT TO BE OPENED" before _____ (The tenderer is to put the date & time of tender opening) are to be written on these envelopes. The inner envelopes are then to be put in a bigger outer envelope, which will also be duly sealed, marked etc. as above. If the outer envelope is not sealed and marked properly as above, the purchaser will not assume any responsibility for its misplacement, premature opening, late opening etc.

- 21.6 TE document seeks quotation following **two Tender System**, in two parts. First part will be known as **'Techno - Commercial Tender'**, and the second part **'Price Tender'** as specified in clause 11 of GIT. Tenderer shall seal **'Techno - Commercial Tender'** and **'Price Tender'** separately and covers will be suitably super scribed. Both these sealed covers shall be put in a bigger cover and sealed and procedure prescribed in Paras 21.1 to 21.5 followed.

D. SUBMISSION OF TENDERS

22. Submission of Tenders

- 22.1 Unless otherwise specified, the tenderers are to deposit the tenders in the tender box kept for this purpose at **HSCC (India) Ltd, Plot E-6 (A), Sector – 1, Noida**.
- 22.2 The tenderers must ensure that they deposit their tenders not later than the closing time and date specified for submission of tenders. It is the responsibility of the tenderer to ensure that their Tenders whether sent by post or by courier or by person, are dropped in the Tender Box by the specified clearing date and time. In the event of the specified date for submission of tender falls on / is subsequently declared a holiday or closed day for the purchaser, the tenders will be received up to the appointed time on the next working day.

23. Late Tender

- 23.1 A tender, which is received after the specified date and time for receipt of tenders will be treated as "late" tender and will be ignored.

24. Alteration and Withdrawal of Tender

- 24.1 The tenderer, after submitting its tender, is permitted to alter / modify its tender so long as such alterations / modifications are received duly signed, sealed and marked like the original tender, within the deadline for submission of tenders. Alterations / modifications to tenders received after the prescribed deadline will not be considered.
- 24.2 No tender should be withdrawn after the deadline for submission of tender and before expiry of the tender validity period. If a tenderer withdraws the tender during this period, it will result in forfeiture of the earnest money furnished by the tenderer in its tender.

E. TENDER OPENING

25. Opening of Tenders

- 25.1 The purchaser will open the tenders at the specified date and time and at the specified place as indicated in the NIT.
- 25.2 Authorized representatives of the tenderers, who have submitted tenders on time may attend the tender opening provided they bring with them letters of authority from the corresponding tenderers.
- 25.3 Two - Tender system as mentioned in Para 21.6 above will be as follows. The **Techno - Commercial Tenders** are to be opened in the first instance, at the prescribed time and date as indicated in NIT. These Tenders shall be scrutinized and evaluated by the competent committee/ authority with reference to parameters prescribed in the TE document. During the Techno - Commercial Tender opening, the tender opening official(s) will read the salient features of the tenders like brief description of the goods offered, delivery period, Earnest Money Deposit and any other special features of the tenders, as deemed fit by the tender opening official(s). Thereafter, in the second stage, the Price Tenders of only the Techno - Commercially acceptable offers (as decided in the first stage) shall be opened for further scrutiny and evaluation on a date notified after the evaluation of the Techno – Commercial

tender. The prices, special discount if any of the goods offered etc., as deemed fit by tender opening official(s) will be read out.

F. SCRUTINY AND EVALUATION OF TENDERS

26. Basic Principle

26.1 Tenders will be evaluated on the basis of the terms & conditions already incorporated in the TE document, based on which tenders have been received and the terms, conditions etc. mentioned by the tenderers in their tenders. No new condition will be brought in while scrutinizing and evaluating the tenders.

27. Scrutiny of Tenders

27.1 The Purchaser will examine the Tenders to determine whether they are complete, whether any computational errors have been made, whether required sureties have been furnished, whether the documents have been properly signed stamped and whether the Tenders are generally in order.

27.2 The Purchaser's determination of a tender's responsiveness is to be based on the contents of the tender itself without recourse to extrinsic evidence.

27.3 The tenders will be scrutinized to determine whether they are complete and meet the essential and important requirements, conditions etc. As prescribed in the TE document. The tenders, which do not meet the basic requirements, are liable to be treated as non – responsive and will be rejected.

27.4 The following are some of the important aspects, for which a tender shall be declared non – responsive and will be summarily ignored;

(i) Tender form as per Section X (Signed and stamped) not enclosed.

(ii) Tender is unsigned.

(iii) Tender validity is shorter than the required period.

(iv) Required EMD (Amount, validity etc.)/ exemption documents have not been provided.

(v) Tenderer has quoted for goods manufactured by other manufacturer(s) without the required Manufacturer's Authorisation Form as per Section XIV.

(vi) Tenderer has not agreed to give the required performance security of required amount in an acceptable form in terms of GCC clause 5, read with modification, if any, in Section – V – “Special Conditions of Contract”, for due performance of the contract.

(vii) Deleted

(viii) Tenderer has not agreed to other essential condition(s) specially incorporated in the tender enquiry like terms of payment, liquidated damages clause, warranty clause, dispute resolution mechanism applicable law.

(ix) Poor/ unsatisfactory past performance.

(x) Tenderers who stand deregistered/banned/blacklisted by any Govt. Authorities.

(xi) Tenderer is not eligible as per GIT Clauses 5.1 & 17.1.

(xii) Tenderer has not quoted for the entire quantity as specified in the List of Requirements in the quoted schedule.

(Xiii) Tenderer has not agreed for the delivery terms & delivery schedule.

28. Minor Infirmary/Irregularity/Non-Conformity

28.1 If during the evaluation, the purchaser find any any minor informality and/or irregularity and/or non-conformity in a tender, the purchaser will convey its observation on such 'minor' issues to the tenderer by registered/speed post etc. asking the tenderer to respond by a specified date. If the tenderer does not reply by the specified date or gives evasive reply without clarifying the point at issue in clear terms, that tender will be liable to be ignored.

29 Discrepancies in Prices

- 29.1 If, in the price structure quoted by a tenderer, there is discrepancy between the unit price and the total price (which is obtained by multiplying the unit price by the quantity), the unit price shall prevail and the total price corrected accordingly, unless the purchaser feels that the tenderer has made a mistake in placing the decimal point in the unit price, in which case the total price as quoted shall prevail over the unit price and the unit price corrected accordingly.
- 29.2 If there is an error in a total price, which has been worked out through addition and/or subtraction of subtotals, the subtotals shall prevail and the total corrected; and
- 29.3 If there is a discrepancy between the amount expressed in words and figures, the amount in words shall prevail, subject to sub clause 29.1 and 29.2 above.
- 29.4 When the rate quoted by the tenderer in figures and in words tallies but the amount is not worked out correctly, the rate quoted by the tenderer shall be taken as correct and not the amount.
If the bidder does not accept the corrected amount of bid, his bid will be rejected, and the bid security may be forfeited.

30. Discrepancy between original and copies of Tender

- 30.1 In case any discrepancy is observed between original & duplicate the tender will be liable to be rejected.

31. Qualification Criteria

- 31.1 Tenders of the tenderers, who do not meet the required Qualification Criteria prescribed in Section IX, will be treated as non - responsive and will not be considered further.

32. Conversion of tender currencies to Indian Rupees

- 32.1 **Deleted**

33. Schedule-wise Evaluation

- 33.1 In case the List of Requirements contains more than one schedule, the responsive tenders will be evaluated and compared separately for each schedule. The tender for a schedule will not be considered if the complete requirements prescribed in that schedule are not included in the tender. However, as already mentioned in GIT sub clause 13.2, the tenderers have the option to quote for any one or more schedules and offer discounts for combined schedules. Such discounts wherever applicable will be taken into account to determine the lowest evaluated cost for the purchaser in deciding the successful tenderer for each schedule, subject to tenderer(s) being responsive.

34. Comparison of Tenders

Unless mentioned otherwise in Section – III – Special Instructions to Tenderers and Section – VI – List of Requirements, the comparison of the responsive tenders shall be carried out on Delivery Duty Paid (DDP) consignee site basis. The quoted turnkey prices and CMC prices will also be added for comparison/ranking purpose for evaluation. **Net Present value (NPV) of the Comprehensive Annual Maintenance charges (CMC) quoted for 5 years after the warranty period shall be added to the bid price for evaluation and will be calculated after discounting the quoted price by a discounting factor of 10% per annum.”**

35. Additional Factors and Parameters for Evaluation and Ranking of Responsive Tenders

35.1 Further to GIT Clause 34 above, the purchaser's evaluation of a tender will include and take into account the following:

- i) In the case of goods manufactured in India or goods of foreign origin already located in India, sales tax & other similar taxes and excise duty & other similar duties, Customs Duties, Service Tax, Works Contract Tax etc which will be contractually payable (to the tenderer), on the goods if a contract is awarded on the tenderer; and
- ii) in the case of goods of foreign origin offered from abroad, customs duty and other similar import duties/taxes, which will be contractually payable (to the tenderer) on the goods if the contract is awarded on the tenderer.

35.2 The purchaser's evaluation of tender will also take into account the additional factors, if any, incorporated in SIT in the manner and to the extent indicated therein.

36. Tenderer's capability to perform the contract

36.1 The purchaser, through the above process of tender scrutiny and tender evaluation will determine to its satisfaction whether the tenderer, whose tender has been determined as the lowest evaluated responsive tender is eligible, qualified and capable in all respects to perform the contract satisfactorily. If, there is more than one schedule in the List of Requirements, then, such determination will be made separately for each schedule.

36.2 The above-mentioned determination will, interalia, take into account the tenderer's financial, technical and production capabilities for satisfying all the requirements of the purchaser as incorporated in the TE document. Such determination will be based upon scrutiny and examination of all relevant data and details submitted by the tenderer in its tender as well as such other allied information as deemed appropriate by the purchaser.

37. Contacting the Purchaser

37.1 From the time of submission of tender to the time of awarding the contract, if a tenderer needs to contact the purchaser for any reason relating to this tender enquiry and / or its tender, it should do so only in writing.

37.2 In case a tenderer attempts to influence the purchaser in the purchaser's decision on scrutiny, comparison & evaluation of tenders and awarding the contract, the tender of the tenderer shall be liable for rejection in addition to appropriate administrative actions being taken against that tenderer, as deemed fit by the purchaser.

G. AWARD OF CONTRACT

38. Purchaser's Right to accept any tender and to reject any or all tenders

38.1 The purchaser reserves the right to accept in part or in full any tender or reject any or more tender(s) without assigning any reason or to cancel the tendering process and reject all tenders at any time prior to award of contract, without incurring any liability, whatsoever to the affected tenderer or tenderers.

39. Award Criteria

39.1 Subject to GIT clause 38 above, the contract will be awarded to the lowest evaluated responsive tenderer decided by the purchaser in terms of GIT Clause 36.

40. Variation of Quantities at the Time of Award/ Currency of Contract

40.1 If the quantity has not been increased at the time of the awarding the contract, the purchaser reserves the right to increase by up to twenty five (25) per cent, the quantity of goods and services mentioned in the contract (rounded off to next whole number) without any change in the unit price and other terms & conditions mentioned in the contract, during the currency of the contract after one year from the Date of Notification of Award.

41. Notification of Award

41.1 Before expiry of the tender validity period, the purchaser will notify the successful tenderer(s) in writing, by registered / speed post or by fax/ telex/cable (to be confirmed by registered / speed post) that its tender for goods & services, which have been selected by the purchaser, has been accepted, also briefly indicating therein the essential details like description, specification and quantity of the goods & services and corresponding prices accepted. The successful tenderer must furnish to the purchaser the required performance security within thirty days from the date of dispatch of this notification, failing which the EMD will be forfeited and the award will be cancelled. Relevant details about the performance security have been provided under GCC Clause 5 under Section IV.

41.2 The Notification of Award shall constitute the conclusion of the Contract.

42. Issue of Contract

42.1 Promptly after notification of award, the Purchaser/Consignee will send the contract form (as per Section XVI) duly completed and signed, in duplicate, to the successful tenderer by registered / speed post.

42.2 Within twenty one days from the date of the contract, the successful tenderer shall return the original copy of the contract, duly signed and dated, to the Purchaser/Consignee by registered / speed post.

42.3 The Purchaser/Consignee reserves the right to issue the Notification of Award consignee wise.

43. Non-receipt of Performance Security and Contract by the Purchaser/Consignee

43.1 Failure of the successful tenderer in providing performance security and / or returning contract copy duly signed in terms of GIT clauses 41 and 42 above shall make the tenderer

liable for forfeiture of its EMD and, also, for further actions by the Purchaser/Consignee against it as per the clause 24 of GCC – Termination of default.

44. Return of E M D

44.1 The earnest money of the successful tenderer and the unsuccessful tenderers will be returned to them without any interest, whatsoever, in terms of GIT Clause 19.6.

45. Publication of Tender Result

45.1 The name and address of the successful tenderer(s) receiving the contract(s) will be mentioned in the notice board/bulletin/web site of the purchaser.

46. Corrupt or Fraudulent Practices

46.1 It is required by all concerned namely the Consignee/Tenderers/Suppliers etc to observe the highest standard of ethics during the procurement and execution of such contracts. In pursuance of this policy, the Purchaser: -

(a) defines, for the purposes of this provision, the terms set forth below as follows:

(i) “corrupt practice” means the offering, giving, receiving or soliciting of any thing of value to influence the action of a public official in the procurement process or in contract execution; and

(ii) “fraudulent practice” means a misrepresentation of facts in order to influence a procurement process or the execution of a contract to the detriment of the Purchaser, and includes collusive practice among Tenderers (prior to or after Tender submission) designed to establish Tender prices at artificial non-competitive levels and to deprive the Purchaser of the benefits of free and open competition

(b) will reject a proposal for award if it determines that the Tenderer recommended for award has engaged in corrupt or fraudulent practices in competing for the contract in question;

(c) will declare a firm ineligible, either indefinitely or for a stated period of time, to be awarded a contract by the purchaser if it at any time determines that the firm has engaged in corrupt or fraudulent practices in competing for, or in executing the contract.

SECTION - III
SPECIAL INSTRUCTIONS TO TENDERERS
(SIT)

Sl. No.	GIT Clause No.	Topic	SIT Provision	Page No.
A	1 to 7	Preamble	No Change	33
B	8 to 10	TE documents	No Change	33
C	11 to 21	Preparation of Tenders	No Change	33
D	22 to 24	Submission of Tenders	No Change	33
E	25	Tender Opening	No Change	33
F	26 to 37	Scrutiny and Evaluation of Tenders	No Change	33
G	38 to 45	Award of Contract	No Change	33

**SPECIAL INSTRUCTIONS TO TENDERERS
(SIT)**

The following Special Instructions to Tenderers will apply for this purchase. These special instructions will modify/substitute/supplement the corresponding General Instructions to Tenderers (GIT) incorporated in Section II. The corresponding GIT clause numbers have also been indicated in the text below:

In case of any conflict between the provision in the GIT and that in the SIT, the provision contained in the SIT shall prevail.

- A Preamble**
No Change

- B TE documents**
No Change

- C Preparation of Tenders**
No Change

- D Submission of Tenders**
No Change

- E Tender Opening**
No Change

- F Scrutiny and Evaluation of Tenders**
No Change

- G Award of Contract**
No Change

SECTION - IV
GENERAL CONDITIONS OF CONTRACT (GCC)
TABLE OF CLAUSES

Sl No.	Topic	Page
1	Application	35
2	Use of contract documents and information	35
3	Patent Rights	35
4	Country of Origin	35
5	Performance Security	36
6	Technical Specifications and Standards	36
7	Packing and Marking	36
8	Inspection, Testing and Quality Control	37
9	Terms of Delivery	38
10	Transportation of Goods	38
11	Insurance	38
12	Spare parts	39
13	Incidental services	39
14	Distribution of Dispatch Documents for Clearance/Receipt of Goods	39
15	Warranty	40
16	Assignment	41
17	Sub Contracts	41
18	Modification of contract	41
19	Prices	42
20	Taxes and Duties	42
21	Terms and mode of Payment	42
22	Delay in the supplier's performance	44
23	Liquidated Damages	46
24	Termination for default	46
25	Termination for insolvency	46
26	Force Majeure	46
27	Termination for convenience	47
28	Governing language	47
29	Notices	47
30	Resolution of disputes	47
31	Applicable Law	48
32	Withholding and Lien	48
33	General/Miscellaneous Clauses	48

GENERAL CONDITIONS OF CONTRACT (GCC)

1. Application

- 1.1 The General Conditions of Contract incorporated in this section shall be applicable for this purchase to the extent the same are not superseded by the Special Conditions of Contract prescribed under Section V, List of requirements under Section VI and Technical Specification under Section VII of this document.

2. Use of contract documents and information

- 2.1 The supplier shall not, without the purchaser's prior written consent, disclose the contract or any provision thereof including any specification, drawing, sample or any information furnished by or on behalf of the purchaser in connection therewith, to any person other than the person(s) employed by the supplier in the performance of the contract emanating from this TE document. Further, any such disclosure to any such employed person shall be made in confidence and only so far as necessary for the purposes of such performance for this contract.
- 2.2 Further, the supplier shall not, without the purchaser's prior written consent, make use of any document or information mentioned in GCC sub-clause 2.1 above except for the sole purpose of performing this contract.
- 2.3 Except the contract issued to the supplier, each and every other document mentioned in GCC sub-clause 2.1 above shall remain the property of the purchaser and, if advised by the purchaser, all copies of all such documents shall be returned to the purchaser on completion of the supplier's performance and obligations under this contract.

3. Patent Rights

- 3.1 The supplier shall, at all times, indemnify and keep indemnified the purchaser, free of cost, against all claims which may arise in respect of goods & services to be provided by the supplier under the contract for infringement of any intellectual property rights or any other right protected by patent, registration of designs or trademarks. In the event of any such claim in respect of alleged breach of patent, registered designs, trade marks etc. being made against the purchaser, the purchaser shall notify the supplier of the same and the supplier shall, at his own expenses take care of the same for settlement without any liability to the purchaser.

4. Country of Origin

- 4.1 All goods and services to be supplied and provided for the contract shall have the origin in India or in the countries with which the Government of India has trade relations.
- 4.2 The word "origin" incorporated in this clause means the place from where the goods are mined, cultivated, grown, manufactured, produced or processed or from where the services are arranged.
- 4.3 The country of origin may be specified in the Price Schedule

5. Performance Security

- 5.1 Within Seven (7) days from date of the issue of notification of award by the Purchaser/Consignee, the supplier, shall furnish performance security to the Purchaser/Consignee for an amount equal to ten percent (10%) of the total value of the contract, valid up to sixty (60) days after the date of completion of all contractual obligations by the

supplier, including the warranty obligations, initially valid for a period of minimum 30 months from the date of Notification of Award.

- 5.2 The Performance security shall be denominated in Indian Rupees or in the currency of the contract as detailed below:
- a) It shall be in any one of the forms namely Account Payee Demand Draft or Fixed Deposit Receipt drawn from any Scheduled bank in India or Bank Guarantee issued by a Scheduled bank in India, in the prescribed form as provided in section XV of this document in favour of the Purchaser/Consignee. The validity of the Fixed Deposit receipt or Bank Guarantee will be for a period up to sixty (60) days beyond Warranty Period.
- 5.3 In the event of any failure /default of the supplier with or without any quantifiable loss to the government including furnishing of consignee wise Bank Guarantee for CMC security as per Proforma in Section XV, the amount of the performance security is liable to be forfeited. The Administration Department may do the needful to cover any failure/default of the supplier with or without any quantifiable loss to the Government.
- 5.4 In the event of any amendment issued to the contract, the supplier shall, within twenty-one (21) days of issue of the amendment, furnish the corresponding amendment to the Performance Security (as necessary), rendering the same valid in all respects in terms of the contract, as amended.
- 5.5 The supplier shall enter into Annual Comprehensive Maintenance Contract as per the 'Contract Form – B' in Section XVI with respective consignees, 3 (three) months prior to the completion of Warranty Period. The CMC will commence from the date of expiry of the Warranty Period.
- 5.6 Subject to GCC sub – clause 5.3 above, the Purchaser/Consignee will release the Performance Security without any interest to the supplier on completion of the supplier's all contractual obligations including the warranty obligations & after receipt of Consignee wise bank guarantee for CMC security in favour of Head of the Hospital/ Institute/ Medical College of the consignee as per the format in Section XV.

6. Technical Specifications and Standards

- 6.1 The Goods & Services to be provided by the supplier under this contract shall conform to the technical specifications and quality control parameters mentioned in 'Technical Specification' and 'Quality Control Requirements' under Sections VII and VIII of this document.

7. Packing and Marking

- 7.1 The packing for the goods to be provided by the supplier should be strong and durable enough to withstand, without limitation, the entire journey during transit including transshipment (if any), rough handling, open storage etc. without any damage, deterioration etc. As and if necessary, the size, weights and volumes of the packing cases shall also take into consideration, the remoteness of the final destination of the goods and availability or otherwise of transport and handling facilities at all points during transit up to final destination as per the contract.
- 7.2 The quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall strictly comply with the requirements as provided in Technical Specifications and Quality Control Requirements under Sections VII and VIII and in SCC under Section V. In case the packing requirements are amended due to issue of any amendment to the contract, the same shall also be taken care of by the supplier accordingly.
- 7.3 Packing instructions:

Unless otherwise mentioned in the Technical Specification and Quality Control Requirements under Sections VII and VIII and in SCC under Section V, the supplier shall make separate packages for each consignee (in case there is more than one consignee mentioned in the contract) and mark each package on three sides with the following with indelible paint of proper quality:

- a. contract number and date
- b. brief description of goods including quantity
- c. packing list reference number
- d. country of origin of goods
- e. consignee's name and full address and
- f. supplier's name and address

8. Inspection, Testing and Quality Control

- 8.1 The purchaser and/or its nominated representative(s) will, without any extra cost to the purchaser, inspect and/or test the ordered goods and the related services to confirm their conformity to the contract specifications and other quality control details incorporated in the contract. The purchaser shall inform the supplier in advance, in writing, the purchaser's programme for such inspection and, also the identity of the officials to be deputed for this purpose. The cost towards the transportation, boarding & lodging will be borne by the purchaser and/or its nominated representative(s) for the first visit. In case the goods are rejected in the first instance and the supplier requests for re-inspection, & if same is accepted by purchaser / consignee / PSA/ PA, all subsequent inspections shall be at the cost of the supplier. The expense will be to and fro. Economy Airfare, Local Conveyance, Boarding and Lodging of the inspection team for the inspection period."
- 8.2 The Technical Specification and Quality Control Requirements incorporated in the contract shall specify what inspections and tests are to be carried out and, also, where and how they are to be conducted. If such inspections and tests are conducted in the premises of the supplier or its subcontractor(s), all reasonable facilities and assistance, including access to relevant drawings, design details and production data, shall be furnished by the supplier to the purchaser's inspector at no charge to the purchaser.
- 8.3 If during such inspections and tests the contracted goods fail to conform to the required specifications and standards, the purchaser's inspector may reject them and the supplier shall either replace the rejected goods or make all alterations necessary to meet the specifications and standards, as required, free of cost to the purchaser and resubmit the same to the purchaser's inspector for conducting the inspections and tests again.
- 8.4 In case the contract stipulates pre-despatch inspection of the ordered goods at supplier's premises, the supplier shall put up the goods for such inspection to the purchaser's inspector well ahead of the contractual delivery period, so that the purchaser's inspector is able to complete the inspection within the contractual delivery period.
- 8.5 If the supplier tenders the goods to the purchaser's inspector for inspection at the last moment without providing reasonable time to the inspector for completing the inspection within the contractual delivery period, the inspector may carry out the inspection and complete the formality beyond the contractual delivery period at the risk and expense of the supplier. The fact that the goods have been inspected after the contractual delivery period will not have the effect of keeping the contract alive and this will be without any prejudice to the legal rights and remedies available to the purchaser under the terms & conditions of the contract.
- 8.6 The purchaser's/consignee's contractual right to inspect, test and, if necessary, reject the goods after the goods' arrival at the final destination shall have no bearing of the fact that the goods

have previously been inspected and cleared by purchaser's inspector during pre-despatch inspection mentioned above.

"On rejection, the supplier shall remove such stores within 14 days of the date of intimation of such rejection from the consignee's premises. If such goods are not removed by the supplier within the period mentioned above, the purchaser / consignee may remove the rejected stores and either return the same to the supplier at his risk and cost by such mode of transport as purchaser / consignee may decide or dispose of such goods at the suppliers risk to recover any expense incurred in connection with such disposals and also the cost of the rejected stores if already paid for."

8.7 Goods accepted by the purchaser/consignee and/or its inspector at initial inspection and in final inspection in terms of the contract shall in no way dilute purchaser's/consignee's right to reject the same later, if found deficient in terms of the warranty clause of the contract, as incorporated under GCC Clause 15.

8.8 **DELETED**

9. Terms of Delivery

9.1 Goods shall be delivered by the supplier in accordance with the terms of delivery and as per the delivery period specified in the schedule of requirement. Please note that the time shall be the essence of the contract.

10. Transportation of Goods

10.1 **DELETED**

10.2 Instructions for transportation of domestic goods including goods already imported by the supplier under its own arrangement:

In case no instruction is provided in this regard in the SCC, the supplier will arrange transportation of the ordered goods as per its own procedure.

11. Insurance:

11.1 Unless otherwise instructed in the SCC, the supplier shall make arrangements for insuring the goods against loss or damage incidental to manufacture or acquisition, transportation, storage and delivery in the following manner:

i) in case of supply of domestic goods on Consignee site basis, the supplier shall be responsible till the entire stores contracted for arrival in good condition at destination. The transit risk in this respect shall be covered by the Supplier by getting the stores duly insured. The insurance cover shall be obtained by the Supplier and should be valid till 3 months after the receipt of goods by the Consignee.

ii) **DELETED**

If the equipment is not commissioned and handed over to the consignee within 3 months, the insurance will be got extended by the supplier at their cost till the successful installation, testing, commissioning and handing over of the goods to the consignee. In case the delay in the installation and commissioning is due to handing over of the site to the supplier by the consignee, such extensions of the insurance will still be done by the supplier, but the insurance extension charges at actuals will be reimbursed.

12. Spare parts

- 12.1 If specified in the List of Requirements and in the resultant contract, the supplier shall supply/provide any or all of the following materials, information etc. pertaining to spare parts manufactured and/or supplied by the supplier:
- a) The spare parts as selected by the Purchaser/Consignee to be purchased from the supplier, subject to the condition that such purchase of the spare parts shall not relieve the supplier of any contractual obligation including warranty obligations; and
 - b) In case the production of the spare parts is discontinued:
 - i) Sufficient advance notice to the Purchaser/Consignee before such discontinuation to provide adequate time to the purchaser to purchase the required spare parts etc., and
 - ii) Immediately following such discontinuation, providing the Purchaser/Consignee, free of cost, the designs, drawings, layouts and specifications of the spare parts, as and if requested by the Purchaser/Consignee.
- 12.2 Supplier shall carry sufficient inventories to assure ex-stock supply of consumable spares for the goods so that the same are used during warranty and CMC period.

13. Incidental services

- 13.1 Subject to the stipulation, if any, in the SCC (Section – V), List of Requirements (Section – VI) and the Technical Specification (Section – VII), the supplier shall be required to perform the following services.
- i) Installation & commissioning, Supervision and Demonstration of the goods
 - ii) Providing required jigs and tools for assembly, minor civil works required for the completion of the installation.
 - iii) Training of Consignee's Doctors, Staff, operators etc. for operating and maintaining the goods
 - iv) Supplying required number of operation & maintenance manual for the goods

14. Distribution of Dispatch Documents for Clearance/Receipt of Goods

The supplier shall send all the relevant despatch documents well in time to the Purchaser/Consignee to enable the Purchaser/Consignee clear or receive (as the case may be) the goods in terms of the contract.

Unless otherwise specified in the SCC, the usual documents involved and the drill to be followed in general for this purpose are as follows.

- A) For Domestic Goods, including goods already imported by the supplier under its own arrangement

Within 24 hours of despatch, the supplier shall notify the purchaser, consignee, and others concerned if mentioned in the contract, the complete details of despatch and also supply the following documents to them by registered post / speed post (or as instructed in the contract):

- (i) Four copies of supplier's invoice showing contract number, goods description, quantity, unit price and total amount;
- (ii) Consignee Receipt Certificate as per Section XVII in original issued by the authorized representative of the consignee;
- (iii) Two copies of packing list identifying contents of each package;
- (iv) Inspection certificate issued by the nominated Inspection agency, if any.
- (v) Certificate of origin;
- (vi) Insurance Certificate as per GCC Clause 11.

(vii) Manufacturers/Supplier's warranty certificate & In-house inspection certificate.

B) DELETED

15. Warranty

- 15.1 The supplier warrants comprehensively that the goods supplied under the contract is new, unused and incorporate all recent improvements in design and materials unless prescribed otherwise by the purchaser in the contract. The supplier further warrants that the goods supplied under the contract shall have no defect arising from design, materials (*except when the design adopted and / or the material used are as per the Purchaser's/Consignee's specifications*) or workmanship or from any act or omission of the supplier, that may develop under normal use of the supplied goods under the conditions prevailing in India.
- 15.2 The **warranty** shall remain valid for the period as mentioned in the list of requirement/ General Technical specification, after the goods or any portion thereof as the case may be, have been delivered, installed and commissioned at the final destination and accepted by the purchaser / consignee (s) in terms of the contract, unless specified otherwise in the SCC
- a. No conditional warranty like mishandling, manufacturing defects etc. will be acceptable.
 - b. Warranty as well as Comprehensive Maintenance contract will be inclusive of all accessories and Turnkey work
 - c. Replacement and repair will be under taken for the defective goods.
 - d. Proper marking has to be made for all spares for identification like printing of installation and repair dates.
- 15.3 In case of any claim arising out of this warranty, the Purchaser/Consignee shall promptly notify the same in writing to the supplier. The period of the warranty will be as per G.C.C clause number 15.2 above irrespective of any other period mentioned elsewhere in the bidding documents.
- 15.4 Upon receipt of such notice, the supplier shall, within 8 hours on a 24(hrs) X 7 (days) X 365 (days) basis respond to take action to repair or replace the defective goods or parts thereof, free of cost, at the ultimate destination. The supplier shall take over the replaced parts/goods after providing their replacements and no claim, whatsoever shall lie on the purchaser for such replaced parts/goods thereafter. The penalty clause for non rectification will be applicable as per tender conditions
- 15.5 In the event of any rectification of a defect or replacement of any defective goods during the warranty period, the warranty for the rectified/replaced goods shall be extended till the completion of the original warranty period of the main equipment.
- 15.6 If the supplier, having been notified, fails to respond to take action to repair or replace the defect(s) within 8 hours on a 24(hrs) X 7 (days) X 365 (days) basis, the purchaser may proceed to take such remedial action(s) as deemed fit by the purchaser, at the risk and expense of the supplier and without prejudice to other contractual rights and remedies, which the purchaser may have against the supplier.
- 15.7 During Warranty period, the supplier is required to visit at each consignee's site at least once in 6 months commencing from the date of the installation for preventive maintenance of the goods
- 15.8 The Purchaser/Consignee reserve the rights to enter into Annual Comprehensive Maintenance Contract between Consignee and the Supplier for the period as mentioned in Section VII, Technical Specifications after the completion of warranty period.
- 15.9 The supplier along with its Indian Agent and the CMC provider shall ensure continued supply of the spare parts for the machines and equipments supplied by them to the purchaser for 10 years from the date of installation and handing over.
- 15.10 The Supplier along with its Indian Agent and the CMC Provider shall always accord most favoured client status to the Purchaser vis-à-vis its other Clients/Purchasers of its

equipments/machines/goods etc. and shall always give the most competitive price for its machines/equipments supplied to the Purchaser/Consignee.

16. Assignment

16.1 The Supplier shall not assign, either in whole or in part, its contractual duties, responsibilities and obligations to perform the contract, except with the Purchaser's prior written permission.

17. Sub Contracts

17.1 The Supplier shall notify the Purchaser in writing of all sub contracts awarded under the contract if not already specified in its tender. Such notification, in its original tender or later, shall not relieve the Supplier from any of its liability or obligation under the terms and conditions of the contract.

17.2 Sub contract shall be only for bought out items and sub-assemblies.

17.3 Sub contracts shall also comply with the provisions of GCC Clause 4 ("Country of Origin").

18. Modification of contract

18.1 If necessary, the purchaser may, by a written order given to the supplier at any time during the currency of the contract, amend the contract by making alterations and modifications within the general scope of contract in any one or more of the following:

- a) Specifications, drawings, designs etc. where goods to be supplied under the contract are to be specially manufactured for the purchaser,
- b) Mode of packing,
- c) Incidental services to be provided by the supplier
- d) Mode of despatch,
- e) Place of delivery, and
- f) Any other area(s) of the contract, as felt necessary by the purchaser depending on the merits of the case.

18.2 In the event of any such modification/alteration causing increase or decrease in the cost of goods and services to be supplied and provided, or in the time required by the supplier to perform any obligation under the contract, an equitable adjustment shall be made in the contract price and/or contract delivery schedule, as the case may be, and the contract amended accordingly. If the supplier doesn't agree to the adjustment made by the Purchaser/Consignee, the supplier shall convey its views to the Purchaser/Consignee within twenty-one days from the date of the supplier's receipt of the Purchaser's/Consignee's amendment / modification of the contract.

19. Prices

19.1 Prices to be charged by the supplier for supply of goods and provision of services in terms of the contract shall not vary from the corresponding prices quoted by the supplier in its tender and incorporated in the contract except for any price adjustment authorised in the SCC.

20. Taxes and Duties

20.1 Supplier shall be entirely responsible for all taxes, duties, fees, levies etc. incurred until delivery of the contracted goods to the purchaser.

20.2 Further instruction, if any, shall be as provided in the SCC.

21. Terms and Mode of Payment

21.1 Payment Terms

Payment shall be made subject to recoveries, if any, by way of liquidated damages or any other charges as per terms & conditions of contract in the following manner.

A) Payment for Domestic Goods Or Foreign Origin Located Within India.

Payment shall be made in Indian Rupees as specified in the contract in the following manner:

a) On delivery:

80 % payment of the contract price shall be paid within 15 days on receipt of goods in good condition and upon the submission of the following documents in all respect:

- (i) Four copies of supplier's invoice user manual, showing contract number, goods description, quantity, unit price and total amount;
- (ii) Consignee Receipt Certificate as per Section XVII in original issued by the authorized representative of the consignee;
- (iii) Two copies of packing list identifying contents of each package;
- (iv) Inspection certificate issued by the nominated Inspection agency, if any.
- (v) Insurance Certificate as per GCC Clause 11
- (vi) Certificate of origin.

b) On Acceptance:

Balance 20 % payment would be made within 30 days after satisfactory installation, testing & commissioning, training & training of personal & against 'Final Acceptance Certificate' as per Section XVIII of goods to be issued by the consignees subject to recoveries, if any, either on account of non-rectification of defects/deficiencies not attended by the Supplier or otherwise.

B) Payment for Imported Goods: DELETED

C) Payment of Turnkey, if any:

Turnkey payment will be made as indicated in the relevant Price Schedule.

D) Payment for Annual Comprehensive Maintenance Contract Charges:

The consignee will enter into CMC with the supplier at the rates as stipulated in the contract. The payment of CMC will be made on six monthly basis after satisfactory completion of said period, duly certified by the consignee on receipt of bank guarantee for an amount equivalent to 2.5 % of the cost of the equipment as per contract in the prescribed format given in Section XV valid till 2 months after expiry of entire CMC period.

21.2 The supplier shall not claim any interest on payments under the contract.

21.3 Where there is a statutory requirement for tax deduction at source, such deduction towards income tax and other tax as applicable will be made from the bills payable to the Supplier at rates as notified from time to time.

21.4 Irrevocable & non – transferable LC shall be opened by the respective consignees. However, if the supplier requests specifically to open confirmed LC, the extra charges would be borne by the supplier. If LC is required to be extended and/or amended for reasons not attributable to the purchaser/consignee, the charges thereof shall be borne by the supplier.

21.5 The payment shall be made in the currency / currencies authorised in the contract.

21.6 The supplier shall send its claim for payment in writing, when contractually due, along with relevant documents etc., duly signed with date, to respective consignees.

21.7 While claiming payment, the supplier is also to certify in the bill that the payment being claimed is strictly in terms of the contract and all the obligations on the part of the supplier for claiming that payment has been fulfilled as required under the contract.

21.8 While claiming reimbursement of duties, taxes etc. (like sales tax, excise duty, custom duty) from the Purchaser/Consignee, as and if permitted under the contract, the supplier shall also certify that, in case it gets any refund out of such taxes and duties from the concerned authorities at a later date, it (the supplier) shall refund to the Purchaser/Consignee forthwith.

- 21.9 In case where the supplier is not in a position to submit its bill for the balance payment for want of receipted copies of Inspection Note from the consignee and the consignee has not complained about the non-receipt, shortage, or defects in the supplies made, balance amount will be paid by the paying authority without consignee's receipt certificate after three months from the date of the preceding part payment for the goods in question, subject to the following conditions:
- (a) The supplier will make good any defect or deficiency that the consignee (s) may report within six months from the date of despatch of goods.
 - (b) Delay in supplies, if any, has been regularized.
 - (c) The contract price where it is subject to variation has been finalized.
 - (d) The supplier furnishes the following undertakings:

"I/We, _____ certify that I/We have not received back the Inspection Note duly received by the consignee or any communication from the purchaser or the consignee about non-receipt, shortage or defects in the goods supplied. I/We _____ agree to make good any defect or deficiency that the consignee may report within three months from the date of receipt of this balance payment.

22. Delivery

- 22.1 The supplier shall deliver of the goods and perform the services under the contract within the time schedule specified by the Purchaser/Consignee in the List of Requirements and as incorporated in the contract. The time for and the date of delivery of the goods stipulated in the schedule shall be deemed to be of the essence of the contract and the delivery must be completed not later than the date (s) as specified in the contract.
- 22.2 Subject to the provision under GCC clause 26, any unexcused delay by the supplier in maintaining its contractual obligations towards delivery of goods and performance of services shall render the supplier liable to any or all of the following sanctions:
- (i) imposition of liquidated damages,
 - (ii) forfeiture of its performance security and
 - (iii) termination of the contract for default.
- 22.3 If at any time during the currency of the contract, the supplier encounters conditions hindering timely delivery of the goods and performance of services, the supplier shall promptly inform the Purchaser/Consignee in writing about the same and its likely duration and make a request to the Purchaser/Consignee for extension of the delivery schedule accordingly. On receiving the supplier's communication, the Purchaser/Consignee shall examine the situation as soon as possible and, at its discretion, may agree to extend the delivery schedule, with or without liquidated damages for completion of supplier's contractual obligations by issuing an amendment to the contract.
- 22.4 When the period of delivery is extended due to unexcused delay by the supplier, the amendment letter extending the delivery period shall, inter alia contain the following conditions:
- (a) The Purchaser/Consignee shall recover from the supplier, under the provisions of the clause 23 of the General Conditions of Contract, liquidated damages on the goods and services, which the Supplier has failed to deliver within the delivery period stipulated in the contract.
 - (b) That no increase in price on account of any ground, whatsoever, including any stipulation in the contract for increase in price on any other ground and, also including statutory increase in or fresh imposition of customs duty, excise duty, sales tax/ VAT, Service Tax and Works Contract Tax or on account of any other tax or duty which may be levied in respect of the goods and services specified in the contract, which takes place after the date of delivery stipulated in the contract shall be admissible on such of the said

goods and services as are delivered and performed after the date of the delivery stipulated in the contract.

(c) But nevertheless, the Purchaser/Consignee shall be entitled to the benefit of any decrease in price on account of reduction in or remission of customs duty, excise duty, sales tax/ VAT, Service Tax and Works Contract Tax or any other duty or tax or levy or on account of any other grounds, which takes place after the expiry of the date of delivery stipulated in the contract.

22.5 The supplier shall not dispatch the goods after expiry of the delivery period. The supplier is required to apply to the Purchaser/Consignee for extension of delivery period and obtain the same before despatch. In case the supplier dispatches the goods without obtaining an extension, it would be doing so at its own risk and no claim for payment for such supply and / or any other expense related to such supply shall lie against the purchaser.

22.6 Passing of Property

22.6.1 The Property in the goods shall not pass to the purchaser unless and until the goods have been delivered to the consignee in accordance with the conditions of the contract.

22.6.2 Where there is a contract for sale of specific goods and the supplier is bound to do something to the goods for the purpose of putting them into a deliverable state the property does not pass until such thing is done.

22.6.3 Unless otherwise agreed, the goods remain at the supplier's risk until the property therein is transferred to the purchaser.

23. Liquidated damages

23.1 Subject to GCC clause 26, if the supplier fails to deliver any or all of the goods or fails to perform the services within the time frame(s) incorporated in the contract, the Purchaser/Consignee shall, without prejudice to other rights and remedies available to the Purchaser/Consignee under the contract, deduct from the contract price, as liquidated damages, a sum equivalent to 0.5% per week of delay or part thereof on delayed supply of goods and/or services until actual delivery or performance subject to a maximum of 10% of the contract price. Once the maximum is reached Purchaser/Consignee may consider termination of the contract as per GCC 24.

During the above-mentioned delayed period of supply and / or performance, the conditions incorporated under GCC sub-clause 22.4 above shall also apply.

24. Termination for default

24.1 The Purchaser/Consignee, without prejudice to any other contractual rights and remedies available to it (the Purchaser/Consignee), may, by written notice of default sent to the supplier, terminate the contract in whole or in part, if the supplier fails to deliver any or all of the goods or fails to perform any other contractual obligation(s) within the time period specified in the contract, or within any extension thereof granted by the Purchaser/Consignee pursuant to GCC sub-clauses 22.3 and 22.4.

24.2 In the event of the Purchaser/Consignee terminates the contract in whole or in part, pursuant to GCC sub-clause 24.1 above, the Purchaser/Consignee may procure goods and/or services similar to those cancelled, with such terms and conditions and in such manner as it deems fit and the supplier shall be liable to the Purchaser/Consignee for the extra expenditure, if any, incurred by the Purchaser/Consignee for arranging such procurement.

24.3 Unless otherwise instructed by the Purchaser/Consignee, the supplier shall continue to perform the contract to the extent not terminated.

25. Termination for insolvency

25.1 If the supplier becomes bankrupt or otherwise insolvent, the purchaser reserves the right to terminate the contract at any time, by serving written notice to the supplier without any compensation, whatsoever, to the supplier, subject to further condition that such termination will not prejudice or affect the rights and remedies which have accrued and / or will accrue thereafter to the Purchaser/Consignee.

26. Force Majeure

26.1 Notwithstanding the provisions contained in GCC clauses 22, 23 and 24, the supplier shall not be liable for imposition of any such sanction so long the delay and/or failure of the supplier in fulfilling its obligations under the contract is the result of an event of Force Majeure.

26.2 For purposes of this clause, Force Majeure means an event beyond the control of the supplier and not involving the supplier's fault or negligence and which is not foreseeable and not brought about at the instance of , the party claiming to be affected by such event and which has caused the non – performance or delay in performance. Such events may include, but are not restricted to, acts of the Purchaser/Consignee either in its sovereign or contractual capacity, wars or revolutions, hostility, acts of public enemy, civil commotion, sabotage, fires, floods, explosions, epidemics, quarantine restrictions, strikes excluding by its employees , lockouts excluding by its management, and freight embargoes.

26.3 If a Force Majeure situation arises, the supplier shall promptly notify the Purchaser/Consignee in writing of such conditions and the cause thereof within twenty one days of occurrence of such event. Unless otherwise directed by the Purchaser/Consignee in writing, the supplier shall continue to perform its obligations under the contract as far as reasonably practical, and shall seek all reasonable alternative means for performance not prevented by the Force Majeure event.

26.4 If the performance in whole or in part or any obligation under this contract is prevented or delayed by any reason of Force Majeure for a period exceeding sixty days, either party may at its option terminate the contract without any financial repercussion on either side.

26.5 In case due to a Force Majeure event the Purchaser/Consignee is unable to fulfil its contractual commitment and responsibility, the Purchaser/Consignee will notify the supplier accordingly and subsequent actions taken on similar lines described in above sub-paragraphs.

27. Termination for convenience

27.1 The Purchaser/Consignee reserves the right to terminate the contract, in whole or in part for its (Purchaser's/Consignee 's) convenience, by serving written notice on the supplier at any time during the currency of the contract. The notice shall specify that the termination is for the convenience of the Purchaser/Consignee. The notice shall also indicate interalia, the extent to which the supplier's performance under the contract is terminated, and the date with effect from which such termination will become effective.

27.2 The goods and services which are complete and ready in terms of the contract for delivery and performance within thirty days after the supplier's receipt of the notice of termination shall be accepted by the Purchaser/Consignee following the contract terms, conditions and prices. For the remaining goods and services, the Purchaser/Consignee may decide:

- a) To get any portion of the balance completed and delivered at the contract terms, conditions and prices; and / or
- b) To cancel the remaining portion of the goods and services and compensate the supplier by paying an agreed amount for the cost incurred by the supplier towards the remaining portion of the goods and services.

28. Governing language

28.1 The contract shall be written in English language following the provision as contained in GIT clause 4. All correspondence and other documents pertaining to the contract, which the parties exchange, shall also be written accordingly in that language.

29. Notices

29.1 Notice, if any, relating to the contract given by one party to the other, shall be sent in writing or by cable or telex or facsimile and confirmed in writing. The procedure will also provide the sender of the notice, the proof of receipt of the notice by the receiver. The addresses of the parties for exchanging such notices will be the addresses as incorporated in the contract.

29.2 The effective date of a notice shall be either the date when delivered to the recipient or the effective date specifically mentioned in the notice, whichever is later.

30. Resolution of disputes

30.1 If dispute or difference of any kind shall arise between the Purchaser/Consignee and the supplier in connection with or relating to the contract, the parties shall make every effort to resolve the same amicably by mutual consultations. However, the decision of Purchaser/ HOD will be final & binding to supplier.

30.2 If the parties fail to resolve their dispute or difference by such mutual consultation within twenty-one days of its occurrence, then, unless otherwise provided in the SCC, either the Purchaser/Consignee or the supplier may give notice to the other party of its intention to commence arbitration, as hereinafter provided the applicable arbitration procedure will be as per the Arbitration and Conciliation Act, 1996 of India. In the case of a dispute or difference arising between the Purchaser/Consignee and a domestic Supplier relating to any matter arising out of or connected with the contract, such dispute or difference shall be referred to the sole arbitration of an officer in the Ministry of Law and Justice, appointed to be the arbitrator by the Director General (Health Services). The award of the arbitrator shall be final and binding on the parties to the contract subject to the provision that the Arbitrator shall give reasoned award in case the value of claim in reference exceeds Rupees One lakhs (Rs. 1,00,000/-)

30.3 Venue of Arbitration: The venue of arbitration shall be the place from where the contract has been issued, i.e., New Delhi, India.

30.4 Jurisdiction of the court will be form the place where the tender enquiry document has been issued, i.e., New Delhi, India.

31. Applicable Law

The contract shall be governed by and interpreted in accordance with the laws of India for the time being in force.

32. Withholding and Lien in respect of sums claimed

Whenever any claim for payment arises under the contract against the supplier the purchaser shall be entitled to withhold and also have a lien to retain such sum from the security deposit or sum of money arising out of under any other contact made by the supplier with the purchaser, pending finalization or adjudication of any such claim.

It is an agreed term of the contract that the sum of money so withheld or retained under the lien referred to above, by the purchaser, will be kept withheld or retained till the claim arising about of or under the contract is determined by the Arbitrator or by the competent court as the case may

be, and the supplier will have no claim for interest or damages whatsoever on any account in respect of such withholding or retention.

32. General/ Miscellaneous Clauses

- 32.1 The Supplier shall notify the Purchaser/Consignee /the Government of India of any material change would impact on performance of its obligations under this Contract.
- 32.2 Each member/constituent of the Supplier/its Indian Agent/CMC Provider, in case of consortium shall be **jointly and severally liable** to and responsible for all obligations towards the Purchaser/Consignee/Government for performance of contract/services including that of its Associates/Sub Contractors under the Contract.
- 32.3 The Supplier/its Indian Agent/CMC Provider shall at all times, indemnify and keep indemnified the Purchaser/Government of India against all claims/damages etc. for any infringement of any Intellectual Property Rights (IPR) while providing its services under CMC or the Contract.
- 32.4 The Supplier/its Agent/CMC Provider shall, at all times, indemnify and keep indemnified the Purchaser/Consignee/Government of India against any claims in respect of any damages or compensation payable in consequences of any accident or injury sustained or suffered by its employees or agents or by any other third party resulting from or by any action, omission or operation conducted by or on behalf of the supplier/its associate/affiliate etc.
- 32.5 All claims regarding indemnity shall survive the termination or expiry of the contract.

SECTION – V

SPECIAL CONDITIONS OF CONTRACT (SCC)

The following Special Conditions of Contract (SCC) will apply for this purchase. The corresponding clauses of General Conditions of Contract (GCC) relating to the SCC stipulations have also been incorporated below.

These Special Conditions will modify/substitute/supplement the corresponding (GCC) clauses. Whenever there is any conflict between the provision in the GCC and that in the SCC, the provision contained in the SCC shall prevail.

The warranty conditions will be as mentioned in the list of requirement as per section VI of the tender enquiry.

SECTION - VI**LIST OF REQUIREMENTS****Part I**

Sl. No.	Equipment Name	Consignee	Department	Quantity	Total Quantity	EMD Details
2						
3						
4						
5						
6						

Part II: Required Delivery Schedule:**a) For Indigenous goods or for imported goods if supplied from India:**

30 days from date of Notification of Award to delivery at consignee site. The date of delivery will be the date of delivery at consignee site (Tenderers may quote earliest delivery period).

Installation and commissioning shall be done within one weeks of receipt of the stores/ goods at site

For delayed delivery and/ or installation and commissioning liquidated damages will get applied as per GCC clause 23.

Part III: Scope of Incidental Services:

Installation & Commissioning, Supervision, Demonstration, Trial run and Training etc. as specified in GCC Clause 13

Part IV:

Turnkey (if any) as per details in Technical Specification.

Part V:

Warranty period as per details in general technical specifications

Comprehensive Maintenance Contract (CMC) as per details in Technical Specification and also specified in part I above.

Part VI:**Required Terms of Delivery and Destination.****a) For Indigenous goods or for imported goods if supplied from India:**

At Consignee Site(S)

Destination/Consignee details are given in Section XXI.

Section – VII
TECHNICAL SPECIFICATIONS

RADIOLOGY (IMAGING)

Ultrasound Machine

The system should be latest fully Digital Color Doppler Ultrasound System and can be used for applications like Abdominal, Obs. / Gynae, small parts, Endocavitary, Pediatric & Vascular applications. The system should have following essential features:

1. The system should have the following image modes: 2D, M mode, PW, Tissue Harmonic mode, Color Doppler, Power Doppler mode.
2. The system should have minimum 1500 or more digital processing channels and 256 or more grey shades.
3. The system should have a very high frame rate for B-mode & Colour mode. Maximum frame rate should be greater than 350 fps for B-mode & colour mode.
4. The system should have a frame rate of at least 300 frames per second (fps) in B mode. Please specify the maximum frame rate in B-mode & M-mode.
5. The system should be able to support all type of transducers (Convex, Endocavitary, Linear, Phased array and Intraoperative Transducers). Frequency range of all transducers should be 2-14Mhz.
6. The system should have advanced measurement packages for all applications.
7. The system should an integrated high resolution TFT/LCD of 17 inches or more with facility of tilt and swivel facility alongwith convenient grip.
8. The system should have minimum three active universal ports & two parking ports. Active ports can be directly selectable from the control panel.
9. The system should have scanning depth in the range of 2- 28cms.
10. The system should have a very high capacity of Hard Disc Drive min.80GB or 1000 images for storage of images
11. The system should have inbuilt CD/DVD R/W and USB ports for image export
12. The system should have zoom facility both in real time and frozen image and it should be minimum 6 times or more in both real time & frozen modes.
13. The system should have minimum 6 steps transmitting focussing (transmit focal zones) and adjustable gain should be available up to 100 dB for B-mode & M-mode.
14. The system should have Directional Power Doppler to define the low blood flow directions.
15. The system should have HD-flow/Advanced dynamic flow to acquire the blood flow with directions in the deeper region at a very high frame rate.
16. The system should have automatic optimization in B-mode and auto adjustment of Doppler base-line & velocity range.

17. The system should have B-mode image steering & Color Doppler steering . Please mention the angle.
18. The system should have the facility of on-screen adjustment for Dynamic range, Frequency selection, Presets, Name of the patient,etc.
19. The system should have the facility to view the Thumbnail images and system can be programmed for various users with the facility of user passwords.
20. The system should have the Trapezoid scan facility for linear probes.
21. The system should have Compound Imaging and Contrast Harmonic Imaging.
22. The system should have the facility of having direct image print out through a B/W thermal printer.
23. The system should be upgradeable to real time 3D (4D) package. Please quote optionally for convex volume probe.
24. System should be offered with the following probes and accessories:
 - (a) Convex probe with frequency range of 2.0-5.0 Mhz.
 - (b) TV/TR probe with frequency range of 5.0-7.5 Mhz. (± 1 MHz) And minimum field of view of 140 degree.
 - (c) Linear probe with frequency range of 6.0-11.0 Mhz. (± 1 MHz)
 - (d) B/w Thermal Printer with 100 paper rolls.Above mentioned probes must have multifrequency selection and THI.
25. Two years complete warranty for the entire equipment, probes and accessories which should include service as well as parts.

Please attach the original manufacture's product catalog and datasheets. Photocopied, computer generated catalogue and datasheet will not be accepted.
26. List of installation – the bidders to provide list of installation of the quoted model in (National and International).

The shortlisted bidders will have to give demonstration of their quoted model before finalizing the evaluation of their bids.
27. The bidder should enclose the original product data sheet, brochure and compliance sheet, without which the bid will be rejected. Computer generated data sheet and brochure will not be accepted. The serial number of specifications must be indicated against the relevant portion of the compliance sheet and data sheet.
28. Valid US FDA or CE (European) approved Product.
29. Essential accessories: Black & White Thermal printer and color laser printer, Suitable UPS with one hour backup to support all functions of the equipment i.e. performing ultrasound procedure, exposure on to films or copy on a CD., mobile cart with transducer holder, jelly bottle holder and space for printer.

HIGH FREQUENCY 300 MA X-RAY MACHINE

High Frequency X – Ray machine suitable for general Radiography.

X-Ray Generator

High frequency X-ray Generator having Frequency of 20 KHz or more suitable for Radiography Should be provided.

Power Output of Generator should be 24 KW or more.

Radiography K range should be 40 to 110 KV ore more.

mAs :- 300 mAs or more

Exposure time (Rad.) : 1 ms to 3 sec . with maximum number of steps.

Control:

A very compact, soft Touch Control Panel having following functions & indications should be provided. The panel can be supplied in Floor or Wall mount with Spill Proff design following features should be available on Control panel.

- Digital display of Kv & mAs
- KV increase & decrease switches
- KV or mAs increase & decrease switches
- Machine ON/OFF switch.
- Tube focal spot selection switch.
- Ready & X ray on switch with indicators.
- Bucky Selection switch.
- Self diagnostic programmed with indicators for :-
 - Earth fault error
 - Kv error
 - Filament error
 - Tube Thermal error.
- Anatomical Programming Radiography should be provided in which KV and mAs are automatically selected depending upon the physique of the patient and part of the body to be X-rayed.

A dual action hand switch retractable cord should be provided if Radiation Protection of Operator.

X- Ray TUBE:-

- One No. Dual focus Rotating Anode BEL / Toshiba / Imported X-ray tube thermally protected having focal spot 1 x 2mm sq
- Anode heat storage capacity of tube should be more than 100 KHU.
- One pair of meter H.V Cable.
- One no manual collimator with aluminium filter & for adjustment of exposure area.

Column Stand:

- It should have floor to ceiling stand with verticlal counter balanced travel.
- It should have 360° Rotation
- It should be provided one chest stand with machine.

Table:-

- Five position manual tilt table having bucky grid ratio of with 60 lines per inches should be provided.
- The bucky tray should accept cassette upto 14'' x 17'' size.

Accessories:-

- Three panel x ray protective barrier for operator with lead glass window in the control panel
- One pair of light lead apron for protection of operator.
- One Vertical Bucky.

Power requirement: -

415 Volt AC Current, 50 H4 15 Amps (with line resist less then 0.4 ohms line regulation+/- 10%) 3 Phase should be available there.

Other terms & conditions:

Guarantee of machine

- 2 years from the date of installation including tube.
- The company should be ISO & European CE Certified.
- The Unit should be Type approved from AERB.
- Minor electrical /civil work for functioning of the machine in the room if any will be the responsibility of the supplier.
- Demonstration may be taken if required

SPECIFICATIONS OF “HEAVY DUTY AUTOMATIC FILM PROCESSOR”

Automatic X-Ray Film Processor for processing of all standard medical X-Ray and Imaging Films. It should have the following features:-

1. Processor should be able to take the maximum film width and infeed width shall be 45 cm and minimum of 10 cm x 10 cm.
 2. High capacity through put of more than 220 Films/Hour of size 14 x 17”.
 3. Transport speed adjustable between atleast 1.0 minute to 5.0 minutes (in the steps of maximum 8 seconds).
 4. The developing and washing tanks should be minimum of 12 Litres capacity for proper developing & washing. Fixer tank may be of higher capacity.
 5. Replenishment should have the individual programmes manual & automatic with tank capacity of 25 Litres.
 6. The roller transport system should have squeeze rollers and the main drive should stop automatically when not in use.
 7. The container assembly should be monoshell and made of material which is non corrosive and of latest technology.
 8. Thermostatic controlled temperature of Developer and it should have range between 28⁰ C to 40⁰ C.
 9. Multiple program memory should be there preferred for different Application. Atleast five program memory must be there.
 10. There should be atleast four infrared sensors for film detection.
 11. Replenishment rate should be adjustable and the adjustment range must be 200-2000 ml. per sq. m.
 12. Driver temperature range must be upto 70⁰ C for better drying.
 13. Water consumption during use should not exceed 2 Lit. per minute, Lower consumption shall be preferred.
 14. It should have anti-oxidation programme option in between adjustable range of 10-90 minute intervals.
-

RADIOLOGY ACCESSORIES:

X-Ray Cassettes with intensifying screening

- Light weight
- Front Side – 1mm Aluminium equivalent
- Backside – 0.1mm lead equivalent. Smooth felt pad fitted to the backside for attaching the intensifying screen.
- Locking system – Push-Lock type

Intensifying Screens:

- High Speed
- Calcium tungstate/ Rare Earth Phosphor (Compatible with double emulsion regular Medical film)
- Blue light emitting (for Ca Tungstate screens)
- Polyester base
- Universal or Front – Back type
- Should be easy to clean.

Sizes: Qty: 6 each

14 x 17
14 x 14
15 x 12
12 x 12
10 x 8
12 x 10

X-Ray Stationary Grid (Grid Ratio 6:1, Parallel)

Size 12" x 15", Size 10" x 12" -1No.

Metal Cassette Passbox (4 Doors)

Lead Lines Box for 4 Packets size LWH (19"x11"x19")

Lead lines box for 6 packets. Size LWH (19"x19"x19")

Lead lines box for 12 packets size LWH (19"x 29½"x19")

Feather Weight Apparel (Lead Aprons):

BARC approved, 0.5mm Lead Equivalent

Medium; (100 x60 cms) 4 Nos.

Regular; (110 x 60cms) 4 Nos.

THYROID SHIELD - 2 Nos.

Protective Gloves - 3 Nos.

Lead Goggles: 1mm Pb front and side - 2 Nos.

Wall Mounted Lead Apron Hanger, capable of keeping multiple aprons (At least 6)

X-ray View Box: (Regular long fluorescent tube type, electronic choke, bright screen, aluminum body). Size: 2 x 1

X-ray View Box: (Regular long fluorescent tube type, electronic choke, bright screen, aluminum body). Size : 5 X 1 (Table mounted)

Lead marker - 4 set

Lead Letter (Big and small sizes) (0-9, A-Z, R & L) – 50sets

For X-ray Films

LABORATORY

HAEMATOLOGY ANALYZER AUTOMATIC 3 PART DIFFERENTIAL

1. Should be a fully automated hematology analyzer providing 18 parameters including a 3- part differential, with user definable settings for RDW – CV or RDW – SD.
2. The system should give the Differential count as Lymphocytes, Mid population & **Neutrophils/Granulocytes** While Mid population should include Eosinophils, Basophils and Monocytes.
3. The system should be capable of processing samples at a speed of 60samples/hour.
4. The system should be **Sample Rotary Valve (SRV) or equivalent** based for the precise sample aliquoting for dilutions.
5. The system should have large LCD display to have a review of all the results along with the three histograms of WBC, RBC and PLT on the screen.
6. The system should have around 200 samples test result memory.
7. The system should have autoprobe wiper to clean the sample probe automatically after sample aspiration.
8. The system should use non-cyanide based reagent for Hb estimation.
9. The system should have an option to print the results with or without histograms also with the option to print only basic 8 parameters.
10. The system should use the proven and approved “volumetric Metering” system of cell counting, for WBC’s, RBC’s & PLT’s for high precision of the results and stability of the calibration.
11. The system should have a system of count and aperture monitoring every 0.5 sec for precision and reliability of the counts.
12. The system should have automatic floating thresholds for the correct separation of RBC’s and PLT’s during overlap in cases of Microcytosis / large platelet.
13. The system should also have additional facility for manual discrimination in order to process veterinary sample.
14. System should not require any daily maintenance except daily shutdown.
15. The system should automatically give an alarm to the operator for doing the maintenance.
16. The system should use high intensity LED for Hb estimation and not the lamp.
17. The system should be open and reagent from other company can also be used
18. All reagents required should be available locally from the Company or its authorized distributor. Cost of consumables shall be considered in financial comparison. Two **vials each of 3 level quality controls (ie total 6 vials) should be provided for initial training and validation of instrument.**
19. Firm will have to supply the UPS with 30 min back up along with the equipment free of cost
20. User’s list should be attached with satisfactory report for the last three years from three users with contact details.
21. Up time & penalty for delays in repair & maintenance: The firm will ensure uptime of 345 days in a year during warranty period & CMC period of both equipment as well as UPS. Whenever there is breakdown the firm will carry out the repair within 48 hours of receipt of such Information (either by telephone or by any other means)

If the down time is exceeded in a year from 20 days then the warranty shall stand extended by double the number of days machine was out of order.

22. Information regarding merger/acquisition/takeover or any change in the production should be submitted at the time of tender by the principal firm. In such case it should be specified who will provide the after sale service, CMC, supply of spare parts etc. failing which the firm shall not be considered for technical evaluation.

**RANDOM ACCESS SMALL THROUGHPUT FULLY AUTOMATED
CLINICAL CHEMISTRY ANALYSER**

1. **SYSTEM:** Floor/Bench top Model, Completely Open, Discreet, Multi-channel, Random Access, With automatic rerun, automatic reflex testing and capable of performing tests like Enzymes, substrates, Serum Proteins, Electrolytes, TDM assays and Immunoturbidimetric etc.
2. **THROUGH PUT:** About 400 Photometric tests/Hour and about 600 Tests /Hour with ISE.
3. **ASSAY MODES:** End point, Rate, fixed point and ISE.
4. **Analytical Methods:** Colorimetry, turbidometry, latex agglutination, homogeneous, ISE
5. **SAMPLE LOADING:** Minimum of 50 sample positions with continuous Loading .Bar code reading facility for positive sample identification, real time test requisition downloading from host should be possible.
6. Cooled compartment for Standards and Controls.
7. **SAMPLE CUPS:** Primary and secondary tubes and paediatric cups.
8. **SAMPLE TYPES:** Plasma, Urine, Serum, CSF etc.
9. **STAT FACILITY:** Facility for continuous loading of stat samples without interrupting the routine run. Minimum 20 STAT sample positions for very urgent samples.
10. **SAMPLE VOLUME:** 2 to 30 micro litres in 1.0 micro litre increment.
11. **SAMPLE PROBE:** Probe liquid level sensor. Sample/Reagent Bubble & clot detection. Crash prevention facility should be available.
12. **REAGENT DISK:** Refrigerated reagent disk with minimum 50 positions.
13. **ON-BOARD PARAMETERS TESTS:** Minimum 50 on-board parameters tests.
14. **REACTION VOLUME:** Should be from 150 ul to 300ul.
15. **REAGENT PROBE:** Probe with liquid level sensors and washing facility. Probe crash detection should be available.
16. **STIRRER:** 2 or more on board variable speed stirrers should be available.
17. **CUVETTES:** may be reusable, permanent or Disposable, specify recurring cost, if any
18. **CUVETTE WASHING:** Automatic on-board washing.
19. **PHOTOMETER:** Wavelength ranging from 340 - 750 nm.
20. **LAMP SOURCE:** Halogen / Xenon Lamp.
21. **QUALITY CONTROL:** Real Time, Individual and cumulative quality control. Automatic QC programming required.

22. Water Plant: Compatible RO/water purification plant to be supplied.
23. SOFTWARE: Window based software or compatible.
24. DATA STORAGE: 50000 patient samples.
25. INTER FACE: Unidirectional and Bidirectional communication possible.
26. REAGENTS: Manufacturing Company if have their own system reagents, controls and calibrators and the price list for the same should be enclosed with the price bid.
27. Accessories, reagents calibrator and control: company shall provide a list of accessories reagents calibrator and control to be use for running the instrument.
28. The equipment to be supplied should have FDA and CE certification.
29. Suitable on-Line UPS with One hour backup.

AUTOCLAVE (VERTICAL)

1. Automatic adjustable working pressure system.
2. Double walled.
3. Inside boiler made of stainless steel & outside mild steel finished in cream enamel.
4. Radial locking system lid.
5. The panel is provided with on/off switch, pressure gauge, steam release valve & indicators to show the working of mains & pressure control system.
6. Electrically operated on 220V A/C with stainless steel basket.
7. Digital Temperature controller with inbuilt timer that will cut off the heaters automatically after lapse of pre set time at the pre set temperature.
8. Microprocessor based PID controller.
9. Height: 600mm.
10. Diameter inside: 450mm(24|x18|)
11. Should be FDA or CE or BIS approved product

GYN. OPD

LED SPOT EXAMINATION LIGHT

Should be with LED Examination light, mounted on a pedestal stand Lux Output 50000

LED lamp with 50000 to 100000 hours lamp life

Cool Beam at similar to that of the day 5500 deg K.

Illumination field 30cm diameter or more.

Arm Flexibility for easy positioning.

Should be European CE / US FDA standard equipment.

DENTAL OPD

ELECTRICALLY OPERATED FULLY PROGRAMMABLE DENTAL CHAIR

Body contoured Self-Adjusting Electrically operated and Multi Programmable. The program should be Micro processor controlled, where doctor can set his own programme. The chair should have all the movements, i.e. up and down. The Back-rest complete flat position to up-right position with foot switch to adjust all the movements of the Chair during Surgery, when Doctor cannot use his hands. Right side arm with lateral rotation for easy exit of the patient. Feather touch Chair control should be on Delivery unit as well as on assistant side.

Dental unit should have latest Continental Overhead Delivery System. Made up of Aluminum alloy body. It should have heat-cure quality paint and base covers should be of Fibre glass material. The chair should be thickly cushioned and covered with high quality Rexene of soothing colour.

Chair Mount Unit fitted with:

- a. Operating Light having Multi surface glass reflector with two intensities control & sensor.
- b. Auto water connection for Spittoon and Tumbler.
- c. Stainless Steel Instrument tray for keeping instruments
- d. X-ray Viewer OPG size led based.
- e. High & Low vacuum motorized suction should with in built system to disinfect with non retraction valve, double bottle auto drain and autoflush system.

The Modular should be fitted with:

- a. Airtor
Airtor units with control box fitted with transparent regulated water tank and foot control. It should be supplied with autoclavable hand piece having a speed of 4,00,000 RPM Clean head with quick-change coupling, non-retraction valve & 360 degree rotation.
- b. Micromotor
Micro-motor (brush less) should be high torque, having range of 2,000 to 40,000 RPM. The Control Unit should have foot and hand control of speed. It should be supplied with autoclavable contra-angle & straight hand – piece.
- c. Ultrasonic scaler (piezo type)
The scaler unit should be supplied with autoclavable hand piece and 3 different working tips.
- d. Three way autoclavable syringe for air & water spray (one on the unit side & other on the Assistant side.
- e. Cordless Light cure LED unit with gun and charger

It should be supplied with the Tip, filter, bulb and eye protection disc. This should give intensity of above 450 NM. Continuous rating, should be more than 2 minutes. Every 10 seconds there should be a beep sound. It should also have built – in voltage regulation for reliable intensity.

- b. Zero back-ache stool (2 in no.)
The surgical stool should be having raising and lowering facility by pneumatic piston. The back-rest should move forward and backward along with the body by pneumatic piston and should be fixable at any convenient position. The back rest should also have height adjustment. The seat should have a piston to move with the body when the surgeon leans forward. It should be freely movable on well-balanced five wheels.
- c. Compressor
The compressor should be Oil-free Medical grade Mono-bloc $\frac{3}{4}$ HP type with air displacement capacity of 100 ltrs./min. and a pressure 60 to 80 Psi. It is to be fitted with

radiator, cooler, vertical column filter, stainless steel tank for retrogrative cleaning and auto drain valve with silencer. The inner surface of the compressor tank should be epoxy coated to prevent rusting.

- Built – in thermo cut-out to save the motor during excess of heat
- Auto head air-release valve
- Safety release valve
- Drain valve and Gauge

d. Intra oral camera

- Direct USB connectivity
- Single cord operated
- Light weight (About 50 gm)
- Four cool white LED light.
- Takes Intra/extra oral images.
- Takes power from USB cable.
- Detachable coupling with softly lock system.
- Auto focus with intra/ extra oral image.
- Operate by hand piece and footswitch.
- RVG software compatibility.
- Having latest microprocessor technology; means image capture should be a function of the camera every image can be freezed directly on Laptop/ computer.
- Having single lens (8.0x5.9mm) focusing adjust intra / Extra oral image .
- Having automatic magnetic on/off and clear all buttons to save time, energy and money and
- Having lightweight chassis (45gm) for easy to use.
- Disposable hand piece sleeves – 10,000 pieces.
- High resolution
- All relevant software.
- LED monitor of 24” screen (minimum), to be mounted on Dental Chair Unit.
- Compatible with Indian power supply.

It should be provided with Servo Stabilizer and Power back up with required capacity.

THE COMPANY MANUFACTURING THESE CHAIRS SHOULD HAVE LATEST ISO CERTIFICATE COMPANY.

All installation including copper pipe for air.

Comprehensive Guaranteed for 2 years including electric, electronic and rubber parts

JUSTIFICATION: For routine surgical work.

DENTAL X-RAY

1 Description of Function

1.1 Dental x-ray machine is used for taking Intra Oral Periapical and Occlusal X-ray

2 Operational Requirements

2.1 High resolution RVG based on CCD/CMOS technology

3 Technical Specifications

3.1 Based on DC current

3.2 Tube voltage, selection: 60-65-70 kVp

3.3 Tube current 6 mA/ 8 mA

- 3.4 Focal spot 0.8 x 0.8 mm
- 3.5 Total filtration > 2 mm Al
- 3.6 Minimum range of exposure time range – 0.02 to 3.2 secs
- 3.7 Manufactured with International Safety standards for radiation leakage
- 3.8 Electronic selection of exposure time/radiation according to tooth number. It should be possible to select exposure time manually.

4 System Configuration Accessories, spares and consumables

- 4.1 X-ray unit should be supplied with lead apron, thyroid collar and gonadal sheath
- 4.2 Should be supplied with lead partition

5 Power Supply

- 5.1 Power input to be 220-240VAC, 50Hz fitted with Indian plug
- 5.2 Servo Voltage stabilizer of appropriate ratings meeting ISI Specifications.(Input 160-260 V and output 220-240 V and 50 Hz)

6 Standards, Safety and Training

- 6.1 Should be US FDA/ CE approved product
- 6.2 Manufacturer/ Supplier should have ISO certification for quality standards.
- 6.3 Electrical safety for dental x-ray unit conforms to standards for electrical safety IEC-60601 / IS-13450

7 Documentation

- 7.1 User/Technical/Maintenance manuals to be supplied in English.
- 7.2 List of important spare parts and accessories with their part number and costing
- 7.3 Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist. The job description of the hospital technician and company service engineer should be clearly spelt out.

INSTRUCTIONS:

1. Vendor will get approval for the site plan from AERB for installation of the equipment.
2. Any electrical work required at the site for installation of machine is to be done by the vendor.

Hand piece sterilization system

Automatic maintenance and sterilization programme for cleaning & lubrication as well as sterilization of all types of hand pieces. Adaptors/ Converters for straight, contra, 2 hole and 4 hole hand pieces. Should be supplied with all accessories and consumables for initial running for 3 months.

PORTABLE FLASH AUTOCLAVE SPECIFICATIONS

- ❖ Autoclave should be operated on 121 degree C and 134 degree C for both wrapped and unwrapped tools.
- ❖ Should also have a provision of Flash Cycle for Rapid-Sterilization Liquid cycle.
- ❖ Should be provided with powerful vacuum pump to eject air pocket from the chamber at the beginning and end of the cycle.
- ❖ The chamber should be of Stainless Steel & ASME Certified.
- ❖ Warm up vacuum, sterilization & drying phases should be fully automatic.
- ❖ Safety features should be controlled by dynamic and electromagnetic.
- ❖ Door should not open unless the chamber is at atmospheric pressure.
- ❖ Autoclave should execute vacuum test, Bowie & Dick test.
- ❖ System should have LCD display which predicts all parameters and error messages.
- ❖ Should have built in digital printing device.
- ❖ Should be Equipped with separate water tanks for clean and contaminated water.
- ❖ System should display water levels, date and hour.

TECHNICAL SPECIFICATIONS:-

- ❖ Chamber Size: not less than 200 (Dia)*500(Depth) mm.
- ❖ Capacity: not less than 20 litres.
- ❖ Operating Voltage: 230 volts, 50/60 Hz.
- ❖ Power consumption is less than 2 Kw.
- ❖ The system complies with 93/42/EEC, 97/23/EEC/PrEN 13060, ASME code, EN 61010-1 FDA Approved. Also TUV, CE IQ, Net, ISO 9001-2000, UL & CE 012 certified.

MINOR OT

ANAESTHESIA MACHINE

1. Should have provision for delivery of oxygen and nitrous oxide with Pressure gauges.
2. Should have independent attachments for connecting central gas supply and pin indexed cylinders.
4. Should have analogue display of cylinder and pipeline gas pressures.
5. Should have provision to attach two cylinder for Oxygen and one for Nitrous Oxide.
6. Oxygen and Nitrous oxide should be linked to ensure a minimum of 25% oxygen delivery at all times to avoid delivery of hypoxic mixture. Lever based anti hypoxic device is not acceptable.
7. Should have depiction of O₂, NO₂. Twin Cascaded Flowmeters with mechanical hypoxic Guard and should preferably have minimum mandatory flow of 250 ml/min of Oxygen.
8. Should have back bar with ISO pin type to attach vaporisers easily.
9. Should have top shelf and a table top to keep drugs and equipments.
10. Machine should have ventilator with battery backup.
11. Castor wheels should be durable and moisture resistant as well as antistatic.
12. Unlockable oxygen flush to deliver oxygen flow of approximately 40-70 l/min.
13. Should have drawers with easy maneuverability.
14. Compliant with ultra-low flow anesthesia.
15. Colour coded cylinder and pipeline pressure gauges
16. Machine should have US FDA / European CE Certified.

B. Standard Circle Absorber System

1. Should have adjustable pressure limiting valve & breathing circuit pressure measuring device.
2. Should have bag / vent selecting valve integrated onto the absorber and should automatically turn on the ventilator when positioned to vent mode.
3. Should be suitable to use ultra low flow techniques.
4. Should have facility to attach oxygen sensor.
5. Should have fully autoclaveable CO₂ absorbent canisters and bellows.

C. Vaporiser

1. Temperature, pressure and flow compensated.
2. Should provide keyed filler based Halothane, Isoflurane, & Sevoflurane vaporisers.
3. Should be easy to mount and dismount from the back bar.
4. Should have ISO pin type (Selectatec) back bar mount.
5. Vaporiser should be maintenance free for ten years.

D. Ventilator

1. Integrated Ventilator
- 2 Volume Controlled Ventilation
3. Integrated Compact breathing system
4. Latex-free and autoclavable bellows
5. Compact autoclavable sodalime canister
6. LED Display
7. Audio Visual alarms
8. Pressure, Volume Monitoring
9. Airway Pressure Gauge
10. Monitor shelf
11. Quick changeover from mechanical to manual Ventilation
12. Compatible with open and semi-closed circuits
13. Should have tidal volume range from 20 ml to 1500 ml.
14. Should be able to set TV, RR and I:E ratio.

15. Ventilator should provide all user alarms.
16. Ventilator should provide Fresh gas compensation and Compliance compensation.
18. Should be supplied with necessary reusable and disposable breathing circuits.

TECHNICAL SPECIFICATIONS FOR ELECTRO-SURGICAL UNIT

1 Description of Function

1.1 Electro-surgical units are required to provide cutting and coagulation electrically during surgery

2 Operational Requirements

2.1 A user friendly, Microcontroller-based multi-programmable by the user, isolated Electro surgical generator, designed for all surgical procedure and with Spray mode for Surgery and essential facility of Vessel Sealing system (Comprehensive unit).

2.2 Technical Demonstration of the quoted model(s) is mandatory for technical evaluation and should be done as per the request & guidelines subscribed by the tenderer.

3 Technical Specifications

3.1 Ensure controlled, precise desiccation with less destruction of peripheral tissue

3.2 Permits two surgeons to operate from a single unit & accessories are activated only when keyed, should not get activated inadvertently

3.3 Low voltage coagulation

3.4 Must have three types of cut modes-Low, Pure& Blend

3.5 Must have four coagulation mode

3.6 Maximum output: 300 watts sinusoid 350-520 kHz.

3.7 Modes: Both Bipolar (minimum 1 port) and Mono-polar (minimum 2 ports), Mono-polar should have cutting, spray, desiccation and fulguration modes. Each monopolar should have separate output setting.

3.9 Under water facility is essential. Convection cooling system

4 System Configuration Accessories, spares and consumables (unit rate to be quoted for individual items)

4.1 Main Unit -01 No

4.2 Pencil Holder -01 No

4.3 Electrode Blades -10 Nos

4.4 Electrode Needle Fine -10

4.5 Electrode – Ball point – (3 different sizes- small, Medium & long) – 5 each

4.6 Pencil Switches with hand control -10 Nos disposable, 2 Nos reusable

4.7 2 Double Foot Control paddle

4.8 chuck handles - 10 Nos disposable, 2 Nos reusable

4.9 Reusable Patient plate with cable – 5 Nos. Disposable patient plate with cable – 100Nos.

4.10 Mobile trolley/cart on castors with brakes -01 Nos

4.11 Power cord of 5 meters length and plug with appropriate rating for standard Indian electrical sockets. -01 No

4.12 Vessel sealing Clamps (re – usable) for both open and Endoscopic surgical procedures – 4 Nos each (Price should be quoted separately)

5 Environmental factors

5.1 The unit shall be capable of being stored continuously in ambient temperature of 0 -50deg C and relative humidity of 15-90%

5.2 The unit shall be capable of operating continuously in ambient temperature of 10 -40deg C and

relative humidity of 15-90%

5.3 Shall meet IEC-60601-1-2:2001(Or Equivalent BIS) General Requirements of Safety for Electromagnetic Compatibility.

6 Power Supply

6.1 Power input to be 220 V AC 50 Hz.

6.2 Resettable over current breaker shall be fitted for protection

7 Standards, Safety and Training

7.1 Patient system should be guaranteed by Return Electrode Contact Quality Monitor System which should automatically switch off the unit together with audiovisual alarms in case of power supply disconnection of the plate in the event of wire break off or loose connection. If the plate is not installed underneath the patient or it has a crack in the system it should not work.

7.2 The supplier should be ISO certified CE European directive / US FDA marked for quality standards. Complete system with all accessories should be from same manufacturer.

7.3 Electrical safety conforms to standards for electrical safety IEC-60601-1 General Requirements

7.4 Certified to be compliant with IEC 60601-2-2 Medical Electrical Equipment Part 2-2: Particular requirements for the safety of High Frequency Surgical Equipments

8 Documentation

8.1 User Manual in English

8.2 Service manual in English

8.3 Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist. The job description of the hospital technician and company service engineer should be clearly spelt out.

8.4 Certificate of calibration and inspection from factory.

8.5 Compliance Report to be submitted in a tabulated and point wise manner clearly mentioning the page/ Para number of original catalogue/data sheet. Any point, if not substantiated with authenticated catalogue/manual, will not be considered.

8.6 List of Equipments available for providing calibration and routine maintenance support as per manufacturer documentation in service / technical manual.

8.7 List of important spare parts and accessories **with their part number and costing**

OPERATION TABLE HYDRAULIC

1. Description of Function

1.1 Hydraulic operating Tables are simple tables for performing surgical procedures and it works without electrical power. It should have basic functions of height, tilting and Trendelenberg positions.

2. Operational Requirements

2.1 OT Table is required for general surgery and should have X-Ray translucent tops.

3. Technical Specifications

1. Four section table top (head section, back section, seat section with perineal cut, divided/foldable/detachable leg section with integrated kidney bridge.

2. Table top should be constructed from a high-pressure laminate/carbon fibre permitting x-ray penetration and fluoroscopy

3. All table positioning, i.e., height, back section, lateral tilt, trendelenburg, and anti-trendelenburg, except foot and head section should be operated hydraulically

4. Should have a manual position selector

5. The casings on the frame and centre supporting column should be made of chromium, nickel steel (CrNi Steel min SS 304) with central locking facility and four anti static double castors.

6. Mattress should be radio lucent and suitable for fluoroscopy

7. Measurements :(all dimensions are approximate)

a. Minimum **600 ±50** mm Max. **1050mm or More** (without mattress)

b. Side tilt: 15 degrees to 20 degrees

c. Back section adjustment: 40 degrees (**Down**) to 70 degrees (**Up**).

d. Foot section adjustment: - 90 to 0 degree/detachable

e. Trendelenburg: 25 to 30 degree

f. Anti trendelenburg: 25 to 30 degree

g. Head section adjustment: -40 to -30 degree/detachable

h. Maximum width: 500-540 mm(without side rails)

i. Length: 1950 – 2200 mm

j. Weight bearing capacity 125-150 Kg.

4. System Configuration Accessories, spares and consumables

4.1 System as specified

4.2 Accessories should include

a. Padded arm rest with straps - pair with damps

b. Anaesthesia screen with clamps

c. Side supports: pair with clamps

d. Shoulder supports: pair with clamps

e. Knee crutches: pair with damps

f. X-ray cassette tray

g. Infusion rod with clamp

5. Standards, Safety and Training

5.1 Should be US – FDA/European CE approved product

5.2 Manufacturer should be ISO certified for quality standards.

6. Documentation

6.1 User/Technical/Maintenance manuals to be supplied in English.

6.2 Certificate of calibration and inspection.

6.3 List of Equipments available for providing calibration and routine Preventive Maintenance Support. as per manufacturer documentation in service/technical manual.

6.4 List of important spare parts and accessories with their part number and costing

6.5 List of installations preferably govt. Institutions/medical colleges with performance certifications (min. Three Nos)

6.6 Demonstration of quoted model before the appropriate committee is necessary. Should provide in house training to all users.

SPECIFICATIONS OF OPERATING THEATRE LIGHT: MOBILE

1.1	Mobile operating light is required for illuminating the operating field in an emergency environment and the system can be moved from place to place.
2.1	State of the art system with shadow less light
3.1	<ul style="list-style-type: none"> a. Mobile light on lockable castors b. Should be LED based microprocessor control technology c. Light output 1,00,000 Lux or more d. Colour temperature 4500K or better e. Colour Rendering Index (CRI) 95 % f. Sterilizable focusing handle g. Should withstand wide voltage fluctuation h. Should have intensity control from 40-100%
3.2	Emergency Power Unit having in-built CVT with automatic change over from Mains to Battery mode in the event of power failure to provide 60 minutes back up
7.1	Should be US FDA , CE,UL or BIS approved product
7.2	Manufacturer should be ISO certified for quality standards.
7.3	Electrical safety conforms to standards for electrical safety IEC 60601-1 General Requirements (or equivalent BIS Standard
7.4	Shall meet internationally recognised standard for Electro Magnetic Compatibility (EMC) for electromedical equipment: IEC-60601-1-2 :latest edition Or Equivalent BIS) or should comply with 89/366/EEC; EMC-directive as amended
7.5	Certified to be compliant with IEC 60601-2-41: Particular requirements for the safety of Operation Theatre Light or equivalent if applicable

Suture Set (4 Sets.)		
1.	METZENBAUM SCISSOR CURVED 150MM	1 No.
2.	METZENBAUM SCISSOR STRAIGHT 200MM	1 No.
3.	SPONGE HOLDING FORCEPS 160MM	1 No.
4.	HALSTED MOSQUITO ARTERY FORCEPS CURVED 125MM	6 Nos.
5.	STANDARD DISSECTING FORCEPS PLAIN 150MM	1 No.
6.	STANDARD DISSECTING FORCEPS TOOTHED 180MM	1 No.
7.	MAYO HEGAR NEEDLE HOLDER 180MM	2 Nos.
8.	HEATHS SUTURE REMOVAL SCISSOR ONE BLADE SERRATED 180MM	1 No.
9.	LARYNGOSCOPE WITH 3 BLADES ADULT	1 Set
10.	LARYNGOSCOPE WITH 2 BLADES PAEDIATRIC	1 Set
11.	BACKHAUS TOWEL CLIPS 4"	6 Nos.
12.	CHAETLE FORCEPS	1 No.
13.	SPONGE HOLDING FORCEPS 10"	1 No.
14.	SS Tray with lid –Length 300mm, Width 200mm, Height 50mm	1 No.
15.	Kidney Tray (Length 300mm, Width 70mm, Height 30mm)	1 No.
16.	Galliot's (10cms)	1 No.
17.	Scalpel Handle No.4	1 No.
18.	Sponge Holding Forceps –8"	1 No.
19.	Tooth Forceps 6"	1 No.
20.	Needle (1/2 Circl, Cutting) (Size 20mm, 30mm)	1 No.
21.	Scissor Mayos – 150mm	1 No.
22.	Needle Holder (Hegar's) -150mm	1 No.
23.	Artery Forceps (Mosquito)- 150mm	6 Nos.

Instruments should be of high quality stainless steel, corrosion free, reusable and rustfree.

Stitch Removal Set (4 Sets.)		
Stitch Removal/ Dressing Instruments		
(i)	Kidney Tray Small size: 8"	1 No.
(ii)	Forceps toothed dressing / dissecting 6"	1 No.
(iii)	Medium sized Artery Forcep 6" Curved	1 No.
(iv)	Suture Cutting Scissor	1 No.

Instruments should be of high quality stainless steel, corrosion free, reusable and rust free .

Minor General Surgical Instruments Set (4 Sets.)

Skeletal Traction Set (4 Sets.)		
1	Kirchner Extension Bow	2 Nos.
2	Wire Strainer	2 Nos.
3	Spanner	2 Nos.
4	Traction Hook	6 Nos.
5	K-wire 200 mm	20 Nos.
6	T – Handle, Steinmann’s Pin, Bohler’s Stirrup, Ropes	2 each
7	Thomas Splint Small, Medium, Large	1 each
8	Crammer Wired Medium	4 Nos.
9	Crammer Wired Large	4 Nos.
10	6. Bone Chisel / Osteotomy	2 each
11	Hammer	1 No.
12	Plier	1 No.
13	K – Wire 200mm	15 Nos.
14	Metzenbaum Dissecting Scissors Curved 8"	2 Nos.

Instrument required for Dressing Room (4 Sets.)		
1	Mosquito Artery Forceps 5"	6 Nos.
2	Toothed & Non Toothed Forceps 6'	1 each
3	Suture Cutting Forceps	1 No.
4	Sponge Holding Forceps 10'	1 No.
5	B.P. Handle No. 3 & 4	1 each
6	Surgical Scissor 7"	1 No.
7	S.S. Kidney Tray 6"	1 No.
8	S.S. Bowl 4"	2 Nos.
9	S.S. Tray with Lid 6" x 8"	1 No.
10	S.S. Sterilization Drum	1 No.
11	Needle Holder 6"	1 No.
12	Cheatele Forceps	1 No.
13	Cheatele Jar	1 No.

Instruments should be of high quality stainless steel, corrosion free, reusable and rustfree.

OPHTHALMOLOGY

SLIT LAMP

1. Should have illumination form top using tungsten filament bulb capable of giving illumination intensity up to 600,00Lux.
2. Magnification from 6.3Xto 40X having continous zoom
3. Slit length 0.2 to 8mm
4. Slit length 0.2 to 8mm
5. Eyepieces of 12.5X
6. Facility to tilt the slit image up to 20 deg. Should be available
7. Should have applanation tonometer
8. Should have the facility for stereoscopic examination of fundus where in Angle on stereoscopic observation can be reduced from 13 Deg. To 4.5 Deg. To have better view of eyes with small pupil or high myopia.
9. Facility for viewing into the microscopic at 20 deg to horizontal to enable the examiner to keep his head in fatigue free position.
10. Slit lamp should be mounted on spring balanced instrument stand

LENSOMETER / FOCIMETER

- For Accurate analysis of the power of any lens including multifocal, prisms, contact lens and centration.
- Specifications
- Light source: LED/ Halogen bulb
- Sphere measuring range: -25 +25D unit 0.25D
- Axis Measuring range 0-180° units 1°
- Prism Measuring range 0-6 unit 1 Prism diopter
- Scale reading : internal -/+ external
- Lens diameter: 24-90 mm
- Eyepiece Focusing: +7D to -10D
- Tilt angle: 0-90° , freely adjustable
- Power supply: AC 220 to 240 volts as well as battery operated
- Accessories: 1 Dust cover, 1 contact lens holder

Should have European CE or US FDA approved.

SECTION-VII

TECHNICAL SPECIFICATIONS GENERAL TECHNICAL SPECIFICATIONS

GENERAL POINTS:

1. Warranty:

- a) Two years Comprehensive Warranty as per Conditions of Contract of the TE document for complete equipment (including X ray tubes, Helium for MRI, Batteries for UPS, other vacuumatic parts wherever applicable) and Turnkey Work from the date of satisfactory installation, commissioning, trial run & handing over of equipment to Hospital/Institution/Medical College.
- b) 98% up time Warranty of complete equipment with extension of Warranty period by double the downtime period on 24 (hrs) X 7 (days) X 365 (days) basis.
- c) All software updates should be provided free of cost during Warranty period.

2. After Sales Service:

After sales service centre should be available nearby at the city of Hospital/Institution/Medical College on 24 (hrs) X 7 (days) X 365 (days) basis. Complaints should be attended properly, maximum within 8 hrs. The service should be provided directly by Tenderer/Indian Agent. Undertaking by the Principals that the spares for the equipment shall be available for at least 10 years from the date of supply.

3. Training:

On Site training to Doctors/ Technicians/ staff is to be provided by Principal/ Indian Agents (if they have the requisite know-how) for operation and maintenance of the equipment to the satisfaction of the consignee.

4. Annual Comprehensive Maintenance Contract (CMC) of subject equipment with Turnkey:

- a) The cost of Comprehensive Maintenance Contract (CMC) which includes preventive maintenance including testing & calibration as per technical/ service /operational manual of the manufacturer, labour and spares, after satisfactory completion of Warranty period may be quoted for next 5 years on yearly basis for complete equipment (including X ray tubes, Helium for MRI, Batteries for UPS, other vacuumatic parts wherever applicable) and Turnkey (if any). The supplier shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, but at least once in six months during the CMC period
- b) The cost of CMC may be quoted along with taxes applicable on the date of Tender Opening. The taxes to be paid extra, to be specifically stated. In the absence of any such stipulation the price will be taken inclusive of such taxes and no claim for the same will be entertained later.
- c) Cost of CMC will be added for Ranking/Evaluation purpose.
- d) The payment of CMC will be made on six monthly basis after satisfactory completion of said period, duly certified by end user on receipt of bank guarantee for 2.5 % of the cost of the equipment as per Section XV valid till 2 months after expiry of entire CMC period.
- e) There will be 98% uptime warranty during CMC period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend CMC period by double the downtime period.
- f) During CMC period, the supplier is required to visit at each consignee's site at least once in 6 months commencing from the date of the successful completion of warranty period for preventive maintenance of the goods.

- g) All software updates should be provided free of cost during CMC.
- h) Failure of the above [4. e) to 4. g)] by the supplier, may lead to the forfeiture of the Bank Guarantee for Annual CMC.
- i) The payment of CMC will be made as stipulated in GCC Clause 21.

Turnkey:

Turnkey is indicated in the technical specification of the respective items, wherever required. The Tenderer shall examine the existing site where the equipment is to be installed, in consultation with HOD of Hospital/Institution/Medical College concerned. Turnkey details of each Hospital/Institution/Medical College are given at the end of Technical Specification. The Tenderer to quote prices indicating break-up of prices of the Machine and Turnkey Job of each Hospital/Institution/Medical College. The Turnkey costs may be quoted in Indian Rupee will be added for Ranking Purpose.

The taxes to be paid extra, to be specifically stated. In the absence of any such stipulation the price will be taken inclusive of such duties and taxes and no claim for the same will be entertained later.

The Turnkey Work should completely comply with AERB requirement, if any.

Note 1: Tenderer's attention is drawn to GIT clause 18 and GIT sub-clause 11.1(c). The tenderer is to provide the required details, information, confirmations, etc. accordingly failing which it's tender is liable to be ignored.

Note 2: General: Bidders are requested to make sure that they should attach the list of equipments for carrying out routine and preventive maintenance wherever asked for and should make sure that Electrical Safety Analyzer / Tester for Medical equipments to periodically check the electrical safety aspects as per BIS Safety Standards IS-13540 which is also equivalent to IEC electrical safety standard IEC-60601 is a part of the equipments. If the Electrical Safety Analyzer/Tester is not available they should provide a commitment to get the equipments checked for electrical safety compliance with Electronic Regional Test Labs / Electronics Test and Development Centres across the country on every preventive maintenance call.

Note 3: OPTIONAL ITEMS: Deleted.

Section – VIII

Quality Control Requirements

(Proforma for equipment and quality control employed by the manufacturer(s))

Tender Reference No.

Date of opening

Time

Name and address of the Tenderer:

Note: All the following details shall relate to the manufacturer(s) for the goods quoted for.

- 01 Name of the manufacturer
 - a. full postal address
 - b. full address of the premises
 - c. telegraphic address
 - d. telex number
 - e. telephone number
 - f. fax number

- 02 Plant and machinery details
- 03 Manufacturing process details
- 04 Monthly (single shift) production capacity of goods quoted for
 - a. normal
 - b. maximum

- 05 Total annual turn-over (value in Rupees)
- 06 Quality control arrangement details
 - a. for incoming materials and bought-out components
 - b. for process control
 - c. for final product evaluation
- 07 Test certificate held
 - a. . type test
 - b. . BIS/ISO certification
 - c. . any other
- 08 Details of staff
 - a. technical
 - b. skilled
 - c. unskilled

Signature and seal of the Tenderer

Section – IX

Qualification Criteria

01. The Tenderer must be a Manufacturer. In case the manufacturer does not quote they shall give reasons for not quoting directly. They may authorise their agent as per proforma of Manufacturer authorization form as given in the tender enquiry document to quote and enter into a contractual obligation.
02. (a) The Manufacturer should have supplied and installed in last **Five** years from the date of Tender Opening, atleast 100% of the quoted quantity of the similar equipment meeting major specification parameters which is functioning satisfactorily, anywhere in India.
02. (b) The Tenderers quoting as authorized representative of the manufacturer meeting the above criteria 02 (a) should have supplied and installed in last **Five** years from the date of Tender Opening, atleast 100% of the quoted quantity of similar equipments which is functioning satisfactorily, any where in India of any manufacturer.

1. The tenderer shall give an affidavit as under:

“We hereby certify that if at any time, information furnished by us is proved to be false or incorrecr, we are liable for any action as deemed fit by the purchaser in addition to forfeiture of the earnest money.

2. In support of 2 (a) & 2 (b), the Tenderer shall furnish Performance statement in the enclosed Proforma ‘A’.

The manufacturer as well as the Tenderer/ Indian Agent shall furnish Satisfactory Performance Certificate in respect of above, duly translated in English and duly notarized in the country of origin, alongwith the tender.

3. The Tenderer shall furnish a brief write-up, packed with adequate data explaining and establishing his available capacity/capability (both technical and financial) to perform the Contract (if awarded) within the stipulated time period, after meeting all its current/present commitments. The Tenderer shall also furnish details of Equipment and Quality Control in the enclosed Section VIII.
4. Notwithstanding anything stated above, the Purchaser reserves the right to assess the Tenderer’s capability and capacity to perform the contract satisfactorily before deciding on award of Contract, should circumstances warrant such an assessment in the overall interest of the Purchaser.
5. The Purchaser reserves the right to ask for a free demonstration of the quoted equipment at a pre determined place acceptable to the purchaser for technical acceptability as per the tender specifications, before the opening of the Price Tender.

PROFORMA 'A'
PROFORMA FOR PERFORMANCE STATEMENT

(For the period of last five years)

Tender Reference No. : _____

Date of opening : _____

Time : _____

Name and address of the Tenderer : _____

Name and address of the manufacturer* : _____

Order placed by (full address of Purchaser/ Consignee)	Order number and date	Description and quantity of ordered goods and services	Value of order (Rs.)	Date of completion of Contract		Remarks indicating reasons for delay if any	Have the goods been functioning Satisfactorily (attach documentary proof)**
				As per contract	Actual		
1	2	3	4	5	6	7	8

“We hereby certify that if at any time, information furnished by us is proved to be false or incorrect, we are liable for any action as deemed fit by the purchaser in addition to forfeiture of the earnest money.

Signature and seal of the Tenderer

** The documentary proof will be a certificate from the consignee/end user with cross-reference of order no. and date in the certificate along with a notarized certification authenticating the correctness of the information furnished.

Section – X
TENDER FORM

Date _____

To

HSCC (India) Ltd, Plot No. 6-A, Block-E, Sector-1, NOIDA (U.P.) – 201 301

Ref. Your TE document No. _____ dated _____

We, the undersigned have examined the above mentioned TE document, including amendment/corrigendum No. _____, dated _____ (if any), the receipt of which is hereby confirmed. We now offer to supply and deliver _____ (Description of goods and services) in conformity with your above referred document for the sum, as shown in the price schedule(s), attached herewith and made part of this tender. If our tender is accepted, we undertake to supply the goods and perform the services as mentioned above, in accordance with the delivery schedule specified in the List of Requirements.

We further confirm that, if our tender is accepted, we shall provide you with a performance security of required amount in an acceptable form in terms of GCC clause 5, read with modification, if any, in Section - V – “Special Conditions of Contract”, for due performance of the contract.

We agree to keep our tender valid for acceptance as required in the GIT clause 20, read with modification, if any in Section - III – “Special Instructions to Tenderers” or for subsequently extended period, if any, agreed to by us. We also accordingly confirm to abide by this tender up to the aforesaid period and this tender may be accepted any time before the expiry of the aforesaid period. We further confirm that, until a formal contract is executed, this tender read with your written acceptance thereof within the aforesaid period shall constitute a binding contract between us.

We further understand that you are not bound to accept the lowest or any tender you may receive against your above-referred tender enquiry.

We confirm that we do not stand deregistered/banned/blacklisted by any Govt. Authorities.

We confirm that we fully agree to the terms and conditions specified in above mentioned TE document, including amendment/ corrigendum if any

(Signature with date)

(Name and designation) Duly authorised to sign tender for and on behalf of

SECTION – XI PRICE SCHEDULE

A) PRICE SCHEDULE FOR DOMESTIC GOODS OR GOODS OF FOREIGN ORIGIN LOCATED WITHIN INDIA

1 Schedule	2 Brief Description of Goods	3 Country of Origin	4 Quantity (Nos.)	5 Price per unit (Rs.)							6 Total Price (at Consignee Site) basis (Rs.)
				Ex - factory/ Ex -warehouse /Ex-showroom /Off - the shelf (a)	Excise Duty (if any) [%age & value] (b)	Sales Tax/ VAT(if any) [%age & value] (c)	Packing and Forwarding charges (d)	Inland Transportation, Insurance for a period including 3 months beyond date of delivery, loading/ unloading and Incidental costs till consignee's site (e)	Incidental Services (including Installation & Commissioning, Supervision, Demonstration and Training) at the Consignee's site (f)	Unit Price (at Consignee Site) basis (g) =a+b+c+d+e+f	
											4 x 5(g)

Total Tender price in Rupees: _____

In words: _____

Note: -

1. If there is a discrepancy between the unit price and total price THE UNIT PRICE shall prevail.
2. The charges for Annual CMC after warranty shall be quoted separately as per Section – XI – Price Schedule C

Name _____

Business Address _____

Place: _____

Signature of Tenderer _____

Date: _____

Seal of the Tenderer _____

SECTION – XI PRICE SCHEDULE

C) **PRICE SCHEDULE FOR ANNUAL COMPREHENSIVE MAINTENANCE CONTRACT AFTER WARRANTY PERIOD**

1	2	3	4					5	6
Schedule No.	BRIEF DESCRIPTION OF GOODS	QUANTITY (Nos.)	Annual Comprehensive Maintenance Contract Cost for Each Unit year wise*.					Total Annual Comprehensive Maintenance Contract Cost for each unit for 5 years (4a+4b+4c+4d+4e)	Annual Comprehensive Maintenance Contract Cost for 5 Years [3 x 5]
			1 st	2 nd	3 rd	4 th	5 th		
			a	b	c	d	e		

* After completion of Warranty period

NOTE:-

1. In case of discrepancy between unit price and total prices, THE UNIT PRICE shall prevail.
2. The cost of Comprehensive Maintenance Contract (CMC) which includes preventive maintenance including testing & calibration as per technical/ service /operational manual, labour and spares, after satisfactory completion of Warranty period may be quoted for next 5 years on yearly basis for complete equipment and Turnkey (if any).
3. The cost of CMC may be quoted along with taxes applicable on the date of Tender Opening. The taxes to be paid extra, to be specifically stated. In the absence of any such stipulation the price will be taken inclusive of such taxes and no claim for the same will be entertained later.
4. Cost of CMC will be added for Ranking/Evaluation purpose.
5. The payment of CMC will be made as per clause GCC clause 21.1 (D).
6. The uptime warranty will be 98 % on 24 (hrs) X 7 (days) X 365 (days) basis or as stated in Technical Specification of the TE document.
7. All software updates should be provided free of cost during CMC period.
8. The stipulations in Technical Specification will supersede above provisions
9. The supplier shall keep sufficient stock of spares required during Annual Comprehensive Maintenance Contract period. In case the spares are required to be imported, it would be the responsibility of the supplier to import and get them custom cleared and pay all necessary duties.

Place: _____
Date: _____

Name _____
Business Address _____
Signature of Tenderer _____
Seal of the Tenderer _____

**SECTION – XI PRICE SCHEDULE
D) PRICE SCHEDULE FOR TURNKEY**

Schedule No.	BRIEF TURNKEY DESCRIPTION OF GOODS	CONSIGNEE CODE	Turnkey price

Note: -

1. The cost of Turnkey as per Technical Specification (Section VII) may be quoted on lump sum along with taxes applicable on the date of Tender Opening. The taxes to be paid extra, to be specifically stated. In the absence of any such stipulation the price will be taken inclusive of such taxes and no claim for the same will be entertained later.
2. Cost of Turnkey will be added for Ranking/Evaluation purpose.
3. The payment of Turnkey will be made as per clause GCC clause 21.1 (c).
4. The stipulations in Technical Specification will supersede above provisions

Name _____

Business Address _____

Place: _____

Signature of Tenderer _____

Date: _____

Seal of the Tenderer _____

**SECTION – XII
QUESTIONNAIRE**

Fill up the Section XX – Check List for Tenderers and enclose with the Tender

1. The tenderer should furnish specific answers to all the questions/issues mentioned in the Checklist. In case a question/issue does not apply to a tenderer, the same should be answered with the remark “not applicable”.
2. Wherever necessary and applicable, the tenderer shall enclose certified copy as documentary proof/ evidence to substantiate the corresponding statement.
3. In case a tenderer furnishes a wrong or evasive answer against any of the question/issues mentioned in the Checklist, its tender will be liable to be ignored.

SECTION – XIII
BANK GUARANTEE FORM FOR EMD

Whereas _____ (hereinafter called the “Tenderer”) has submitted its quotation dated _____ for the supply of _____ (hereinafter called the “tender”) against the purchaser’s tender enquiry No. _____ Know all persons by these presents that we _____ of _____ (Hereinafter called the “Bank”) having our registered office at _____ are bound unto _____ (hereinafter called the “Purchaser) in the sum of _____ for which payment will and truly to be made to the said Purchaser, the Bank binds itself, its successors and assigns by these presents. Sealed with the Common Seal of the said Bank this _____ day of _____ 20____. The conditions of this obligation are:

- (1) If the Tenderer withdraws or amends, impairs or derogates from the tender in any respect within the period of validity of this tender.
- (2) If the Tenderer having been notified of the acceptance of his tender by the Purchaser during the period of its validity:-
 - a) fails or refuses to furnish the performance security for the due performance of the contract.
 - or
 - b) fails or refuses to accept/execute the contract.
 - or
 - c) if it comes to notice that the information/documents furnished in its tender is incorrect, false, misleading or forged

We undertake to pay the Purchaser up to the above amount upon receipt of its first written demand, without the Purchaser having to substantiate its demand, provided that in its demand the Purchaser will note that the amount claimed by it is due to it owing to the occurrence of one or both the two conditions, specifying the occurred condition(s).

This guarantee will remain in force for a period of forty-five days after the period of tender validity and any demand in respect thereof should reach the Bank not later than the above date.

(Signature of the authorised officer of the Bank)

Name and designation of the officer

Seal, name & address of the Bank and address of the Branch

SECTION – XIV
MANUFACTURER’S AUTHORISATION FORM

To

HSCC (India) Ltd
E-6(A), Sector-1, Noida

Dear Sirs,

Ref. Your TE document No _____, dated _____

We, _____ who are proven and reputable manufacturers of _____ (*name and description of the goods offered in the tender*) having factories at _____, hereby authorise Messrs _____ (*name and address of the agent*) to submit a tender, process the same further and enter into a contract with you against your requirement as contained in the above referred TE documents for the above goods manufactured by us.

We further confirm that no supplier or firm or individual other than Messrs. _____ (*name and address of the above agent*) is authorised to submit a tender, process the same further and enter into a contract with you against your requirement as contained in the above referred TE documents for the above goods manufactured by us.

We also hereby extend our full warranty, CMC as applicable as per clause 15 of the General Conditions of Contract, read with modification, if any, in the Special Conditions of Contract for the goods and services offered for supply by the above firm against this TE document.

Yours faithfully,

[Signature with date, name and designation]
for and on behalf of Messrs _____

[Name & address of the manufacturers]

- Note: 1. This letter of authorisation should be on the letter head of the manufacturing firm and should be signed by a person competent and having the power of attorney to legally bind the manufacturer.*
- 2. Original letter may be sent.*

SECTION – XV

BANK GUARANTEE FORM FOR PERFORMANCE SECURITY/ CMC SECURITY

**To
HSCC (India) Ltd
E-6(A), Sector-1,
Noida**

WHEREAS _____ (Name and address of the supplier) (Hereinafter called “the supplier”) has undertaken, in pursuance of contract no_____ dated _____ to supply (description of goods and services) (herein after called “the contract”).

AND WHEREAS it has been stipulated by you in the said contract that the supplier shall furnish you with a bank guarantee by a scheduled commercial bank recognised by you for the sum specified therein as security for compliance with its obligations in accordance with the contract;

AND WHEREAS we have agreed to give the supplier such a bank guarantee;

NOW THEREFORE we hereby affirm that we are guarantors and responsible to you, on behalf of the supplier, up to a total of. _____ (Amount of the guarantee in words and figures), and we undertake to pay you, upon your first written demand declaring the supplier to be in default under the contract and without cavil or argument, any sum or sums within the limits of (amount of guarantee) as aforesaid, without your needing to prove or to show grounds or reasons for your demand or the sum specified therein.

We hereby waive the necessity of your demanding the said debt from the supplier before presenting us with the demand.

We further agree that no change or addition to or other modification of the terms of the contract to be performed there under or of any of the contract documents which may be made between you and the supplier shall in any way release us from any liability under this guarantee and we hereby waive notice of any such change, addition or modification.

This guarantee shall be valid up to 30 (thirty) months from the date of Notification of Award i.e. up to ----- (indicate date)

.....
(Signature with date of the authorised officer of the Bank)

.....
Name and designation of the officer

.....

Seal, name & address of the Bank and address of the Branch

**SECTION – XVI
CONTRACT FORM - A**

CONTRACT FORM FOR SUPPLY, INSTALLATION, COMMISSIONING, HANDING OVER, TRIAL RUN, TRAINING OF OPERATORS & WARRANTY OF GOODS

(Address of the Purchaser's/Consignee's office issuing the contract)

Contract No _____ dated _____

This is in continuation to this office's Notification of Award No _____ dated _____

1. Name & address of the Supplier: _____
2. Purchaser's TE document No _____ dated _____ and subsequent Amendment No _____, dated _____ (if any), issued by the purchaser
3. Supplier's Tender No _____ dated _____ and subsequent communication(s) No _____ dated _____ (if any), exchanged between the supplier and the purchaser in connection with this tender.
4. In addition to this Contract Form, the following documents etc, which are included in the documents mentioned under paragraphs 2 and 3 above, shall also be deemed to form and be read and construed as integral part of this contract:

- (i) General Conditions of Contract;
- (ii) Special Conditions of Contract;
- (iii) List of Requirements;
- (iv) Technical Specifications;
- (v) Quality Control Requirements;
- (vi) Tender Form furnished by the supplier;
- (vii) Price Schedule(s) furnished by the supplier in its tender;
- (viii) Manufacturers' Authorisation Form (if applicable for this tender);
- (ix) Purchaser's Notification of Award

Note: The words and expressions used in this contract shall have the same meanings as are respectively assigned to them in the conditions of contract referred to above. Further, the definitions and abbreviations incorporated under clause 1 of Section II – 'General Instructions to Tenderers' of the Purchaser's TE document shall also apply to this contract.

5. Some terms, conditions, stipulations etc. out of the above-referred documents are reproduced below for ready reference:

- (i) Brief particulars of the goods and services which shall be supplied/ provided by the supplier are as under:

Schedule No.	Brief description of goods/services	Accounting unit	Quantity to be supplied	Unit Price	Total price	Terms of delivery

Any other additional services (if applicable) and cost thereof: _____

Total value (in figure) _____ (In words) _____

2. Delivery schedule
- (iii) Details of Performance Security
- (iv) Quality Control
 - (a) Mode(s), stage(s) and place(s) of conducting inspections and tests.
 - (b) Designation and address of purchaser's inspecting officer
- (v) Destination and despatch instructions
- (vi) Consignee, including port consignee, if any
 3. Warranty clause
 4. Payment terms
 5. Paying authority

**(Signature, name and address
of the Purchaser's/Consignee's authorised official)
For and on behalf of** _____

Received and accepted this contract

(Signature, name and address of the supplier's executive
duly authorised to sign on behalf of the supplier)

For and on behalf of _____

(Name and address of the supplier)

(Seal of the supplier)

Date: _____

Place: _____

SECTION – XVI
CONTRACT FORM – B
CONTRACT FORM FOR ANNUAL COMPREHENSIVE MAINTENANCE CONTRACT

Annual CM Contract No. _____ dated _____
 Between _____

(Address of Head of Hospital/Institute/Medical College)
 And _____

(Name & Address of the Supplier)

Ref: Contract No _____ dated _____ (Contract No. & date of Contract for supply, installation, commissioning, handing over, Trial run, Training of operators & warranty of goods)

In continuation to the above referred contract

6. The Contract of Annual Comprehensive Maintenance is hereby concluded as under:

-

1	2	3	4					5
Schedule No.	BRIEF DESCRIPTION OF GOODS	QUANTITY. (Nos.)	Annual Comprehensive Maintenance Contract Cost for Each Unit year wise*.					Total Annual Comprehensive Maintenance Contract Cost for 3 Years [5 x (4a+4b+4c+4d+4e)]
			1 st	2 nd	3 rd	4 rd	5 th	
			a	b	c	d	e	

Total value (in figure) _____ (In words) _____

- b) The CMC commence from the date of expiry of all obligations under Warranty i.e. from _____ (date of expiry of Warranty) and will expire on _____ (date of expiry of CMC)
- c) The cost of Annual Comprehensive Maintenance Contract (CMC) which includes preventive maintenance, labour and spares, after satisfactory completion of Warranty period may be quoted for next 5 years as contained in the above referred contract on yearly basis for complete equipment (including X ray tubes, Helium for MRI, Batteries for UPS, other vacuumatic parts, _____ & _____) and Turnkey (if any).
- d) There will be 98% uptime warranty during CMC period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend CMC period by double the downtime period.
- e) During CMC period, the supplier shall visit at each consignee's site for preventive maintenance including testing and calibration as per the manufacturer's service/ technical/ operational manual. The supplier shall visit each consignee site as recommended in the manufacturer's manual, but at least once in 6 months commencing from the date of the successful completion of warranty period for preventive maintenance of the goods.
- f) All software updates should be provided free of cost during CMC.

- g) The bank guarantee valid till _____ [(fill the date) 2 months after expiry of entire CMC period] for an amount of Rs. _____ [(fill amount) equivalent to 2.5 % of the cost of the equipment as per contract] shall be furnished in the prescribed format given in Section XV of the TE document, along with the signed copy of Annual CMC within a period of 21 (twenty one) days of issue of Annual CMC failing which the proceeds of Performance Security shall be payable to the Purchaser/Consignee.
- h) If there is any lapse in the performance of the CMC as per contract, the proceeds Annual CMC bank guarantee for an amount of Rs. _____ (equivalent to 2.5 % of the cost of the equipment as per contract) shall be payable to the Consignee.
- i) **Payment terms:** The payment of Annual CMC will be made against the bills raised to the consignee by the supplier on six monthly basis after satisfactory completion of said period, duly certified by the HOD concerned. The payment will be made in Indian Rupees.
- j) **Paying authority:** _____ (name of the consignee i.e. Hospital/ Institute /Medical College's authorised official)

**(Signature, name and address
of Hospital AIIMS authorised official)**
For and on behalf of _____

Received and accepted this contract

(Signature, name and address of the supplier's executive
duly authorised to sign on behalf of the supplier)

For and on behalf of _____

(Name and address of the supplier)

(Seal of the supplier)

Date: _____

Place: _____

SECTION – XVII
CONSIGNEE RECEIPT CERTIFICATE
(To be given by consignee’s authorized representative)

The following store (s) has/have been received in good condition:

- 1) Contract No. & date : _____
- 2) Supplier’s Name : _____
- 3) Consignee’s Name & Address with
telephone No. & Fax No. : _____
- 4) Name of the item supplied : _____
- 5) Quantity Supplied : _____
- 6) Date of Receipt by the Consignee : _____
- 7) Name and designation of Authorized
Representative of Consignee : _____
- 8) Signature of Authorized
Representative of Consignee with
date : _____
- 9) Seal of the Consignee : _____

SECTION – XVIII
Proforma of Final Acceptance Certificate by the Consignee

No _____

Date _____

To

M/s _____

Subject: Certificate of commissioning of equipment/plant.

This is to certify that the equipment(s)/plant(s) as detailed below has/have been received in good conditions along with all the standard and special accessories and a set of spares (subject to remarks in Para no.02) in accordance with the contract/technical specifications. The same has been installed and commissioned.

- (a) Contract No _____ dated _____
- (b) Description of the equipment(s)/plants: _____
- (c) Equipment(s)/ plant(s) nos.: _____
- (d) Quantity: _____
- (e) Bill of Loading/Air Way Bill/Railway Receipt/ Goods Consignment Note no _____ dated _____
- (f) Name of the vessel/Transporters: _____
- (g) Name of the Consignee: _____
- (h) Date of commissioning and proving test: _____

Details of accessories/spares not yet supplied and recoveries to be made on that account.

Sl. No.	Description of Item	Quantity	Amount to be recovered No.
---------	---------------------	----------	----------------------------

The proving test has been done to our entire satisfaction and operators have been trained to operate the equipment(s)/plant(s).

The supplier has fulfilled its contractual obligations satisfactorily ## or

The supplier has failed to fulfil its contractual obligations with regard to the following:

He has not adhered to the time schedule specified in the contract in dispatching the documents/drawings pursuant to 'Technical Specifications'.

He has not supervised the commissioning of the equipment(s)/plant(s) in time, i.e. within the period specified in the contract from date of intimation by the Purchaser/Consignee in respect of the installation of the equipment(s)/plant(s).

The supplier as specified in the contract has not done training of personnel.

The extent of delay for each of the activities to be performed by the supplier in terms of the contract is

The amount of recovery on account of non-supply of accessories and spares is given under Para no.02.

The amount of recovery on account of failure of the supplier to meet his contractual obligations is _____ (here indicate the amount).

Signature
Name
Designation with stamp

Explanatory notes for filling up the certificate:

- i) He has adhered to the time schedule specified in the contract in dispatching the documents/drawings pursuant to 'Technical Specification'.**
- ii) He has supervised the commissioning of the equipment(s)/plant(s) in time, i.e. within the time specified in the contract from date of intimation by the Purchaser/Consignee in respect of the installation of the equipment(s)/plant(s).**
- iii) Training of personnel has been done by the supplier as specified in the contract.**
- iv) In the event of documents/drawings having not been supplied or installation and commissioning of the equipment(s)/plant(s) having been delayed on account of the supplier, the extent of delay should always be mentioned in clear terms.**

**SECTION – XIX
ANNEXURES**

Annexure 1

Deleted being domestic tender

SECTION – XX**CHECKLIST**

Name of Tenderer:

Name of Manufacturer:

Sl No.	Activity	Yes/ No/ NA	Page No. in the TE document	Remarks
1. a.	Have you enclosed EMD of required amount for the quoted schedules?			
b.	In case EMD is furnished in the form of Bank Guarantee, has it been furnished as per Section XIII?			
c.	In case Bank Guarantee is furnished, have you kept its validity of 165 days from Techno Commercial Tender Opening date as per clause 19 of GIT?			
2. a.	Have you enclosed duly filled Tender Form as per format in Section X?			
b.	Have you enclosed Power of Attorney in favour of the signatory?			
3.	Are you a SSI unit, if yes have you enclosed certificate of registration issued by Directorate of Industries/NSIC			
4. a.	Have you enclosed clause-by-clause technical compliance statement for the quoted goods vis-à-vis the Technical specifications?			
b.	In case of Technical deviations in the compliance statement, have you identified and marked the deviations?			
5. a.	Have you submitted satisfactory performance certificate as per the Proforma for performance statement in Sec. IX of TE document in respect of all orders?			

Sl No.	Activity	Yes/ No/ NA	Page No. in the TE document	Remarks
b.	Have you submitted copy of the order(s) and end user certificate?			
6.	Have you submitted manufacturer's authorization as per Section XIV?			
7.	Have you submitted prices of goods, turnkey (if any), CMC etc. in the Price Schedule as per Section XI?			
8.	Have you kept validity of 120 days from the Techno Commercial Tender Opening date as per the TE document?			
9. a.	In case of Indian Tenderer, have you furnished PAN No. as allotted by the Income Tax Department of Government of India?			
b.	In case of Foreign Tenderer, have you furnished PAN No. of your Indian Agent as allotted by the Income Tax Department of Government of India?			
10.	Have you intimated the name and full address of your Banker (s) along with your Account Number			
11.	Have you fully accepted payment terms as per TE document?			
12.	Have you fully accepted delivery period as per TE document?			
13.	Have you submitted the certificate of incorporation?			
14.	Have you accepted the warranty as per TE document?			
15.	Have you accepted terms and conditions of TE document?			

Sl No.	Activity	Yes/ No/ NA	Page No. in the TE document	Remarks
16.	Have you furnished documents establishing your eligibility & qualification criteria as per TE documents?			
17.	Have you furnished Annual Report (Balance Sheet and Profit & Loss Account) for last three years prior to the date of Tender opening?			

N.B.

1. All pages of the Tender should be page numbered and indexed.
2. The Tenderer may go through the checklist and ensure that all the documents/confirmations listed above are enclosed in the tender and no column is left blank. If any column is not applicable, it may be filled up as NA.
3. It is the responsibility of tendered to go through the TE document to ensure furnishing all required documents in addition to above, if any.

(Signature with date)

**(Full name, designation & address of the person duly authorised sign on behalf of the Tenderer)
For and on behalf of**

(Name, address and stamp of the tendering firm)

Section – XXI
Consignee List

Consignee Code	Medical Institutions	Contact Address.	AirPort	Sea Port
	All India Institute of Medical Science, Raebareli	All India Institute of Medical Science, Raebareli		

NB: The consignee will ensure timely issue of NMIC, Octroi Exemption Certificates, Road Permits & Entry Tax Exemption Certificates, wherever applicable, to the suppliers.