

AMENDMENT – X

Dated: - 06.06.2016

Subject: Amendment to the tender Enquiry Document**Ref: Tender Enquiry No.:** HSCC/KCGMC/ Medical Equipment / 2015-06/03 Dated 03.03.2016

With reference to the aforesaid IFB, Bidders are requested to note the following amendment:

Bid sale, submission and opening date for the items 1,2,8,10,11,12,14 has been extended as per details given in Table -1:

Table -1

Sl. No.	Description	Revised Schedule
i.	Sale date of the tender	13.06.2016, 2.00 P.M.
ii.	Closing date & time for receipt of tender	13.06.2016, 2.30 P.M.
iii.	Time and date of Opening of Techno – Commercial Tenders	13.06.2016, 3.00 P.M.

Bidders are also requested to note the following technical Amendment:

Item No. 2**Name of Item: Operation Theatre Table – Major OT**

S. No.	Specification as per tender	Amended as
2	It should be 5 section hydraulic (Oil free / Oil Leakage Proof - Spill free) operation table with split leg section.	It should be 5 Section Hydraulic (Oil Leakage proof - Spill free) operation table with split Leg Section.
12	The maximum permissible patient carrying weight to be min. 250 Kg or above all over the table area.	The maximum permissible patient carrying weight to be 250 KG or above in all positions.
	The following essential items must be quoted:	
1	All additional essential attachments for Orthopaedics Operating Table – 2 Nos. complete sets for making it fully functional as Orthopaedics Operation Table.	Additional attachments for Orthopaedics Operating Table – 2 Nos. complete sets for making it fully functional as Orthopaedics Operation Table should be quoted as per following: 1. Extension device 2 Extension bar supports. 2 Foot plates for adults. 2 Foot plates for child. 2 Traction boots for adults. 2 Traction boots for Child 1 Countertraction post for Dorsal position. Pad.---1 No. 2.Mounting fixture. ---1 No. Mounting arbor.---1 No. 3.Transport trolley---1 No. 4.Countertraction post for lateral position ---1 No. 5.Kirschner wire device--1 No. 6.Traction device for tibial fractures ---1 No. 7.Upper arm positioning device ---1 No. 8.Countertraction post for upper arm---1 No. 9.Weinberger hand traction device---1 No. 10.Upper arm positioning plate---1 No.
2	All additional essential attachments for Urology Operating Table – 1Nos. complete set for making it fully functional as Urology Operation Table.	Additional attachments for Urology Operating Table – 1Nos. complete sets for making it fully functional as Urology Operation Table should be quoted as per following: Swivel - mounted rinsing basin with holder - 1 No. TUR sieve - 1 No. Urological adaptor - 1 No. Elbow rests (Pair) - 1 No.

S. No.	Specification as per tender	Amended as
3	All additional essential attachments for ENT Operating Table – 1Nos. complete set for making it fully functional as ENT Operation Table.	Additional attachments for ENT Operating Table – 1Nos. complete sets for making it fully functional as ENT Operation Table should be quoted as per following: Shaped Head Rest - 1 No Universal adaptor - 1 No. Shoulder Plate - 1 No Gel Heading - 1 No.
4	All additional essential attachments for Endoscopy Procedure Operating Table – 1Nis. complete set for making it fully functional as Endoscopy Procedure Operation Table.	No change
5	All additional essential attachments for General Surgery Operating Table – 3 Nos. complete set for making it fully functional as General Surgery Operating Table out of which one no. shall be used for Dental Surgery.	No change

Item No. 11

Name of Item: Digital Flat Panel Radiography System

S. No.	Specification as per tender	Amended as
	Out of X – ray tube, Generator & Flat Panel detector at least 2 should be from the main manufacturer. In order to ensure the quality of the equipment.	Deleted.
6.	X-Ray Tube: 6.1 One number ceiling suspended Dual focus X ray tube with rotating anode with 9000 rpm or more.	6. X-Ray Tube: 6.1 One number ceiling suspended/ floor mounted Dual focus X ray tube with rotating anode with 9000 rpm or more.

Item No. 14

Name of Item: MRI (1.5 Tesla)

	Description of Function:	Amendment by HSCC/ KCGMC
8. COIL SYSTEM C.	Neuro – Vascular coil with 20 or more channels or Head / Neck coil combined, capable of high resolution neuro – vascular imaging.	Neuro – Vascular coil with 16 or more channels or Head / Neck coil combined, capable of high resolution neuro – vascular imaging.
8. COIL SYSTEM H.	Bilateral Breast Coil with at least 16 channels. The best available coil with vendor should be supplied.	Bilateral Breast Coil with at least 8 channels. The best available coil with vendor should be supplied.

All other terms and conditions of the tender enquiry document shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/ CPPP website for corrigendum /amendments etc. if any, as these will be notified on these portals only. No separate advertisement will published in the news papers in this regard.

s/d
CGM, HSCC India Limited
For and on behalf of DGMER, Panchkula