

Amendment-V**Ref.: IFB No. HSCC/SJH/Medical Equipment/2015 Dated 10.04.2015.**

Sub.: Procurement of Medical Equipment for New Emergency Block & Super-Specialty Block at Safderjung Hospital, New Delhi – Pre-bid Meeting held on 30.4.2015 regarding (Microbiology Eqpt.).

To be read as:**Item No. 22 Automatic Tissue Processor**

| S. No. | Tender Specifications | Amended Specifications |
|---------------|---|--|
| 1 | Point No. 2 Bench top unit movable on roller/castors provided with vacuum function and fume control. | Point No. 2 Bench top unit movable on roller/castors provided with vacuum/spiral agitation function and fume control. |

Item No. 23 Rotary Microtome with 50 blades

| S. No. | Tender Specifications | Amended Specifications |
|---------------|---|---|
| 1 | Point No. 3 (3.1) Specimen advance 1 to 1000 μm | Point No. 3 (3.1) Specimen advance min. 1 to 3000 μm |
| 2 | 3.5 Disposable blade holder for high & low profile blade. | 3.5 Disposable blade holder common or for both high & low profile blades. |

Item No. 32 Automated Slide Stainers

| S. No. | Tender Specifications | Amended Specifications |
|---------------|--|---|
| 1 | Point No. 2 (2.1) Should be programmable for routine H & E & other special stains with facility for imunohistochemical stains & memory of various staining procedures | Point No. 2 (2.1) Should be programmable for routine H & E & other special stains with memory of various staining procedures |
| 2 | 3.5 Can be connected with any make automatic cover-slipper. | 3.5 Can be compatible with any make automatic cover-slipper. |

Item No. 33 Tissue Embedding Station

| S. No. | Tender Specifications | Amended Specifications |
|---------------|---|---|
| 1 | Point No. 4 (4.2) Large field Magnifying lens with cold light source | Point No. 4 (4.2) Large field Magnifying lens with built-in white illumination |
| 2 | 4.5 Halogen Bulb – 12 Nos. | Deleted |

Contd...2

Item No. 34 GROSSING STATION WITH FUME HOOD

| S. No. | Tender Specifications | Amended Specifications |
|---------------|--|--|
| 1 | Point No. 5 Foot operated faucet control for hot and cold water. | Deleted |
| 2 | Point No. 12 Adjustable shelves. | Deleted |
| 3 | Point No. 13 Digital platform scale, 2 kg capacity | Deleted |
| 4 | Point No. 20 Dictation equipment with foot control or voice operation. | Deleted |
| 5 | Point No. 21 Camera mount facilities for use of a digital or 35mm SLR camera. | Point No. 21 Camera mount facilities for use of a digital camera. |
| 6 | Point No. 22 Eyewash assembly. | Deleted |

Item No. 40 Laser (Clinical CE/IVD) Flow Cytometer

| S. No. | Tender Specifications | Amended Specifications |
|---------------|--|---|
| 1 | Point No. 1 Should have simultaneous minimum 06 fluorescent (6 color) parameters analysis plus forward & side scatter. For each parameter the flow cytometer should be capable of measuring area, height and width. System should have PMTs on both forward & side scattered. | Point No. 1 Should have simultaneous minimum 06 fluorescent (6 color) parameters analysis plus forward & side scatter. For each parameter the flow cytometer should be capable of measuring area, height and width. System should have PMT and/or Photo Diode for both forward & for side scatter. |
| 2 | 16. Clinical Antibodies for 100 Tests with below panel :- 17. Fluorochrome labelled Antibodies for Acute leukemias, chronic lymphoproliferative disorders, Multiple Myeloma should be available with the company and compatible with the machine. | Deleted Same |

The date of submission of bids for Item no. 22, 23, 32, 33, 34 & 40 is 25.5.2015.

Amendment to be issued will be uploaded on websites www.tenderwizard.com/HSCC & www.hsccltd.com.

All other tender terms and conditions remain unchanged.

**Medical Superintendent
Safderjung Hospital & VMMC,
New Delhi.**