

**AMENDMENT – V**

Dated: - 25.05.2017

**Subject: Amendment to the tender Enquiry Document.**

Tender Ref: HSCC/KCGMC/Medical Equipment/2015-06/13 dated 11.04.17

Bidders are requested to note the following technical Amendments:

**Item No. 4 Phaco Emulsification System**

Point No	Tender Specification	Amended as
14	Should have a visco mode to remove viscoelastic from the eye. Should have aspiration flow rate capabilities up to 100 CC / min in both linear & fixed modes through all phaco/IA /VIT sub modes. Should have peristaltic system and the facility to use vacuum level upto 650 mm hg in both linear and fixed modes through all phaco/IA/ VIT submodes.	Should have a visco mode to remove viscoelastic from the eye. Should have aspiration flow rate capabilities from 50 CC / min and above in both linear & fixed modes through all phaco/IA /VIT sub modes. Should have peristaltic system and the facility to use vacuum level upto 650 mm hg in both linear and fixed modes through all phaco/IA/ VIT submodes.
15	Should have ability to regulate irrigating pressure by height of I/V pole and Continuous irrigation controlled by foot pedal and on touch panel.	Should have ability to regulate irrigating pressure by height of I/V pole and Continuous irrigation controlled by foot pedal.

**Item No. 10 OPG Machine**

Technical Specification	Amended as
A Partition with Lead lining of 2 mm thickness Lead, to be constructed. This will make an OPG chamber of dimension 7 feet x 7 feet. A Door of 3 feet width & 7 feet height will be available for patient entry in OPG chamber. The door will have 2 mm Lead barrier in it. This partition wall will have lead glass of 2 feet x 2 feet. All painting and finishing , civil, electrical HVAC work inside the room will be the responsibility of the vendor.	3 walls need to be Lead lined with 2mm thickness of Lead upto 7 feet height & same to be finished with paint. This should meet AERB guidelines. A Partition with Lead lining of 2 mm thickness Lead, to be constructed. This will make an OPG chamber of dimension 7 feet x 7 feet. A Door of 3 feet width & 7 feet height will be available for patient entry in OPG chamber. The door will have 2 mm Lead barrier in it. This partition wall will have lead glass of 2 feet x 2 feet. All painting and finishing , civil, electrical HVAC work inside the room will be the responsibility of the vendor.

**Bid sale, submission and opening date for item no. 4,6,10 has been extended as per details given in Table -1:**

**Table -1**

Sl. No.	Description	Revised Schedule
i.	Sale date of the tender	05.06.17, 3.00 P.M.
ii.	Closing date & time for receipt of tender	05.06.17, 3.30 P.M.
iii.	Time and date of Opening of Tenders	05.06.17, 4.00 P.M.

s/d  
**CGM, HSCC India Limited**  
**For and on behalf of DGMR, Panchkula**