

HSCC I Ltd.

AMENDMENT – VI

Dated: - 21.03.2017

Subject: Amendment to the tender Enquiry Document.

Tender Ref: HSCC/KCGMC/Medical Equipment/2015-06/12 dated 03.02.17

KCGMC has revised the specification of Item no.2 - Laprocat for Tubal Ligation as per the following:

Technical specifications for LAPROCATOR		
Sl.No	Specifications	Units
1	1) Forward-Oblique Tele- scope 30°, enlarged 1 view, diameter 6.5 mm, length 35 cm, autoclavable, fiber optic light transmission incorporated, color code: red	1
	2) Trocar, size 7mm, color code: yellow, consisting of 1 Trocar only with pyramidal tip Cannula without valve, with insufflation stop-cock, length 10.5cm ,Automatic valve.	2
	3) Ring Applicator for use with Operating Laparoscope	1
	4) Mounting Cone.	2
	5) Pusher.	2
	6) VERESS Pneumoperitoneum Needle with spring-loaded blunt stylet, LUER-lock, length 10 cm.	1
	7) VERESS Pneumoperitoneum Needle with spring loaded blunt stylet LUER-lock, length 15 cm.	1
	8) Electronic Endoflator Set, SCB operating voltage: 100- 1 240 VAC, 50/60 Hz consisting of Electronic Endoflator, with integrated SCB Module 400A Power cord , Silicone tube, sterilizable Universal wrench SCB Connecting cable Sterile filter, Package of 10 pieces.	1
	9) High Pressure Hose. American connection - pin-index 1 connection length 55 cm.	1
	10) Universal Wrench.	1
	11) Cold Light Fountain LED with one LED lamp and one light outlet Power Supply: 230 VAC, 50/60 Hz Consisting of: LED , Mains Cord.	1
	12) Fiber Optic Light Cable, size: 4.8 mm, length 180 cm.	1
	13) Formaline Chamber (for Sterilization of Laparoscopes) of appropriate size for sterilization of Laparoscopes.	1
	14) Tray for Sterilization of Laparoscopes of appropriate size for Sterilization of Laparoscopes	2

HSCC I Ltd.

Bid sale, submission and opening date for all items has been extended as per details given in Table -1:

Table -1

Sl. No.	Description	Revised Schedule
i.	Sale date of the tender	28.03.17, 3.00 P.M.
ii.	Closing date & time for receipt of tender	28.03.17, 3.30 P.M.
iii.	Time and date of Opening of Tenders	28.03.17, 4.00 P.M.

All other terms and conditions of the tender enquiry document shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/ CPPP website for corrigendum /amendments etc. if any, as these will be notified on these portals only. No separate advertisement will be published in the news papers in this regard.

s/d
CGM, HSCC India Limited
For and on behalf of DGMER, Panchkula