

AMENDMENT – XIII

Dated: - 31.10.2016

Subject: Amendment to the tender Enquiry Document.

Ref: Tender Enquiry No.: HSCC/KCGMC/Medical Equipment/2015-06/06 Dated 26.07.16

Bidders are requested to note the following technical amendment for the item no. 13 Temporal Bone Lab:

Specification of Microscope:

1. Integrated / Clamped on main temporal bone work station
2. 3 Step magnification 5x,10x & 20x
3. Manual fine focussing
4. Working Distance F= 200mm objective lens
5. Inter pupillary distance adjustment 55mm to 75mm
6. Light source : Twin halogen reflector lamp
7. Power Supply : 220v -240v

Specification of Suction Machine

1. Integrated in the main unit
2. High vacuum pressure
3. Noiseless
4. Automatic on/off switchfor vacuum function
5. One set(04of each) of suction tips of different caliber
7. Liquid container with two bottles of minimum one litre capacity.

Specification of hydraulic chair

1. Made of chrome plated steel
2. Adjustable height
3. Revolving
4. Comfortable seating
5. Upholstery of the seat should be fully synthetic leather and easy to clean

Image of Hydraulic Chair



Bid sale, submission and opening date for the item no. 13 (Temporal Bone Lab) has been extended as per details given in Table -1:

Table -1

Sl. No.	Description	Revised Schedule
i.	Sale date of the tender	07.11.2016, 3.00 P.M.
ii.	Closing date & time for receipt of tender	07.11.2016, 3.30 P.M.
iii.	Time and date of Opening of Techno – Commercial Tenders	07.11.2016, 4.00 P.M.

All other terms and conditions of the tender enquiry document shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/ CPPP website for corrigendum /amendments etc. if any, as these will be notified on these portals only. No separate advertisement will published in the news papers in this regard.

CGM, HSCC India Limited

**For and on behalf of DGMER,
Panchkula**