

**AMENDMENT – XIV**

**Dated: - 29.06.2016**

**Subject:** Amendment to the tender Enquiry Document.

**Ref: Tender Enquiry No.:** HSCC/KCGMC/ Medical Equipment / 2015-06/03 Dated 03.03.2016

With reference to the aforesaid IFB, Bidders are requested to note the following technical amendment in item no 12:

**CT Scan Machine**

<b>S. No.</b>	<b>Technical specification</b>	<b>Amended by KCGMC/ HSCC</b>
1	C) X-ray Tube • Anode Heat Storage Capacity of at least 7.0 MHU or more direct cooling tube	C) X-ray Tube • Anode Heat Storage Capacity of at least <b>8.0 MHU</b> or more direct cooling tube
2	B) X-ray Generator • High Frequency type. • Power output: 70 kW or higher • mA Range: 20-600 mA (With incremental steps of 10 mA) • KV Range: 80-110 or more	B) X-ray Generator • High Frequency type. • Power output: <b>80 kW</b> or higher • mA Range: 20-600 mA (With incremental steps of 10 mA) • KV Range: 80-110 or more
3	Scan Time should be 0.35 sec or less for full 360 degree rotation	Scan Time should be 0.4 sec or less for full 360 degree rotation

**All other terms and conditions of the tender enquiry document shall remain unchanged.**

**Prospective bidders are advised to regularly visit HSCC website/ CPPP website for corrigendum /amendments etc. if any, as these will be notified on these portals only. No separate advertisement will published in the news papers in this regard.**

s/d  
CGM, HSCC India Limited  
For and on behalf of DGMER, Panchkula