

HSCC/Saf./SSB/Hosp. Fur/181/2016

Dated: 18.05.2016

All Bidders

Amendment-1

Subject: **Tender for** Package-I : Supply, installation, testing & commissioning of Hospital Beds for Super Specialty Block at Safdarjung Hospital & VMMC New Delhi

Dear Sir,

This has reference to the pre bid meeting held on 16.05.2016 for the captioned subject works,

The following AMMENDMENT-1/CLARIFICATION-1 may be noted, which shall be treated as a part of tender **to be submitted online duly signed & stamp along with tender :**

1. Full Fowler With Mattress

“Bed should have bottom frame of ERW rectangular /oval tube of section 30 mm X 60 mm with thickness of 1.6 mm and ERW rectangular/round tube of section 35 mm and thickness 3 mm Bed should have top frame of ERW rectangular/oval tube of section 30 mm X 60 mm with thickness of 1.6 mm and ERW rectangular /Round tube with section 35 mm and thickness 3 mm.”

Other all terms & conditions of tender shall remain unchanged.

Prospective bidders are advised to regularly scan through HSCC e-tender portal <http://www.tenderwizard.com/HSCC> as corrigendum/amendments etc, if any, will be notified on this e- tender portal only and separate advertisement will not be made for this.

- sd-
(A.K.Sinha)
Project Coordinator