All Bidders

Amendment-1

Subject: Tender for Package 2: Supply, installation, testing & commissioning of ICU Beds for Super Specialty Block at Safdarjung Hospital & VMMC New Delhi

Dear Sir.

This has reference to the pre bid meeting held on 16.05.2016 for the captioned subject works,

The following AMMENDMENT-1/CLARIFICATION-1 may be noted, which shall be treated as a part of tender to be submitted online duly signed & stamp along with tender:

1. ICU BED with Mattress

- 1. There should be 4 nos of bumpers given at the four corners made up of neoprene/rubber bumpers with excellent shock absorbing property.
- 2. The ratchet should be adjustable in 7-8 different positions.
- 3. The unit should have the total load bearing capacity of 250-300kg with patient load bearing capacity of 135-200 kg.

Other all terms & conditions of tender shall remain unchanged.

Prospective bidders are advised to regularly scan through HSCC e-tender portal http://www.tenderwizard.com/HSCC as corrigendum/amendments etc, if any, will be notified on this e- tender portal only and separate advertisement will not be made for this.

- sd-(A.K.Sinha) Project Coordinator

Dated: 18.05.2016