

HSCC (India) Ltd

AMENDMENT – XXIV

Dated: - 06.1.2017

Subject: Amendment to the tender Enquiry Document.

Ref: Tender Enquiry No.: HSCC/KCGMC/Medical Equipment/2015-06/05

Dated 09/07/2016

Bidders are requested to note the following technical amendment:

Sl. No. - 8 FESS Instrument

Tender Specification	Amended as
Full HD Digital Camera	Full HD Digital Camera Additional requirement apart from exiting: Camera Head: The minimum functions like magnification, filters and intensity of light source have been controlled from the camera head in addition to the light source. Camera Head should be autoclavable.
Endoscope <ul style="list-style-type: none">45 degree, 2.7mm, 18cm straight forward telescope, autoclavable	Endoscope 45/30 degree, 2.7mm, 14-18cm straight forward telescope, autoclavable

Bid sale, submission and opening date for item no.8 has been extended as per details given in Table -1:

Table -1

Sl. No.	Description	Revised Schedule
i.	Sale date of the tender	13.01.2017, 3.00 P.M.
ii.	Closing date & time for receipt of tender	13.01.2017, 3.30 P.M.
iii.	Time and date of Opening of Tenders	13.01.2017, 4.00 P.M.

All other terms and conditions of the tender enquiry document shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/ CPPP website for corrigendum /amendments etc. if any, as these will be notified on these portals only. No separate advertisement will published in the news papers in this regard.

CGM, HSCC India Limited
For and on behalf of DGMER, Panchkula