

**Amendment no II Dated 11.05.2017**  
**IFB no HSCC/LGBRIMH/Medical Equipment /2017/02 dt. 26.04.17**

All bidders are requested to note the following:

Item	Tender Specifications	Tender Specifications may please read as
Item no 02, Ambulatory EEG	2. Must have minimum 24 referential EEG Channels, 4 differential channels and 4 DC Channels	Must have minimum 21-24 referential EEG Channels with full range of polygraphic/DC channels for upgradation to PSG
	9. Amplifier must offer high speed USB connectivity to acquisition laptop.	Amplifier must offer high speed USB <b>connectivity/Wifi</b> to acquisition laptop.
	10. Should be supplied with White LED photic stimulator so that photic artefacts doesn't interfere EEG signals.	Should be supplied with <b>White/Red</b> LED photic stimulator so that photic artefacts doesn't interfere EEG signals.
	13. Should have facility to record patient video using HD (High Definition), High resolution color Day/ Night camera with fully synchronized MPEG-4 Video.	Should be <b>supplied</b> with video camera (2MP, IR, PTZ Network Camera) to record patient video using HD (High Definition), High resolution color Day / Night camera with fully synchronized MPEG-4 Video.
	19. EEG Electrode Starter Kit to consist of : 500 Nos. of Gold EEG electrodes, Conductivity Paste – 10000 Gms and Skin Prepping Gel – 10Nos of Tube	EEG Electrode Starter Kit to consist of : 500 Nos. of Gold EEG electrodes, Conductivity Paste –228 Gms and Skin Prepping Gel – 20Nos of each
	22. System must conform to international certifications such as US FDA.	System must conform to international certifications such as US FDA/ <b>European CE</b>
	23. Acquisition laptop should have i7 core Processor, 4 GB RAM, 1 TB Hard Disk, DVD Writer, 14" Screen Size, Set of Speakers, Laser Printer, UPS, Metallic Trolley with Caster.	Acquisition laptop should have <b>windows 10 operating system</b> , i7 core Processor, 4 GB RAM, 1 TB Hard Disk, DVD Writer, 14" Screen Size, Set of Speakers, Laser Printer, UPS, Metallic Trolley with Caster.

➤ **Tender document fee may please read as fee of cost whereas mentioned in the TE document.**

All other terms and conditions of the tender enquiry document shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/ CPP as corrigendum /amendments etc. if any, will be notified on this portal only, no separate advertisement will published in the news papers.

s/d

CGM HSCC (I) Ltd.

For & on behalf of Director LGBRIMH- TEZPUR