

# HSCC (India) Ltd

---

## AMENDMENT – VII

Dated: - 12.09.2016

**Subject:** Amendment to the tender Enquiry Document.

**Ref: Tender Enquiry No.:** HSCC/KCGMC/Medical Equipment/2015-06/05      Dated 09/07/2016

Bidders are requested to note the following Technical Amendment in the item no. 9 Digital Radiography Machine:

| <b>Tender Specification</b>  | <b>Decision</b>  |
|--|--|
| <b>2. Flat Panel Detector:</b><br>II. Detector Panel should be made of amorphous Silicon with Csl or Gadox.  | <b>2. Flat Panel Detector:</b><br>Detector Panel should be made of amorphous Silicon with Csl  |
| <b>14. Approvals:</b><br><br>The bidder should provide USFDA and European CE approved and AERB Type Approval. Please enclose any other certificate required for installation of the machine. | <b>14. Approvals</b><br><br>The bidder should provide USFDA and European CE approved and AERB Type Approval/ NOC. Please enclose any other certificate required for installation of the machine. |

Rest of the technical specification of the item no. 9 will remain same.

All other terms and conditions of the tender enquiry document shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/ CPPP website for corrigendum /amendments etc. if any, as these will be notified on these portals only. No separate advertisement will published in the news papers in this regard.

**CGM, HSCC India Limited  
For and on behalf of DGMR,  
Panchkula**