

# HSCC (India) Ltd

## AMENDMENT - I

Dated: - 15.03.2016

**Subject:** Amendment to the tender Enquiry Document.

**Ref: Tender Enquiry No.:** HSCC/KCGMC/ Medical Equipment / 2015-06/02 Dated 25.02.2016

The pre-bid meeting for the referred tender enquiry was held on 04.03.2016. The following amendments are being incorporated in the referred tender enquiry document.

### Item No. 01 - a). Diagnostic Cystoscope, b) Cystoscope - Paediatrics, c) Resectoscope

S.No.	Description	Amendment by HSCC/ KCGMC
	<b>Technical Specification:-</b>	
<b>Part A</b>	<b><u>Instruments for (a) Diagnostic Cystoscope, (b) Telescope and Instruments for Cystoscope - Paediatrics (Paed Surgery), (c) Instruments for Resectoscope Set</u></b>	
<b>3.a.</b>	<b><u>Diagnostic Cystoscope</u></b>	
<b>i</b>	<b>TELESOPES</b>	
<b>•</b>	Forward Viewing 4mm Telescope 0, 30 & 70 degrees, enlarged view, autoclavable, fiber optic transmission incorporated with HD	Forward Viewing 4mm Telescope 0, 30 & 70 degrees, enlarged view, autoclavable, fiber optic transmission incorporated.
	Telescope and Instruments for Cystoscopy should be of same manufacturer for system compatibility, all items except trolley should be USFDA or European CE (EN type) certified, certificate should be submitted with tender documents.	All items from Part A and Part B (Expect Trolley & plastic tray / cage) should be from same/single manufacturer for better adaptability & Maintenance. All items should be USFDA or European CE (EN type) approved.
<b>3. b.</b>	<b><u>Telescope and Instruments for Cystoscope - Paediatrics (Paed Surgery)</u></b>	

## HSCC (India) Ltd

	Telescope and Instruments for Cystoscopy should be of same manufacturer for system compatibility, all items except trolley should be USFDA or European CE (EN type) certified, certificate should be submitted with tender documents.	All items from Part A and Part B (Expect Trolley & plastic tray / cage) should be from same/single manufacturer for better adaptability & Maintenance. All items should be USFDA or European CE (EN type) approved.
<b>3. c.</b>	<b>Instruments for Resectoscope Set</b>	
•	4Fr 30 degree Cutting Loop Electrode, .012 (minimum 10 Nos. each).	24Fr 30 degree Cutting Loop Electrode, .012 (minimum 10 Nos. each).
•	All items should be ideally / Mandatory (Except Trolley) from same manufacturer for better adaptability & Maintenance, should be USFDA or European CE (EN type) approved. If the products are from different Manufacturer, all items should be adaptable / interphaseable otherwise will not be accepted in final acceptance.	All items from Part A and Part B (Expect Trolley & plastic tray / cage) should be from same/single manufacturer for better adaptability & Maintenance. All items should be USFDA or European CE (EN type) approved.
<b>Part B</b>	<b>High Definition Camera System for Resectoscope</b>	
	<b>Each set should consists of following items:</b>	
1	<b>Full High Definition Digital Camera</b>	
•	The system should have Optical zoom to enhance the quality of Image size & cross specialty standardization of the camera system, regardless of the telescope used having focal length anything between 15 - 31 mm (2X) or better. Camera & Coupler should be one piece (Integrated)	The system should have Optical zoom to enhance the quality of Image size & cross specialty standardization of the camera system, regardless of the telescope used having focal length anything between 15 - 31 mm (2X) or better.
•	Video Outputs: two DVI, one SVHS and one direct fiber optic output	Video Outputs: two DVI, one SVHS and one direct fiber optic output/SDI
2	<b>LED Light Source</b>	
•	LED light source 220volts, 225-300watts OR Xenon light source 300 watts or more	Xenon light source 300 watts or more.

## HSCC (India) Ltd

•	Increased patient safety & added protection in OR with safelight technology Intuitive simple user interface with LCD touch screen.	Increased patient safety & added protection in OR with safelight technology Intuitive simple user interface with LCD touch screen/Touch button on front panel (non sterile area)
•	Should have more than 5000 to 25000 hrs bulb life (with one or spare bulbs)	Lamp life 500 Hours or more (With 2 spare bulbs)
	All the items except trolley & plastic tray/cage should be of same manufacturer for system compatibility, all items except trolley & plastic tray/cage should be USFDA or European CE certified, certificate should be submitted with tender documents.	All items from Part A and Part B (Expect Trolley & plastic tray / cage) should be from same/single manufacturer for better adaptability & Maintenance. All items should be USFDA or European CE (EN type) approved.
	<b>NOTE:</b>	
	All the items at Part A comprising of Diagnostic Cystoscopes, Paediatric Cystoscope & Resectoscopes should be ideally from same manufacturer.	All items from Part A and Part B (Expect Trolley & plastic tray / cage) should be from same/single manufacturer for better adaptability & Maintenance. All items should be USFDA or European CE (EN type) approved.

### Item no. 3 Pulse Oximeter, Qty : 17

a	<b><u>Description of Function:</u></b>	<b>Amendment by HSCC/ KCGMC</b>
	The Display should be LCD / LED with Backlight.	The Display should be 5" or more LCD / LED with Backlight.
g	<b><u>Standard, Safety &amp; Training:</u></b>	
	Should be US FDA / European CE /UL / BIS approved product.	Should be US FDA or European CE

## HSCC (India) Ltd

---

Bid Security submitted in the form of Bank Guarantee should be in the name of HSCC I Ltd., E-6(A), Sector -1 Noida

Bid sale, submission and opening date for following items has been extended as per details given in Table -1:

Table -1

Sl. No.	Description	Revised Schedule
i.	Sale date of the tender	23.03.2016, 2.00 P.M
ii.	Closing date & time for receipt of tender	23.03.2016, 2.30 P.M.
iii.	Time and date of Opening of Techno – Commercial Tenders	23.03.2016, 3.00 P.M.

**Bid Document Fee:** Micro and small scale enterprises (MSEs) are exempted from Bid Document fee, The MSEs fulfilling the prescribed eligibility criteria and participating in the tender shall enclose with their tender a copy of their valid registration certificate with District Industries Centres or Khadi and Village Industries Commission or Khadi and Village Industries Board or Coir board or National Small Industries Corporation or any other body specified by Ministry of Micro and Small enterprises in support of their being on MSE, failing which their tender will be liable to be ignored.

**All other terms and conditions of the tender enquiry document shall remain unchanged.**

**Prospective bidders are advised to regularly visit HSCC website/ CPPP website for corrigendum /amendments etc. if any, as these will be notified on these portals only. No separate advertisement will published in the news papers in this regard.**

CGM, HSCC India Limited  
For and on behalf of DGMER,  
Panchkula