

AMENDMENT No. VIII – Dated 28/2/2018

Sub.: Procurement of Medical Equipment for Kalpana Chawala Government Medical College, Karnal.

Tender Enquiry No.: HSCC/KCGMC/Medical Equipment/2017-06/16 uploaded on 14.12.17

Bidders are requested to note the following technical Amendment:

ITEM no.03 – R F Generator

- Cooling unit is required as it is significantly beneficial to the patient. Since required accessories have been asked, so it should be understood and clarified that compatible RF machine is required.
- Transducer Probe Kit – 10
- Equipment should be USFDA or CE certified

Rest of the specification shall remain unchanged

Item No. 5 Flexible Video Bronchoscope

- 150 watt Xenon light source is required instead of LED light source
- The adaptor for the use of the said equipment with the LED light source of Richard Wolf make (model already available with the department) has to be provided by the bidders for back up.

Rest of the specification shall remain unchanged

ITEM no.10-Automated ESR Analyser

Point no 18. The instrument should be provided with online UPS with half an hour back up.

Point no.20 Consumables, reagents and cards for 15000 samples to be provided at the time of delivery and installation.(bidder to quote unit price of each consumable which will be valid for next seven yaers)

Point no 21 Control both normal and abnormal should be provided for one year and calibrator to be provided once after six months.

Cost of all reagents, consumables ,controls and cards to be frozen for next seven years.

Rest of the specification shall remain unchanged

Item no.11 Fully automated Coagulometer

Point no.22.Initially to start with minimum test menu should be PT, APTT, Fibrinogen ,Factor 8 and 9,and d -dimer followed by other tests later on as mentioned in the specifications and Point no.30 Consumables, Reagents for 5000 PT and APTT tests to be provided. One bottle of reagent for Fibrinogen ,Factor 8,9,d -dimer with a long expiry to be provided preferably one year.

Point no .31 Only Controls of PT and APTT to be provided for 1 year and calibrator to be provided once after six months..

Rest of the specification shall remain unchanged

Item No. 15 Semi Automated Tissue Processor for Tissue Processing

- Tissue processor should be of carousal type with 12 station with at least three paraffin baths Instrument should have vacuum function with fume control system.
- Tissue baskets should be of metal type , glass or high grade plastic.
- Capacity of equipment should be 100-120 tissue cassettes.
- There should be an ergonomic control panel with LCD.
- Infiltration time should be individually programmable for each station.
- There should be programmable delayed start time.
- There should be option for interrupting the process before the end of a run.
- Cassette basket should spin counter clockwise and clockwise within reagent container.
- Inbuilt Battery backup system in case of power failure.
- Voltage requirement should be 220-240V.
- Provided with UPS with 12hrs backup.
- Should have USFDA and CE Europe certification.
- Track record and satisfactory performance report from institute of national importance.(at least one)
- Company should provide 1000 plastic cassettes of good quality along with 1 extra tissue basket bath.

Bid sale, submission and opening date of item No. 3,4,5,10,11,12,13,15 has been extended as per details given in Table -1:

Table -1

Sl. No.	Description	Revised Schedule
i.	Sale date of the tender	08.03.18, 3.00 P.M.
ii.	Closing date & time for receipt of tender	08.03.18, 3.30 P.M.
iii.	Time and date of Opening of Tenders	08.03.18, 4.00 P.M.

All other terms and conditions of the tender enquiry document shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/ CPPP website for corrigendum /amendments etc. if any, as these will be notified on these portals only. No separate advertisement will published in the news papers in this regard.

S/d
CGM, HSCC India Limited
For and on behalf of DGMER, Panchkula