

**Amendment – IV**

**04.09.2013**

**INVITATION FOR BIDS FOR SUPPLY, INSTALLATION, TESTING & COMMISSIONING OF MEDICAL EQUIPMENTS AND CSSD EQUIPMENTS FOR 200 BEDDED NEPAL- BHARAT MAITRI EMERGENCY AND TRAUMA CENTRE (NBMETC), KATHMANDU, NEPAL**

**Bid No.: BID REFERENCE: IFB No. HSCC/PUR/NEPAL/2013-14/1 Dated 06.07.2013**

Bidders are requested to note the following amendment in the technical terms:

**Name of the equipment: Cautery Item No. 22**

**Name of the Firm: Bidder 1**

<b>Amended by MEA/HSCC in Amendment -II</b>	<b>Bidder Query after Amendment</b>	<b>Amended by MEA/HSCC in this amendment</b>
<p>The machine should have bipolar generators designated to produce simultaneous dual user monopolar outputs.</p>	<p><b>Objection :</b> It can't be possible simultaneous bipolar and monopolar outputs, because if once Doctor use monopolar and beep sound comes, in the same time if Doctor use bipolar and beep sound comes how can he recognize that whether beep sound is for bipolar or monopolar.</p> <p><b>Suggestion:</b> The machine should have bipolar generators designated to produce bipolar and monopolar outputs with switchover.</p>	<p>No Change</p>
<p>The machine should be supplied along with 3 sets of each of the following- Metal Foot Switch</p>	<p><b>Objection :</b> Metal Switch is not long lasting as compare to water proof switch, when we use metal pedal , it can't get stuck after using some time whether it is bipolar mode or monopolar mode and it can sent return electrode to the unit.</p> <p><b>Suggestion:</b> Water proof/ Explosion proof/Foot Switch</p>	<p>The machine should be supplied along with 3 sets of each of the following- Water proof/ Explosion proof/Foot Switch</p>

**Item No. 32- Automatic Steam Sterilizer**  
**Name of the supplier: Bidder -1**

Amended by MEA/HSCC in Amendment -II	Bidder query after Amendment	Amended by MEA/HSCC in this amendment
<p>Turnkey work pertaining to Civil, Electrical, Plumbing, sanitary, furniture, servo stabilisers, U.P.S. etc. required for successful installation testing and commissioning of the system will be the responsibility of the bidder.</p>	<p>In tender specification you have mentioned "Turnkey work pertaining to Civil, Electrical, Plumbing, sanitary, furniture, servo stabilizers, U.P.S. etc. required for successful installation testing and commissioning of the system will be the responsibility of the bidder"</p> <p>You are requested to kindly provide us with the auto cat drawings with room dimensions where these sterilizers need to be installed, so that we can calculate cost of civil work.</p> <p>The same things were also discussed at pre bid meeting and surprisingly noting has been amended, if still specification is not amended then what is the need of open tender inquiry and pre bid meeting.</p>	<p>Additional information:</p> <p>The room size where the equipment is to be installed is 19 ft X 19 ft. the source of power from the room is around 50 mtrs.</p>

**All other terms and conditions of the bid document shall remain unchanged.**

**S/d**  
**Joint Secretary –DPA (III)**  
**Ministry of External Affairs,**  
**New Delhi**