

AMENDMENT No - VII Dated 06.06.2016

Subject: Amendment to the tender Enquiry Document

Ref: Tender Enquiry No.: HSCC/SJH/Med. Eqpt./15 dated 08.04.2016

Existing as :

Sl. No.	Description	Schedule
i.	Closing date & time for receipt of Tender	06.06.2016, 1430 hrs IST
ii.	Time and date of opening of Techno - Commercial tenders	06.06.2016, 1500 hrs IST

To be read as :

Sl. No.	Description	Schedule
i.	Closing date & time for receipt of Tender	14.06.2016, 1430 hrs IST
ii.	Time and date of opening of Techno - Commercial tenders	14.06.2016, 1500 hrs IST

Section - VIII

Item No. 7 Endo-Bronchial Ultrasound System (EBUS)

1.1 Existing as: Field of view At least 100° (at least 45° forward oblique)

To be read as : - Field of view 80° and above (at least 35° forward oblique)

Existing as 1.13 Scan Angle 75°

1.13 Amended as Scan Angle 65° or more

2. Digital Color Video Processor

Existing as: 2.1 Single CCD color, high resolution HDTV & narrow board imaging compatibility

2.1 **Amended as:** - Single CCD color, high resolution HDTV & narrow board imaging compatibility / I scan

2.2 **Existing as:** xenon lamp/LED with a spare bulb

2.2 **Amended as:** -xenon lamp 300 watts

3. Monitor

3.1 **Existing as:** 19" High resolution LCD or higher (LED) color monitor compatible with processor with full range of colors & inputs including viewing angle as needed for proper functioning

3.1 **Amended as:** - 21" High resolution LCD or higher (LED) color monitor compatible with processor with full range of colors & inputs including viewing angle as needed for proper functioning

3.2 Power consumption 230V

4.61 **Existing as:** High definition dynamic tissue harmonic imaging

4.6 1 **Amended as:** - Deleted.

5. Essential Accessories

5.5. **Point Added:** At least 100 Needles and 100 balloons

Item No. 8 POTABLE ULTRASOUND CUM ECHO (COLOUR DOPPLER SYSTEM) FOR USE IN ICU

Existing as: Point 10. The unit must be compact, portable and lightweight, weighing less than 5 kg to use in ICU.

To be Read as: Point 10 The unit must be compact, portable and lightweight, weighing less than 7 kg to use in ICU.

Existing as: Point 12. Flat LCD/ TFT monitor of at least 10 inches with flicker free image.

To be Read as: Flat LCD/ TFT monitor of at least 12 inches and above with flicker free image.

Existing as: Point 16 Must be able to operate both on AC and inbuilt battery. Inbuilt battery pack should be self-recharging and should last at least for 1 hour when fully charged.

To be Read as: Must be able to operate both on AC and inbuilt battery. Inbuilt battery pack should be self-recharging and should last at least for 45 minutes & above when fully charged.

16. Transducers to be supplied as standard

a. **Existing as:** Phased Array Probe 2-5MHz for abdominal, cardiac, FAST applications.

To be read as: Phased Array Probe 2-5MHz (± 1 MHz) for abdominal, cardiac, FAST applications

b. **Existing as :** High Frequency Linear transducer 6-13 MHz (± 1 MHz) with less than 40 mm size for vascular access & vascular, small parts Imaging.

To be read as: High Frequency Linear transducer 6-13 MHz (± 1 MHz) with less than 50 mm size for vascular access & vascular, small parts Imaging.

c. **Existing As:** Convex probe 2-5 for abdominal & ob/gyn applications.

To be read as: Convex probe 2-5(± 1 MHz) for abdominal & ob/gyn applications

f. **Existing As:** Carry bag, Trolley, with a provision of locking the main system.

To be read as: Carry bag, Trolley, with a provision of locking the main system from the same manufacture.

18 New Point Added as: At least 1 lac channels.

Optional probes to be quoted

- **Existing As:** Phased array 4-8 MHz for paediatric cardiac Applications with PW & CW facility.
- **To be read As:** Phased array 4-8 MHz (± 1 MHz) for paediatric cardiac Applications with PW & CW facility.

- High Frequency Linear transducer 6-13 MHz (+/- 1MHz) or more with less than 26 mm footprint size for nerve blocks, vascular access, Vascular Imaging in Pediatric patients. Higher frequency will be preferred.

ESSENTIAL REQUIREMENT: Existing As: The firm must have minimum number of 50 installations of the same model in India, attach list of installations, and also provide performance certificates.

To be read as: The firm must have minimum number of 50 installations of similar model in India, attach list of installations, and also provide performance certificates.

SECTION – IX

QUALIFICATION CRITERIA

Existing as :

2. (a) A Tenderer quoting as Manufacturer should have supplied and installed in last Five years from the date of Tender Opening, at least 50% of the quoted quantity of the similar equipment which is functioning satisfactorily in Government Hospitals / Private Hospitals / PSU Hospital/ UN Agencies/Laboratories anywhere in India. Tenders shall submit Performance Certificate / Installation reports & Order Copies in respect of the above.

2. (b) A Tenderer quoting as authorized representative of the manufacturer should have executed at least one contract in the last five years from the date of tender opening of similar equipment which is functioning satisfactorily, in Government Hospitals/Private Hospitals/PSU Hospital/UN Agencies/Laboratories anywhere in India. Tenders shall submit Performance Certificate / Installation reports & order copies in respect of the above.

To be read as :

2. (a) The Manufacturer should have supplied and installed in last Five years from the date of Tender Opening, at least 33 % of the quoted quantity of the similar equipment meeting major parameters of technical specification which is functioning satisfactorily.

2 (b). The Tenderer quoting as authorized representative of the manufacturer meeting the above criteria 2 (a) should have executed at least one contract in the last five years from the date of tender opening of similar equipment meeting major parameters of Technical

specification which is functioning satisfactorily, anywhere in India of the same manufacturer.

SECTION - IV GENERAL CONDITIONS OF CONTRACT (GCC)

Liquidated damages:

Existing as :

23.1 Subject to GCC clause 26, if the supplier fails to deliver any or all of the goods or fails to perform the services within the time frame(s) incorporated in the contract, the Purchaser/Consignee shall, without prejudice to other rights and remedies available to the Purchaser/Consignee under the contract, deduct from the contract price, as liquidated damages, a sum equivalent to 0.5% per week of delay or part thereof on delayed supply of goods and/or services until actual delivery or performance subject to a maximum of 10% of the contract price. Once the maximum is reached Purchaser/Consignee may consider termination of the contract as per GCC 24.

During the above-mentioned delayed period of supply and / or performance, the conditions incorporated under GCC sub-clause 22.4 above shall also apply.

To be read as:

23.1 Subject to GCC clause 26, if the supplier fails to deliver any or all of the goods or fails to perform the services within the time frame(s) incorporated in the contract, the Purchaser/Consignee shall, without prejudice to other rights and remedies available to the Purchaser/Consignee under the contract, deduct from the contract price, as liquidated damages, a sum equivalent to 0.5% per week of delay or part thereof on delayed supply of goods and/or services until actual delivery or performance subject to a maximum of 10% of the contract price. Once the maximum is reached Purchaser/Consignee may consider termination of the contract as per GCC 24.

During the above-mentioned delayed period of supply and / or performance, the conditions incorporated under GCC sub-clause 22.4 above shall also apply.

23.2 In the event of delay in submission of Proforma Invoice beyond 7 working days from the date of notification of award, the delay shall be to the account of supplier & Purchaser shall deduct Liquidated damages, as per clause 23.1. Proforma Invoice should be strictly as per the terms & conditions mentioned in Notification of Award / Tender Conditions.

23.3 Proforma Invoice submitted by supplier is found to be deficient, because of which purchaser is unable to open the letter of credit, delay shall be to the account of supplier & purchaser shall deduct liquidated damages as per clause 23.1.

All other terms & conditions remains unaltered.

Medical Superintendent

**Safdarjung Hospital &
VMMC,**

New Delhi.