

Amendment no I Dated 06.06.2018
HSCC/PUR/AIIA/Blood Bank /A dt. 10.05.2018

Procurement of Medical Equipment under AIIA, Sarita Vihar

All bidders are requested to note the following:

Item No	Last date & time sale/downloading of Tender document	Last closing/submission date & time for receipt of tender	Date of opening of Techno – Commercial bids.	Earlier date of opening	Amendment Status for opening date
1 - 19	18.06.2018, up to 14.00hrs IST	18.06.2018, 14.00hrs IST	18.06.2018, 14.30hrs IST	11.06.2018	Date extended with Technical specification updating

Item no 03, Deep Freezer (-80 deg C)

Tender Specification	May please read as
8. Compatible Automatic Voltage Stabilizer of suitable rating	8. Compatible Automatic Servo Voltage Stabilizer of suitable rating

Item no 04, Deep Freezer (-40 deg C)

Tender Specification	May please read as
18. Compatible Automatic Voltage Stabilizer of suitable rating	18. Compatible Automatic Servo Voltage Stabilizer of suitable rating

Item no 09, Refrigerated Centrifuge

Tender Specification	May please read as
3. Swing bucket blood bank rotor: Should have swing bucket blood bank rotor with metal buckets, 6 x 2000 ml. with or without wind shield.	3. Swing bucket blood bank rotor: Should have swing bucket blood bank rotor with metal buckets, 6 x 2000 ml. with wind shield.

All other terms and conditions of the tender enquiry document shall remain unchanged. Prospective bidders are advised to regularly visit HSCC website/ CPP as corrigendum /amendments etc. if any, will be notified on this portal only, no separate advertisement will published in the news papers.

Sr. CGM-I, HSCC (I) Ltd
For & on behalf of Director, AIIA, Sarita Vihar