

All Bidders**Amendment –I**

Project: Supply, Installation, Testing & Commissioning of Anaesthesia Pendant for OTs at Mother & Child Block at All India Institute of Medical Sciences (AIIMS), New Delhi.

IFB No. HSCC/SES/Pendant/AIIMS/M&C/2019 dated 12.09.2019

This has reference to above IFB.

The following Amendment may be noted which shall be treated as part of the tender document and to be submitted duly signed & stamp along with tender.

| S.No | Bidders' Queries | Reply |
|------|---|--|
| 1. | <p>Point No.c) It should have electromagnetic/pneumatic brakes. Request - It should have electromagnetic brakes only with brake light. The provision for Air pipeline is not sufficient and there is no dedicated pipeline for Pneumatic brakes asked in the tender, hence it should be deleted from tender. Request you to kindly amend the same to "Electromagnetic brakes only with brake light".</p> | Tender terms and conditions prevail. |
| 2. | <p>Point no. e) – The Pendant service Heads should be modular with minimum 800 mm head. Request -The Pendant Service Head should be modular with minimum 1500mm head. Since you require 10 Nos. Gas Outlets , 10 Nos. Electrical Sockets. 2 Nos. RJ Data Socket. Earthing Points, Manometer and other accessories , 800mm service head will not be going to provide solution for fixing the above referred accessories. You are therefore requested to kindly increase the size of Pendant Service Head (Column Length) to minimum 1500 mm.</p> | The Pendant service heads shall be modular with min. 1500 mm head. |
| 3. | <p>Point No. i) – Pendant supplier should provide cut outs for Patch panels. Request – To be deleted from technical specifications of the tender. There is no integration system in the Mother & Child Block , AIIMS, New Delhi Project, hence Patch Panels are not required and</p> | Tender terms & conditions prevail. |

| | | |
|----|--|---|
| | needs to be deleted from the technical specifications of tender. You are requested to kindly delete the same. | |
| 4 | <p>Responsibility of Bidder :</p> <p>Point No. 5 – Training should be imparted to the hospital staff for 2 weeks by the contractor.</p> <p>Request – Training should be imparted to the hospital staff for 1 -2 days by the contractor. It is invalid to ask for two weeks training for Pendants. Kindly amend this clause of tender.</p> | Training should be imparted to the hospital staff for one week by the contractor. |
| 5. | <p>Volume 1, 1.2 Completion Period</p> <p>Page 5</p> <p>2 months from the date of order of commencement.</p> <p>Request – kindly amend to 4 months providing us enough buffer for production and shipment.</p> | Tender terms & conditions prevail. |
| 6. | <p>Volume IV- Technical Specification</p> <p>Pendants for Anaesthetist</p> <p>Double arm movable pendant for Anaesthetist</p> <p>Shelf of the Pendant should be made of high quality Bio Clean Material.</p> <p><u>Request</u></p> <p>Bio clean is a trade mark of a particular company and only one pendant manufacturer uses this. All the pendant manufacturers worldwide manufactures Pendants including its shelves which are easily cleanable with any hospital disinfectants. In case you want an anti- bacterial coating , the same should be specified as Bio Clean/anti bacterial coating on the Pendant shelf. The current specifications are very restrictive and will allow particular bidder to participate. Request you to amend accordingly in order to allow more manufacturers to participate.</p> | <p>Double arm movable pendant for Anaesthetist</p> <p>Shelf of the Pendant should be made of high quality Bio Clean Material/ anti bacterial coating.</p> |

All other terms & conditions remain unchanged.

Sr. Chief General Manager -I, HSCC (I) Ltd.
For and behalf of Director, AIIMS, New Delhi