

HSCC (India) Ltd

AMENDMENT No –I dated 24.11.2014

Sub.: Procurement of Medical Equipment for AIIMS, Raebareli.

Ref: Tender Enquiry No.: HSCC/PUR/AIIMS/Raebareli/Temp OPD/2014 dated 5.11.2014

Based on the bidders queries raised during pre- bid meeting held on 14.11.2014, all bidders are requested to note the following amendments in the aforesaid tender:

Sl. No.	Description	Revised Schedule
i.	Last date and sale of tender enquiry documents	on 10.12.2014, 12.00 hrs IST
ii.	Closing date & time for receipt of tender	on 10.12.2014, 14.30 hrs IST
iii.	Time and date of Opening of Techno – Commercial Tenders	on 10.12.2014, 15.00 hrs IST

ITEM NO 1: Biochemistry Analyzer

Point No.	Existing Tender Specification	Amended As
13	ON-BOARD PARAMETERS TESTS: Minimum 50 on-board parameters tests.	ON-BOARD PARAMETERS TESTS: Minimum 45 or more on-board parameters tests.
17	Cuvettes : may be reusable, permanent or disposable, specify recurring cost, if any	Cuvettes: Must have permanent/reusable type of hard glass. It should have the facility to change individual cuvettes.
28	The equipment to be supplied should have FDA & CE certification.	The equipment to be supplied should have US-FDA or European CE certification.

ITEM NO 3: Anaesthesia Machine

Point No.	Existing Tender Specification	Amended As
16	16. Machine should have US FDA Certified.	Machine should have US FDA or European CE Certified.
D (6)	6. LED Display	LED/LCD/TFT display of at least 6.4 inch or more.
D (13)	13. Should have tidal volume range from 20 ml to 1500 ml.	Should have tidal volume range from 20 ml to 1400 ml or more.
D (2)	Volume Controlled Ventillation	Should have Volume Controlled Ventilation, Pressure Controlled Ventilation (PCV), pressure Support and Synchronous intermittent Mandatory ventilation (SIMV)

HSCC (India) Ltd

ITEM NO 4: Operation Table Hydraulic for Minor OT

Point No.	Existing Tender Specification	Amended As
3 (4)	Should have a manual position selector	Should have a manual position selector Manual Position Selector to position table to different required position for surgery.
3 (6)	a. Minimum 600 ±50 mm Max. 1050mm or More (without mattress)	Minimum 600 ±50 mm Max. 950mm or More (without mattress)
3(6)	e. Trendelenburg: 25 to 30 degree f. Anti trendelenburg: 25 to 30 degree	e. Trendelenburg: ±25 degree or more f. Anti trendelenburg: ±25 degree or more

The Bid Submission date revised from 28/11/2014 to 10/12/2014.

The EMD in case submitted in the form of BG (Bank Guarantee)/ Banker cheque it shall be valid for 165 days from the extended Techno-Commercial tender opening date i.e. 10/12/2014.

All other terms and conditions of the tender enquiry document shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/ CPP as corrigendum /amendments etc. if any, will be notified on this portal only, no separate advertisement will published in the news papers.

Deputy Secretary (MOHFW)