

All Bidders**Amendment –I**

Project: Supply, Installation, Testing & Commissioning of CSSD for Hospital Block, All India Institute of Medical Sciences (AIIMS), Raebareli.

IFB No. HSCC/SES/CSSD/AIIMS/Raebareli/2020 Date: 03.12.2020

This has reference to above IFB. The following Amendment may be noted which shall be treated as part of the tender document and to be submitted duly signed & stamp along with tender.

Sr. No.	Bidders' Query	Reply
1.	<p>Volume 1 – Prequalification Documents Page No. 5: Prequalification Qualification Criteria 2.2 * Similar nature of works means successful completion of supply, installation, testing and commissioning of CSSD equipment.</p> <p>Volume 1 – Prequalification Documents Page No. 5: Prequalification Qualification Criteria 2.2 *Similar nature of works means successful completion of supply, installation, testing and commissioning of CSSD equipment with turnkey services.</p> <p>Kindly amend the same.</p>	<p>Tender terms remain unchanged</p>
2.	<p>Volume 3 – Special Conditions of Contract Page SCC-38: Payment Terms</p> <p>- 70% of the BOQ contract rates on delivery of equipment/items at site after inspection, passing and issue of dispatch clearance on pro-data basis.</p> <p>- 20%% of BOQ contract rates on satisfactory take over certificate by client after erection and installation, testing and commissioning of equipment on pro-data basis.</p> <p>- 10% of BOQ contract rates after successful completion of trial run of 30 days from the date of handover to the client on pro-data basis.</p> <p>Volume 3 – Special Conditions of Contract Page SCC-38: Payment Terms</p>	<p>- 75% of the BOQ contract rates on delivery of equipment/ items at site after inspection, passing and issue of dispatch clearance on pro-data basis.</p> <p>- 15%% of BOQ contract rates on satisfactory take over certificate by client after erection and installation, testing and commissioning of equipment.</p> <p>- 10% of BOQ contract rates after successful completion of trial run of 30 days from</p>

	<p>- 80% of the BOQ contract rates on delivery of equipment/items at site after inspection, passing and issue of dispatch clearance on pro-data basis.</p> <p>- 10%% of BOQ contract rates on satisfactory take over certificate by client after erection and installation, testing and commissioning of equipment on pro-data basis.</p> <p>- 10% of BOQ contract rates after successful completion of trial run of 30 days from the date of handover to the client on pro-data basis.</p> <p>Kindly amend the same.</p>	the date of takeover by the client
3.	<p>Volume – 03 SCC, Page No. SCC-14, The Contractor must fill up price in Indian Rupees against each item of BOQ (Volume V) on line both in words and figures in the blank spaces provided in the respective columns.</p> <p>Volume – 03 SCC, Page No. SCC-14, The Contractor must fill up price in Indian Rupees / Foreign Currency against each item of BOQ (Volume V) on line both in words and figures in the blank spaces provided in the respective columns.</p> <p>You are requested to kindly give provision for Letter of credit to execute the project smoothly.</p>	Tender terms remain unchanged
4.	<p>Item No. 6 – Ultrasonic Cleaner</p> <p>Page No. 10 of Technical Specification: h. Ultrasonic cleaner’s manufacturer should be ISO 13485 certified with declaration of conformity to BIS/USFDA/European CE should be either ISO-13485 certified with declaration of conformity as per BIS/US FDA/European CE certified with four digit notified body number.</p> <p>Page No. 8 of Technical Specification: h. Ultrasonic cleaner should be either ISO- 13485 with declaration of conformity as per US FDA/European CE certified.</p>	Tender terms remain unchanged
5.	<p>Page No. 13 of Technical Specification: Item No. 23 - Modular Sterilizing Basket Size: 585x395x195 Item No. 24 – Modular Sterilizing Basket Size: 585x395x100</p> <p>Page No. 13 of Technical Specification:</p>	<p>Size: 585x395x195 or Size: 600 x 300 x 290 For item no.-23 & 20</p> <p>Specification of Modular Sterilizing Basket is as per SPRI/ STU standard Size</p>

	<p>Item No. 20 - Modular Sterilizing Basket Size: 600 x 300 x 290</p> <p>Kindly note that Specification of Modular Sterilizing Basket asked in tender is as per SPRI standard Size while Sterilizer has been asked as per STU standard. We request you to kindly amend the size as per STU.</p>	
6.	<p>Furniture Items</p> <p>All CSSD Furniture items should be accepted with $\pm 10\%$ deviations. Kindly amend the same.</p>	$\pm 10\%$ deviations accepted.
7.	<p>We would like to request the budget is not sufficient it may please be enhanced as also in previous bid the budget was not in accordance to the kind of turnkey work asked, also you have added the some of the equipment in this tender like instrument trays, RO water, sterile containers with good quantity and etc. And the exchange rate of European currency has risen. And this also need to enhance the budget so please look into the matter and enhance the budget of 1.5 more Crore excluding CMC from current budget of 6 crore. i.e. approx. 7.5 Crore excluding cmc. This shall help you to conclude this project and enable you to further retendering of such a time bound project.</p>	Tender terms remain unchanged.
8.	<p>The budget is not sufficient it may please be enhanced as also in previous bid the budget was not in accordance to the kind of turnkey work asked, further the cost after this covid -19 pandemic also increased at every point right from the manufacturing till execution for additional safety and limitation of facilities. Also you have added the some of the equipment in this tender like instrument trays, RO water, sterile containers with good quantity and etc. this also need to enhance the budget so please look into the matter and enhance the budget of 1.5 Crore more excluding CMC from current budget of 6 crore. i.e. approx. 7.5 Crore excluding CMC. This shall help you to conclude this project and enable you to avoid further retendering of such a time bound project.</p>	Tender terms remain unchanged.
9.	<p>Further we would like to request you that kindly extend the tender at least 10 days.</p>	The Tender submission date is extended from 14.12.2020 to 17.12.2020 and bid should be valid for 180 days from the date of bid submission i.e. from 17.12.2020.

10.	<p>Foreign currency quoting should be allowed for major equipment's i.e., Steam Sterilizer for 8 STU and 4 STU and Washer Disinfector and Plasma sterilizer and smaller equipment Table top sterilizer may also be at least allowed for foreign currency and the BOQ may please be revised accordingly. Also, please clarify about the exemptions certificates may please made available.</p> <p>You have also allowed above points in your previous tender no. HSCC/SES/CSSD/AIIMS/Raebareli/2020 Dated: 15.09.2020. Kindly consider the above points and upload revised BOQ.</p> <p>Time of completion of project is 4 months, which should be amended to 5 months from the date of site handover to vendor.</p>	Tender terms remain unchanged
11.	<p>We would like to bring to your notice that there are certain tender specifications for some equipment where BIS certificate is required. But for equipment like Instrument Double Door washer disinfector (item /Sr.no. 4 of the tender), the BIS body itself does not have any standards and hence the BIS body does not/ cannot certify this product under BIS.</p> <p>Hence, we would request you to consider other equivalent ISO/Indian standards like ISO 15883 for this equipment.</p>	IS/ISO 15883 /US FDA/European CE certified with four digit notified body number
12.	<p>Kindly note that you have asked for BIS Certification for Washer Disinfector and Plasma Sterilizer also.</p> <p>We confirm that there is no BIS Certification for both these products.</p> <p>Therefore, we request you to amend the same, so that we can Quote for the above Tender.</p>	<p>Washer Disinfector- IS/ISO 15883 /US FDA/European CE</p> <p>Plasma Steriliser –Tender terms remain unchanged</p>

All other terms & conditions remain unchanged.

Chief General Manager, HSCC (I) Ltd.
As Executing Agency of Ministry of Health & Family Welfare