

HSCC (I) Ltd

Amendment –I dated 04.01.2019

IFB No. HSCC/PUR/CNCI-Kolkata/Medical Equipment/101 dated 18.12.2018

Bid sale, submission and opening date for above IFB has been extended as per details given in **Table -1**

Sl. No.	Description	Existing Schedule	Revised Schedule
i.	Sale date of the tender	18.12.18 to 08.01.2019, 13.00 Hrs	18.12.18 to 21.01.2019, (up to 13.00 Hrs)
ii.	Closing date & time for receipt of tender	09.01.19, 2.00 P.M.	22.01.19, 2.00 P.M.
iii.	Time and date of Opening of	09.01.19, 2.30 P.M.	22.01.19, 2.30 P.M.

The following amendment as under:

1. **General Information:**

Sl	Tendered Technical Specification	Amendment as
1	Tender Specifies under Section 1 Point (A) page 7:- Original Offline EMD, Affidavit and Bid Summary Sheet is to be submitted physically.	All offline Documents (like Original Offline EMD, Affidavit etc, as per Bid Document) are to be submitted in sealed envelope at HSCC, NOIDA Office:E-6(A), Sector - I, NOIDA - 201 301.
	OTHER INFORMATIONS	The EMD shall be in favour of HSCC (I) Ltd.

2. **Technical Amendment:**

Item No. 2 : Cytospin

Sl	Tendered Technical Specification	Amendment as
1	Warranty – 2 - 3 yrs.	GCC Clause 15.2 supersede in this case, wherein Warrantee shall be for a Period of 5 Years & CMC to be quoted for 5 Years.

Item No. 5: Blood Glass Analyzer (ABG)

Sl	Tendered Technical Specification	Amendment as
1	Should have pH, pCO ₂ , pO ₂ , Sodium, Potassium, Calcium, Chloride, total hemoglobin, Hematocrit, oxygen saturation, Glucose, Lactate, Urea, Neonatal Bilirubin, Barometric Pressure, O ₂ Hb, HHb, COHb, MetHB&Thb Co ox as a measured parameter. The parameters	"Should have pH, pCO ₂ , pO ₂ , Sodium, Potassium, Calcium, Chloride, total hemoglobin, Hematocrit, oxygen saturation, Glucose, Lactate, Neonatal Bilirubin, Barometric Pressure, O ₂ Hb, HHb, COHb, MetHB&Thb Co ox as a measured and calculated parameter. The parameters can be activated individually for measurement"

Item No. 9: Automatic Glass Washer

SI	Tendered Technical Specification	Amendment as
1	Lower spindle rack of type 304 stainless steel with 36 detachable large spindles, 18 glassware holders, 36 dips and 10 additional interchangeable small spindles.	As per tender "It is desirable
2	Pump to recirculate 96 gal/min (363 litres) when operating on 230 volt. 50 Hz and 112 gal/min (424 litres/min) on 230 volt. 60 Hz.	Now be read as "up to 500litres/min"

Item No. 10: Automatic Microbial Identification system

SI	Tendered Technical Specification	Amendment as
1	Maldi-Tof system	
2	Should have minimum of 1 m Flight tube	As per tender upto 1m Flight Tube is accepted

Item No. 12: Automatic Slide Strainer

SI	Tendered Technical Specification	Amendment as
1	Autostainer XL- H & E stainer Compact, space saving Imported H&E stainer for routine applications	Now be read as "Autostainer H & E stainer"
3	Exact incubation times	It denotes incubation time for each reagent station.
17	Reagent container volume: 450 ml	Min 400 ml
24	Should have CE certificate	CE/USFDA and ISO 9001 certificate
25	Should have minimum 10 installation in government institution	Government Institute installation is preferable, but installation in non-government institute shall also be considered. However, installation in atleast 5 Government Institute is required.

Item No. 13: Automatic TB Culture system

SI	Tendered Technical Specification	Amendment as
1	AUTOMATED SYSTEM FOR MICRO BACTERIAL CULTURE & SENSITIVITY TESTING (BD MAKE BACTEC MGIT 960) :-	Higher throughput of more than 900 station of any make will be accepted.

Item No. 14: Automatic TB Culture system

SI	Tendered Technical Specification	Amendment as
2	Configurations: – Basic instrument – Vacuum function – Fume control system	Now read as "Configurations: – Basic instrument --Vacuum function with fume control system".

	– Vacuum function with fume control system	
3	Option: two basket loading	Two basket loading.
17	Number: 10 (9)	Both 9 and 10 will suffice the requirement.
17.2	Should have CE and DIN EN ISO 9001 certificate	Should have CE/ USFDA and ISO 9001 certificate
	Standard tissue basket:	
18	Number: 1 (2 optional)	Tissue basket 1 is must. 2 is desirable.
	Other conditions:	
	Should have minimum 20 installation in Government Institutes in Eastern India.	Government Institute installation is preferable, but installation in non-government institute shall also be considered. However, installation in atleast 5 Government Institute are required.

Item No. 21: Bio Safety Cabinet

Sl	Tendered Technical Specification	Amendment as
2	Approximate dimensions of cabinet: Exterior 1500 H x 1300 W x 800 D; Interior 800 H x 1200 W x 500D; Front of cabinet angled by 10° for user ergonomics	All dimensions are desirable

Item No. 49: Glassware, Media

Sl	Tendered Technical Specification	Amendment as
1	Borosil Glassware made 90 & 120 mm plates, Borosil flask of 500 ml, 1, 2 Ltr & 1, 2 Ltr calibrated jar. Dehydrated media:	The procurement of Glassware, Media (Item No. 49) may be cancelled from this tender. The same shall be procured later.

Item No. 50: Grossing Station with Fume Hood

Sl	Tendered Technical Specification	Amendment as
1	The Following may be added in the Tendered Specification:	The Following may be added in the Tendered Specification: <i>"Material of Construction: The body should be made of Stainless Steel 304 Grade".</i>

Item No. 62: Miscellaneous Glassware strains etc

Sl	Tendered Technical Specification	Amendment as
1	Borosil Glassware made different sized glass products used in Laboratory.	The procurement of Miscellaneous Glassware strains etc (Item No. 62) may be cancelled from this tender. The same shall be procured later.

Item No. 74: Rotary Microtome with 50 spare knives

Sl	Tendered Technical Specification	Amendment as
11	Should have CE certificate	Both CE, USFDA will be considered

Item No. 78: Tissue embedding station

Sl	Tendered Technical Specification	Amendment as
5	Operating temperatures: 50 °C to 75 °C, adjustable in 1 ° C increments	All this temperature cut off is of desirable parameters.
9	Display: 5.7 inch capacitive LCD touch screen	5.7 inch is desirable however minor variation may be considered. Not below 5.5 inch will be considered
14	Foot switchHeatable Forceps	As per tender
15	Should have CE certificate	Both CE, USFDA will be considered

All other terms and conditions of the tender enquiry document shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/ CPPP website for corrigendum/amendments etc. if any, as these will be notified on these portals only. No separate advertisement will published in the news papers in this regard.

s/d

**Sr.CGM-1, HSCC(India)Limited
For and on behalf of Director, CNCI, Kolkata**