

HSCC (I) Ltd

Amendment –I Dated ,27.02.2020

Tender Ref : HSCC/SJH/Medical Equipment /2020/44 Dated : 29.01.2020

Subject: Extension of Bid submission Date.

Technical Amendment

Item No. 4 RO Plant Nephrology Department

Please find enclosed the technical amendment for Item No. 04, RO Unit (For Nephrology Department).

Sr. No.	Tendered Specification :	To be Read As:
Pre Treatment Point III	Raw Water Tank having food grade quality for at least 5000 Liters capacity to store Raw Water.	Raw Water Tank having food grade quality for at least 2000 Liters capacity to store Raw Water.
Pre Treatment Point V	Particle filter, cartridge filter type of 50 microns to 10 microns.	Particle filter, cartridge filter type of 20 microns only.
Pre Treatment Point VIII	Should have fine filter, cartridge type of 5 micron & 1 micron	Should have fine filter, cartridge type of 5 micron only
RO Unit Point IV	Should supply 1500 Liter /hour of permeate, which should be upgradable to 2000 liters/hour	Should supply 1500 liter/hour of permeate for 20 Hemodialysis Machine and 2 dialyzer reprocessor only No upgradability required.
RO Unit IX	Should have fully automatic volume controlled heat and chemical disinfection cycle.	Should be fully automatic microprocessor controlled heat or chemical disinfection
Post Treatment System Point 1	Should have appropriate material and shape Permeate Storage Tank of at least 5000 Liters capacity with level control system.	Should have appropriate material and shape permeate storage Tank of up to 2000 liters capacity with level control system.

The bid submission date for Item No. 04, RO Unit is extended from 28.02.2020 to **06.03.2020**.

The reply of the pre bid queries of other following items tendered are awaited Meanwhile, their bid submission date is also extended from 28.02.2020 to **06.03.2020**

1. Item No. 1,2,5, (For CTVS Department)
2. Item No. 03, (For ICU & Pulmonary Department)

All other terms and conditions of the tender enquiry documents including Amendments issued so far shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website and CPPP Website for the Corrigendum/amendments etc. if any, as these will be notified on these portals only. No separate advertisement will published in the newspaper in this regard.

**CGM (Proc)
HSCC I LTD**