

Amendment no I Dated 20.09.2017

HSCC/PUR/AIIA/Blood Bank Dated: 01.09.2017

All bidders are requested to note the following:

Item	Tender Specifications	Tender Specifications read as
<p align="center">Item no 09, Refrigerated Centrifuge</p>	<p>11. Power Requirement: Input Voltage 220/240 V single or three phase as per machine requirement.</p>	<p>Power Requirement: Input Voltage 220-400 V single or three phase as per machine requirement.</p>
	<p>20. Automatic Line voltage corrector/Voltage Stabilizer: A line voltage corrector of appropriate rating should form part of standard configuration. Copper wound single phase automatic line voltage corrector conforming to IS: 9815(Pt.I)/94 with latest amendments or equivalent international standards fitted with a voltmeter and switch to indicate output/input voltage as under: Capacity/rating: 10 KVA: As per the requirement of the equipment. Input voltage: 140 to 280 volts, 50 cycles. Output voltage: 220 volts \pm 10% volts. Input-out voltmeter and amperemeter. Protection: high-low voltage cut-off, overload and short circuit protection. The equipment should be supplied with 2 meter chord at input and fitted with plugs of appropriate rating (15Amp.). Make of the line voltage corrector shall be indicated.</p>	<p>Suitable Automatic Line Voltage Corrector/Voltage Stabilizer as per machine requirement should be supplied.</p>
<p align="center">Item no 13 Cell Counter (Automated Hematology Analyzer)</p>	<p>6. The system should give the differential count as lymphocytes, mix population and neutrophils in percentage as well as absolute count.</p> <p>7. Rapid result turnaround time upto 60 samples-per-hour throughput.</p>	<p>The system should give the differential count as lymphocytes, eosinophils, basophils, monocytes and neutrophils in percentage as well as absolute count.</p> <p>7. Rapid result turnaround time minimum 60 samples-per-hour throughput.</p>

All bidder are requested to refer the attached Affidavit/Undertaking (Section XIX)

All other terms and conditions of the tender enquiry document shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/ CPP as corrigendum /amendments etc. if any, will be notified on this portal only, no separate advertisement will published in the news papers.

s/d

CGM, HSCC (I) Ltd.

For & on behalf of Director, AIIA, Sarita Vihar

SECTION - XIX
AFFIDAVIT/UNDERTAKING

IFB No.

We have read and understood the all instructions and all terms and conditions contained in the TE document.

We are fully agreed all the terms and conditions of TE document including SIT, SCC, amendment/ corrigendum, technical specification issued till opening of bid. In case, anywhere any conditional terms found in our bid, the same shall be treated as deleted/cancelled/ withdrawn from our bid. Whenever there is a conflict, the acceptance of all terms and conditions of TE document in the tender form/ bid form / affidavit shall prevail only.

We (manufacturer and /or manufacture authorised agent) shall jointly and severally liable to perform all contractual obligations under the agreement.

We (manufacturer and /or manufacture authorised agent) confirm that we do not stand deregistered/ banned/ blacklisted/ debarred by any Govt. Authorities in India.

We hereby confirm and certify that the prices offered by us in this tender is not higher than the prices we had offered to any other Govt. of India Organisation(s)/PSU(s) during the last one year and shall provide the justification for reasonableness of our offered price whenever asked during evaluation of our submitted bid.

We hereby certify that all information and documents submitted in this tender are true to the best of our knowledge and belief and that nothing material has been concealed/ misrepresented. We are solely responsible for its accuracy.

In case, at any stage, any of the information/ document is found to be false/ misrepresentation, we (manufacturer and /or manufacture authorised agent) shall be fully liable and the purchaser/HSCC shall have full right to reject my bid/ cancel the purchase order and / or stop payment / recover the liabilities/ loss if any, from our balance payment /EMD/ performance security etc. We are liable for any action as deemed fit by the purchaser/HSCC in addition to forfeiture of the earnest money/ performance security.

We are fully agreed all the terms and conditions of TE document including amendment/ corrigendum /technical specification issued till opening of bid. In case, anywhere any conditional terms found in our bid, the same shall be treated as deleted/cancelled/ withdrawn from our bid. Whenever there is a conflict, the acceptance of all terms and conditions of TE document in the tender form/ bid form / affidavit shall prevail only.

Signature:

Name:

Designation

Seal:

Note:

- **Original copy of Affidavit is to be submitted as instructed in the tender. The original document duly signed and stamped is to be scan & upload**
- **To be submitted on non-judicial stamp paper of Rs. 10/- duly certified by Public Notary**
