

**Amendment –II dated 10.09.2020**  
 Reply to Bidder queries (Pre bid held on 17.03.2020)  
 IFB no. HSCC/PUR/AIIA /Blood Bank/C dated 03.03.2020

All bidders are requested to note the following:-

**1. Item no2 –Elisa Reader and Washer:-**

Sr.	Tender technical Specification	AIIA decision
1	Should have 8-12 measuring channels and one reference channel.	Tender technical Specification prevail
2	Should have wave length range of 340-750 nm (6 standard filters- 340,405,450,492,590,540,620 nm with extra blank filter positions)	Tender technical Specification prevail
<b>Elisa Washer</b>		
3	Should have touch screen with two line LCD display.	Tender technical Specification prevail
4	Should have automatic calibration, well depth detection and last row detection features.	Tender technical Specification prevail

**2. Item no. 4, Bench top Di – electric tube sealer:-**

Sr.	Tender technical Specification	AIIA decision
1	Point no.8:- Should also have extended portable hand unit Sealing hand should be with coaxial of 1.5 -2.0 meter	Should not have extended portable hand unit.
2	Point no. 12:- System should run on both mains and battery (more than 10hrs. back up and charger)	Should run only on mains.

Name of Item	Last date & time sale/downloading of Tender document	Last date & for closing/submission date & time for receipt of tender	Date of opening of Techno – Commercial bids.
1. Platelet Incubator with Agitator 2. Elisa Reader & Washer 3. Blood Collection Monitor 4. Bench top Di – electric tube sealer 5. Cell Counter (Automated Haematology Analyzer)	22.09.2020, 14.00 hrs.	22.09.2020, 14.30 hrs.	22.09.2020, 14.35 hrs.

AIIA vide letter no. G -19004/5/2/2017/AIIA/Vol-V/1235 dated 08.09.2020, letter no. G -19004/5/2/2017/AIIA dated 16.06.2020 & email dated 19.05.2020 (copy is attached) conveyed their decision on the bidders queries, therefore, it is propose that the same may be upload on the website and CPP.

Submitted for approval please

CGM (Proc.)