

Amendment No. II Dated 16.08.2018
HSCC/PUR/CNCI/Kolkata/Medical Equipment/05 dt. 13.07.2018

Procurement of Medical Equipment CNCI 2nd Campus.

All bidders are requested to note the following:

Name of Item	Last date & time sale/downloading of Tender document	Last closing/submission date & time for receipt of tender	Date of opening of Techno – Commercial bids.	Earlier date of opening	Amendment Status for opening date
1. ENT Work Center with Endoscope system 2. Dental Chair with X-ray 3. Fully Automated Urine Analyzer 4. Fully Automated Electrolyte Analyzer 5. Multipara Monitors for Day Care 6. Multipara Monitors for Emergency 7. Multipara Monitors for Post Operative 8. Multipara Monitors for Pre Operative	28.08.2018, up to 13.00 hrs IST	28.08.2018, 14.00 hrs IST	28.08.2018, 14.30 hrs IST	16.08.2018	Date Extended

Item No.3, Fully Automated Urine Analyzer

Tender Specification	May Please Read As
4. The analyzer should be based on fluorescence flow cytometer for accurate measurement of urine parameters such as rbc, wbc, epithelial cells, cast and bacteria	The analyzer should be based on fluorescence flow cytometer/ Digital Imaging (Auto Particle Recognition)/flat flow cytometer for accurate measurement of urine parameters such as rbc, wbc, epithelial cells, cast and bacteria
5. The instrument should provide scatter grams and histograms for easy interpretation.	The instrument should provide scatter grams, histograms/ images imaging capacity in 800 frame for easy interpretation

Item No.4, Fully Automatic Electrolyte Analyzer

Tender Specification	May Please Read As				
System Analysis: Measurement of k,Na,Cl,Ca,TCO ₂ (by monometric method) working independently or at the same time (available for different combinations).: Technical Index	System Analysis: Na+, K+, Ca++,CL (ISE method) Technical Index: Deleted				
<table border="1" style="width: 100%;"> <tr> <td>CO₂</td> <td>≤3.0%</td> <td>≤3.0%</td> <td>6.5 ~ 50.0</td> </tr> </table>	CO ₂	≤3.0%	≤3.0%	6.5 ~ 50.0	
CO ₂	≤3.0%	≤3.0%	6.5 ~ 50.0		

All other terms and conditions of the tender enquiry document shall remain unchanged. Prospective bidders are advised to regularly visit HSCC website/ CPP as corrigendum /amendments etc. if any, will be notified on this portal only, no separate advertisement will published in the newspapers.

Sr. CGM-I, HSCC (I) Ltd
For & on behalf of Director CNCI, Kolkata