

**AMENDMENT-II****Ref.: IFB No. HSCC/SJH/Medical Equipment/2016/32 Dated 26.08.2017**

Sub.: Procurement of Medical Equipment for New Emergency Block & Super- Specialty Block at Safderjung Hospital, New Delhi.

Regarding Item No. 1,(Spine OT Table) Amended Specification has been received from Safdarjung Hospital for Spine OT Table. It is therefore extended the bid submission date from 22.09.2017 to 29.09.2017

**Amended Specification Specification for Spine Operation Table**

Specification :

Fully radiolucent modular table for spinal surgery. Should be versatile for lateral positioning, prone and supine of all types of cervical, dorsal and lumbar spine surgeries. Anterior and posterior approach by 180 / 360 rotating of patient on table top. C- Arm compatible offering free unhindered movement on the C-Arm. Battery Operated, Remote Controlled. The mechanism of operating table should be completely oil free motor driven. Should have interchangeable and replaceable table tops for specific surgery procedures.

Dimensions and limits:

1. Patient maximum safe working load : 200 kg or more.
2. Table height : 650 mm to 1000 mm & above
3. Position for Trendelenburg / Anti Trendelenburg
4. Lateral Tilt : 20 - 25 °

The table must come with the following equipment's and accessories :

- 1) Radiolucent Table Top
- 2) Centrally locking castors
- 3) Should have a corded / cordless remote
- 4) Patient Chest Pad – 2
- 5) Cervical traction Vector – 1
- 6) Universal Side Rail adaptor – 2
- 7) Cervical Management Base Unit – 1
- 8) Anterior Extension Positioner / Supine Positioner pad – 1
- 9) Skull Clamp – Aluminium – 1
- 10) Medical pillow for positioning
- 11) Universal Clamp / Socket – 2
- 12) Wilson Frame Cover– 1

- 13) Lateral Arm Board set with Pads – 1
- 14) Cross Arm support with Pads – 1
- 15) 3 piece Lateral Positioner / brace along with 2 side Rail Clamps / Socket– 1 Set
- 16) Mayfield / Doro adaptor for skull clamp with positioner
- 17) Large, Medium & Small sizes attachment with pad
- 18) Body positioners for prone position including lap supports
- 19) Safety straps to be provided
- 20) Equipment Cart – 1
- 21) Rates of all consumables if any to be freezed for 5 years, and to be quoted in tender document.
- 22) European CE & USFDA approved
- 23) Physical demonstration (if required)
- 24) Warranty 5 years + CMC 5 years
- 25) L1 will include CMC cost also.
- 26) Preventive maintenance every 3 months with log book entry
- 27) Company should have service centre in Delhi/ India and should provide service 24x7.
- 28) Spares parts should be available for 10 years
- 29) Should have installation at minimum two centres (Govt or Private) and should submit installation and performance certificate.

Amendment to be issued will be uploaded on websites [www.tenderwizard.com/HSCC](http://www.tenderwizard.com/HSCC) & [www.hsccltd.com](http://www.hsccltd.com).

All other tender terms and conditions remain unchanged.

Medical Superintendent  
VMMC& Safdarjung Hospital