

Amendment -III dated 16.11.2015

IFB: HSCC/KCGMC/Medical Equipment /2015-03 dated 22.10.2015

Item no.	Tender Specification	Tender Specification may please read as
Item no. 04 Abdominal Hysterectomy Set Main OT		Added: Instrument should be USFDA/European CE/ or Equivalent
Item no. 05 Vaginal hysterectomy set		Added: Instrument should be USFDA/European CE/ or Equivalent
Item no.7 Co2	<ol style="list-style-type: none"> 1. Should use treatment laser type Nd YAG. 2. Energy should be continuously variable from 0.5 to 10 mJ. 3. Should have a cone angle of 16° or greater. 4. Should have a wavelength of 1064nm. 5. Should have fundamental Q-switched mode. 6. Should have a spot size of 8 microns. 7. Should have pulse width of 4 nano-sec (typical). 8. Should have burst mode of 1,2,3 Pulses and each burst separation between pulse shall be of 25 micron- sec. 9. Should have repetition rate of better than 1 pulse per second. 10. Should have continuously adjustable YAG laser offset shift of up to 500Tm both anterior and posterior around aiming focal plane. 11. The aiming beam laser shall be diode of red colour. 12. Should have 3/5 step magnification for the laser delivery system. 13. Should have a focus of 250 microns posterior offset. 14. Should have an eye piece of at least 12.5X. 15. Should have an inter pupillary distance of 48 to 78 mm. 16. Should have slit width from at least 0 to 14mm continuously variable. 17. Should have slit length 0 to 14 mm continuously variable. 18. Should have slit rotation from 0 to 180°. 19. Should have red-free, cobalt blue filters. 20. Should operate from 200 to 240Vac, 50 Hz input supply. 21. Should be supplied with motorized table. 22. Should have safety certificate from a competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid. <p>Comprehensive warranty for 5 years and 5 years CMC/AMC after warranty. The cost of CMC/AMC must be quoted in the price bid.</p>	Deleted

Amendment -III dated 16.11.2015

IFB: HSCC/KCGMC/Medical Equipment /2015-03 dated 22.10.2015

Item no. 13 CONTACT MICROHYSTROSCOPE	-Cold Light fountain Halogen 150 Watts - 1 No Color Temp 3400k with 1 light outlet	Cold Light fountain LED 300 watt./ Xenon 300 watts-01 nos -Color Temp 3400k or more with 1 light outlet