

HSCC (I) Ltd

AMENDMENT –III Dated 14.10.2020

IFB No.: **HSCC/PUR/CNCI/KOLKATA/MEDICAL EQUIPMENT/108 Dt. 14.09.2020**

The reply to the pre bid queries are being uploaded as Annexure-III, as received from CNCI Kolkata for the above the referred IFB. It is proposed that the Bid sale, submission and opening date for the above the referred IFB remains **unchanged** (as mentioned in Amendment-II dt. 09.10.2020) as per details given below:

Sl. No.	Description	Existing Schedule
i.	Sale date of the tender	14.09.2020 (10:00 Hrs) to 16.10.2020 (upto 13:00 Hrs)
ii.	Closing date & time for receipt of tender	16.10.2020, 14:00 Hrs.
iii.	Time and date of Opening of Tenders	16.10.2020, 14:30 Hrs.

All other terms and conditions of the tender enquiry document shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/ CPPP website for corrigendum/ amendments etc. if any, as these will be notified on these portals only. No separate advertisement will published in the newspapers in this regard.

- sd -

CGM, HSCC (India) Limited
For and on behalf of Director, CNCI, Kolkata

REPLY TO PRE-BID QUERIES

Wide Bore 4D CT – Simulator (1 No.)

Sr. No.	TENDERED SPECIFICATIONS	AMENDMENT REQUESTED BY BIDDERS	TO BE NOW READ AS:
	<i>TECHNICAL CLARIFICATIONS:</i>		
3.	X-ray Tube		
3.6	The x-ray tube should have dual focal spot (please specify the size of each focal spot). The automatic selection of focal spot should be possible.	The x-ray tube should have single/dual focal spot (please specify the size of each focal spot).	The x-ray tube should have dual focal spot or equivalent (please specify the size of each focal spot). The automatic selection of focal spot should be possible.
6.	CT scanning parameters		
6.3	Maximum true scan field of view should be at least 60 cm or more.	Maximum true scan field of view should be at least 50 cm or more.	Maximum true scan field of view should be at least 50 cm or more.
	<i>COMMERCIAL CLARIFICATIONS:</i>		
8.2	Clause 8.2- Inspection, Testing and Quality Control: The Technical Specification and Quality Control Requirements incorporated in the contract shall specify what inspections and tests are to be carried out and, also, where and how they are to be conducted. If such inspections and tests are conducted in the premises of the supplier or its subcontractor(s), all reasonable facilities and assistance, including access to relevant drawings, design details and production data, shall be furnished by the supplier to the purchaser's inspector at no charge to the purchaser.	The Technical Specification and Quality Control Requirements incorporated in the contract shall specify what inspections and tests are to be carried out and, also, where and how they are to be conducted. If such inspections and tests are conducted in the premises of the supplier or its subcontractor(s), all reasonable facilities and assistance, including access to relevant drawings, design details and production data which are not confidential and proprietary information of the manufacturer, shall be furnished by the supplier to the purchaser's inspector at no charge to the purchaser.	No Changes. Tendered Terms & Condition prevails.
12.1. b.ii	GCC Clause 12.1.b.ii - Spare Parts: Immediately following such discontinuation, providing the Purchaser / Consignee free of cost, the designs, drawings, layouts and specifications of the spare parts, as and if requested by the Purchaser/Consignee.	We request to replace this clause with requirement for the supplier to provide an undertaking to provide spare parts for specific period of time (e.g. 10 years).	No Changes. Tendered Terms & Condition prevails.