HSCC/SES/CSSD/AIIMS/Raebareli/2020

Date : 05.10.2020

## All Bidders

## Amendment –III

Project: Supply, Installation, Testing & Commissioning of CSSD for Hospital Block, All India Institute of Medical Sciences (AIIMS), Raebareli.

IFB No. HSCC/SES/CSSD/AIIMS/Raebareli/2020 Dated 15.09.2020

This has reference to above IFB.

The following Amendment may be noted which shall be treated as part of the tender document and to be submitted duly signed & stamp along with tender.

The bid submission date is extended from **05.10.2020 to 20.10.2020** and bid security should be valid for 180 days from the date of original bid submission ie. from 30.09.2020.

All other terms & conditions remain unchanged.

Chief General Manager, HSCC (I) Ltd. As Executing Agency of Ministry of Health & Family Welfare

| 1. | 5. PLASMA/ H2O2 /LOW TEMPERATURE<br>STRILIZER (Double door)-120-150 L | PLASMA/ H2O2 /LOW<br>TEMPERATURE<br>STRILIZER (Double door)- |
|----|---|--|
|    | Page No. 8 of Technical Specification:                                | 120-150 L  |
|    | 1. Sterilizer process should be suitable for                          |  |
|    | sterilization of medical devices like flexible                        |  |
|    |   |  |
|    | endoscopes, rigid endoscopes- both single                             | Tender terms & conditions                                    |
|    | channel and also dual channel and non                                 | prevail  |
|    | lumen, metal, nonmetal heat & moisture                                |  |
|    | sensitive instruments, like defib paddles etc. The                    |  |
|    | sterilizer process must have maximum material                         |  |
|    | device manufacturers' recommendations from                            |  |
|    | major endoscopic equipment manufacturers.                             |  |
|    | Request for amendment as :  |  |
|    | Page No. 8 of Technical Specification:                                |  |
|    | 1. Sterilizer process should be able to process                       |  |
|    | materials sensitive to temperatures above 50                          |  |
|    | deg C and suitable for sterilization of medical                       |  |
|    | devices like rigid endoscopes- both single                            |  |
|    | channel and also dual channel and non                                 |  |
|    | lumen, metal, nonmetal heat & moisture                                |  |
|    | sensitive instruments, like defib paddles etc.                        |  |
|    | The sterilizer process must have maximum                              |  |
|    | material device manufacturers'  |  |
|    |   |  |
|    | recommendations from major endoscopic                                 |  |
|    | equipment manufacturers.  |  |
|    | We request you to kindly delete sterilization                         |  |
|    | for flexible endoscope as it requires high                            |  |
|    | level disinfection.   |  |
| 2. | Item No. 6 – Ultrasonic Cleaner                                       | Ultrasonic Cleaner   |
|    | Page No. 8 of Technical Specification:                                | Tender terms & conditions                                    |
|    | <b>h.</b> Ultrasonic cleaner should be BIS/European                   | prevail  |
|    | CE /US FDA certified  | prevan   |
|    | CE/05 TDA certified   |  |
|    | Request for amendment as :  |  |
|    | Page No. 8 of Technical Specification:                                |  |
|    | <b>h.</b> Ultrasonic cleaner should be either ISO-                    |  |
|    | 13485 with declaration of conformity as per US                        |  |
|    | FDA/European CE certified.  |  |
| 3. | Item No. 7 – Heat Sealing Machine                                     | Size: 585x395x195 or 600 x                                   |
| 5. | item i tot / i item beuning ivitemite                                 | 300 x 290 or 600 x 300 x 300                                 |
|    | Page No. 13 of Technical Specification:                               |  |
|    | Item No. 23 - Modular Sterilizing Basket                              |  |
|    | Size : 585x395x195  |  |
|    | Item No. 24 – Modular Sterilizing Basket                              |  |
|    | Size : 585x395x100  |  |
|    | 5120 . 30383738100  |  |
|    | Request for amendment as :  |  |
|    | Item No. 23 & 24 - Modular Sterilizing Basket                         |  |
|    | should be   |  |
|    |   |  |

|    | Size : 600 x 300 x 290 or 600 x 300 x 300<br>Kindly note that Specification of Modular<br>Sterilizing Basket asked in tender is as per<br>SPRI standard Size while Sterilizer has been<br>asked as per STU standard. We request you to<br>kindly amend the size as per STU.  |   |
|----|--|---|
| 4. | <b>Furniture Items</b><br>Request for amendment as :<br>All CSSD Furniture items should be accepted<br>with ±10% deviations.   | All CSSD Furniture items should be accepted with $\pm 10\%$ deviations. |
|    |  |   |
| 5. | Kindly amend the same.<br>The budget is not sufficient it may please be<br>enhanced as also in previous bid the budget was not<br>in accordance to the kind of turnkey work asked,<br>further the cost after this covid -19 pandemic also<br>increased at every point right from the<br>manufacturing till execution for additional safety<br>and limitation of facilities. Also you have added the<br>some of the equipment in this tender like instrument<br>trays, RO water, sterile containers with good<br>quantity and etc. this also need to enhance the<br>budget so please look into the matter and enhance<br>the budget of 1.5 more Crore excluding CMC from<br>current budget of 6 crore. i.e. approx. 7.5 Crore<br>excluding cmc. This shall help you to conclude this<br>project and enable you to avoid further retendering<br>of such a time bound project.  | Tender terms & conditions<br>prevail                                    |
| 6. | <b>TURNKEY WORKS</b><br>1. The area of CSSD in sq. mtrs. To be developed is not<br>specified in the tender. It should be clearly specified in<br>sq. mtrs. The CAD drawing should also be provided for<br>the area available to develop CSSD.<br>2. Also please note the kind of turnkey work asked in<br>the tender attracts a good cost for quality work to be<br>executed in accordance to the specification.so if any<br>point of time HSCC still not get convince even after<br>becoming L-1 and submission of justified orders by the<br>L-1 bidder. The award criteria must be like that the<br>award may be done for equipment supply without<br>turnkey work to the L-1 bidder to avoid delay in<br>execution of this tender as also our experience that we<br>have quoted and became L1 in equipment pricing and<br>total bid pricing too in previous tender with submission<br>of justified turnkey orders along with the bid which<br>made clear that we were the L-1 bidder for complete<br>overall bid. Further if any of the expert in evaluation<br>committee finds the turnkey job is not justified it may<br>be taken over by HSCC itself. | Tender terms & conditions<br>prevail                                    |

| 7.       In the BOQ Part-II CMC, the consumables including accessories may please be excluded from the CMC charges as the same has limited life and cannot be estimated in CMC.       Tender terms & conditions prevail         8.       Foreign currency quoting should be allowed for major equipment's i.e., Steam Sterilizer for 8 STU and 4 STU and Washer Disinfector and Plasma sterilizer and smaller equipment Table top sterilizer may also be at least allowed for foreign currency and the BOQ may please be revised accordingly. Also, please clarify about the exemptions certificates may please made available.       Tender terms & conditions prevail         9.       Time of completion of project is 4 months, which should be amended to 5 months from the date of site handover to vendor.       Tender terms & conditions prevail |
|---|
| charges as the same has limited life and cannot be<br>estimated in CMC.Tender terms & conditions<br>prevail8.Foreign currency quoting should be allowed for<br>major equipment's i.e., Steam Sterilizer for 8 STU<br>and 4 STU and Washer Disinfector and Plasma<br>sterilizer and smaller equipment Table top sterilizer<br>may also be at least allowed for foreign currency<br>and the BOQ may please be revised accordingly.<br>Also, please clarify about the exemptions certificates<br>may please made available.Tender terms & conditions<br>prevail9.Time of completion of project is 4 months, which<br>should be amended to 5 months from the date of site<br>handover to vendor.Tender terms & conditions<br>prevail  |
| estimated in CMC.Tender terms & conditions8.Foreign currency quoting should be allowed for<br>major equipment's i.e., Steam Sterilizer for 8 STU<br>and 4 STU and Washer Disinfector and Plasma<br>sterilizer and smaller equipment Table top sterilizer<br>may also be at least allowed for foreign currency<br>and the BOQ may please be revised accordingly.<br>Also, please clarify about the exemptions certificates<br>may please made available.Tender terms & conditions<br>prevail9.Time of completion of project is 4 months, which<br>should be amended to 5 months from the date of site<br>handover to vendor.Tender terms & conditions<br>prevail   |
| 8.       Foreign currency quoting should be allowed for<br>major equipment's i.e., Steam Sterilizer for 8 STU<br>and 4 STU and Washer Disinfector and Plasma<br>sterilizer and smaller equipment Table top sterilizer<br>may also be at least allowed for foreign currency<br>and the BOQ may please be revised accordingly.<br>Also, please clarify about the exemptions certificates<br>may please made available.       Tender terms & conditions         9.       Time of completion of project is 4 months, which<br>should be amended to 5 months from the date of site<br>handover to vendor.       Tender terms & conditions  |
| major equipment's i.e., Steam Sterilizer for 8 STU<br>and 4 STU and Washer Disinfector and Plasma<br>sterilizer and smaller equipment Table top sterilizer<br>may also be at least allowed for foreign currency<br>and the BOQ may please be revised accordingly.<br>Also, please clarify about the exemptions certificates<br>may please made available.prevail9.Time of completion of project is 4 months, which<br>should be amended to 5 months from the date of site<br>handover to vendor.Tender terms & conditions<br>prevail  |
| and 4 STU and Washer Disinfector and Plasma<br>sterilizer and smaller equipment Table top sterilizer<br>may also be at least allowed for foreign currency<br>and the BOQ may please be revised accordingly.<br>Also, please clarify about the exemptions certificates<br>may please made available.Image: Completion of project is 4 months, which<br>should be amended to 5 months from the date of site<br>handover to vendor.Tender terms & conditions<br>prevail  |
| sterilizer and smaller equipment Table top sterilizer<br>may also be at least allowed for foreign currency<br>and the BOQ may please be revised accordingly.<br>Also, please clarify about the exemptions certificates<br>may please made available.Tender terms & conditions<br>prevail9.Time of completion of project is 4 months, which<br>should be amended to 5 months from the date of site<br>handover to vendor.Tender terms & conditions<br>prevail  |
| may also be at least allowed for foreign currency<br>and the BOQ may please be revised accordingly.<br>Also, please clarify about the exemptions certificates<br>may please made available.Tender terms & conditions9.Time of completion of project is 4 months, which<br>should be amended to 5 months from the date of site<br>handover to vendor.Tender terms & conditions<br>prevail  |
| and the BOQ may please be revised accordingly.<br>Also, please clarify about the exemptions certificates<br>may please made available.Tender terms & conditions9.Time of completion of project is 4 months, which<br>should be amended to 5 months from the date of site<br>handover to vendor.Tender terms & conditions<br>prevail   |
| Also, please clarify about the exemptions certificates<br>may please made available.       Time of completion of project is 4 months, which<br>should be amended to 5 months from the date of site<br>handover to vendor.       Tender terms & conditions<br>prevail  |
| 9.Time of completion of project is 4 months, which<br>should be amended to 5 months from the date of site<br>handover to vendor.Tender terms & conditions<br>prevail  |
| should be amended to 5 months from the date of <b>site</b> prevail <b>handover to vendor.</b>   |
| handover to vendor.   |
|   |
|   |
| 10. <b>The payment terms for Indigenous/Locally supplied</b> 75% of the BOQ contract ra   |
| items part of BOQ may please be amended as on delivery of equipment   |
| follows: site after satisfactor   |
| $\square$ 80 % of the BOQ contract rates on delivery of inspection report and issue   |
| equipment's at site after inspection, passing and issue dispatch clearance on pro-rat   |
| of dispatch clearance. basis.   |
| $\Box$ 10 % of BOQ contract rates on satisfactory take over $-15\%$ of BOQ contract rates   |
| certificate by client after erection and installation, on satisfactory take-or  |
| testing and commissioning of equipment's. certificate by client af  |
| □ 10 % of BOQ contract rates after successful successful completion   |
| completion of trial run of 30 days from the date of erection and installation   |
| handover to the client. testing and commissioning   |
| CSSD.   |
| The payment terms for foreign currency items part<br>of BOO (for Steam Sterilizer (SSTU & 4STU) and after successful  |
| of bog (for Steam Stermizer (0510 & 4510) and   |
| Washer Disinfector, Thasha Stermizer) should be   |
|   |
| B of // of the Dog contract futes of the goods simpped  |
| shall be paid through irrevocable, non-transferable   |
| Letter of Credit (LC) opened in favour of the supplier  |
| in a bank in his country. $\Box$ 10 % $\Box$ DOO  |
| $\Box$ 10 % of BOQ contract rates of the goods shipped  |
| shall be paid through irrevocable, non-transferable   |
| Letter of Credit (LC) opened in favour of the supplier  |
| in a bank in his country after satisfactory take over   |
| certificate by client after erection and installation,  |
| testing and commissioning of equipment's.   |
| $\Box$ 10 % of BOQ contract rates after successful  |
| completion of trial run of 30 days from the date of   |
| handover to the client shall be paid through irrevocable,   |
| non-transferable Letter of Credit (LC) opened in favour<br>of the supplier in a bank in his country.  |
| of the supplier in a bank in his country.   |
| 11.       The payment clause mentioned in GCC of the tender       Tender terms & conditions   |
| document may please be amended as per the nature of prevail   |
| the project.  |
|   |
| Payment of duties and CDC exemption are not   |
| specified in the tender. Kindly clarify whether CDC   |

|   | 1  |
|---|--|
| exemption will be provided or not and custom duty will<br>be reimbursed to us or directly paid by you.as in the bid<br>if the BOQ shall not be revised the clarity to pay in INR<br>apart from the foreign bidder shall not be clear. As also<br>we raised issues in previous tender BOQ and<br>amendments were made up to some extent. |  |
| DOUBLE DOOR WASHER DISINFECTOR 280-<br>350 Litre (12-15 DIN Trays) WITH<br>ACCESSORIES  | DOUBLE DOOR WASHER<br>DISINFECTOR 280-350<br>Litre (12-15 DIN Trays)<br>WITH ACCESSORIES   |
| <b>Chamber Capacity:</b> Chamber capacity: Operational Volume should be able to process minimum 12-15 DIN trays (Approx480X250X50) with 4 levels in single process. Should supply minimum 12 Nos of standard Stainless Steel DIN trays.   | Chamber capacity: Operational<br>Volume should be able to<br>process minimum 12-15 DIN<br>trays<br>(Approx480X250X50)with 5/6<br>levels in single process.   |
| Request for amendment as :  |  |
| DOUBLE DOOR WASHER DISINFECTOR 300-<br>350 Litre (minimum 12 DIN Trays) WITH<br>ACCESSORIES   |  |
| Chamber Capacity: Chamber capacity:<br>Operational Volume should be able to process<br>minimum 12 DIN trays (Approx480X250X50)<br>with 6 levels in single process. Should supply<br>minimum 12 Nos of standard Stainless Steel DIN<br>trays.  |  |
| The washer disinfector shall be supplied with<br>universal rack, 4 level racks for instrument tray, full<br>size instrument tray as well as stop valves, anti-<br>suction device and plastic water trap.  | The washer disinfector shall be<br>supplied withuniversal rack,<br>5/6 level racks or more for<br>instrument tray,full size<br>instrument tray as well as stop<br>valves,anti-suction device and   |
| The washer disinfector shall be supplied with<br>universal rack, <b>6 level racks</b> for instrument tray, full<br>size instrument tray as well as stop valves, anti-   | plastic water trap.  |
| PLASMA/LOW TEMPERATURE STRILIZER<br>(Double door)-120-150L  | PLASMA/LOW<br>TEMPERATURE<br>STRILIZER (Double door)-<br>120-150L<br>Sterilizer process should be<br>suitable for sterilization of<br>medical devices like flexible  |
|   | <ul> <li>if the BOQ shall not be revised the clarity to pay in INR apart from the foreign bidder shall not be clear. As also we raised issues in previous tender BOQ and amendments were made up to some extent.</li> <li>DOUBLE DOOR WASHER DISINFECTOR 280-350 Litre (12-15 DIN Trays) WITH ACCESSORIES</li> <li>Chamber Capacity: Chamber capacity: Operational Volume should be able to process minimum 12-15 DIN trays (Approx480X250X50) with 4 levels in single process. Should supply minimum 12 Nos of standard Stainless Steel DIN trays.</li> <li>Request for amendment as :</li> <li>DOUBLE DOOR WASHER DISINFECTOR 300-350 Litre (minimum 12 DIN Trays) WITH ACCESSORIES</li> <li>Chamber Capacity: Chamber capacity: Operational Volume should be able to process minimum 12 DIN Trays) WITH ACCESSORIES</li> <li>Chamber Capacity: Chamber capacity: Operational Volume should be able to process minimum 12 DIN trays (Approx480X250X50) with 6 levels in single process. Should supply minimum 12 Nos of standard Stainless Steel DIN trays.</li> <li>The washer disinfector shall be supplied with universal rack, 4 level racks for instrument tray, full size instrument tray as well as stop valves, antisuction device and plastic water trap.</li> <li>Request for amendment as :</li> <li>The washer disinfector shall be supplied with universal rack, 6 level racks for instrument tray, full size instrument tray as well as stop valves, antisuction device and plastic water trap.</li> <li>PLASMA/LOW TEMPERATURE STRILIZER</li> </ul> |

|     | Sterilizer process should be suitable for sterilization<br>of medical devices like flexible endoscopes, rigid<br>endoscopes- both single channel and also dual<br>channel and non-lumen, metal, non-metal heat &<br>moisture sensitive instruments, like defib paddles etc.<br>The sterilizer process must have maximum<br>Request for amendment as :<br><b>LOW TEMPERATURE PLASMA STRILIZER</b><br>(Double door)-120-150L<br>Sterilizer process should be suitable for<br>sterilization of medical devices like flexible<br>endoscopes, rigid endoscopes- both single channel<br>and also dual channel and non-lumen, metal, non-<br>metal heat & moisture sensitive instruments, like<br>defib paddles etc. The sterilizer process must have<br>maximum material device manufacturers'<br>recommendations for major endoscopic equipment<br>manufacturers/backed by clinical testing. | endoscopes, rigid endoscopes-<br>both single channel and also<br>dual channel and non-lumen,<br>metal, non-metal heat &<br>moisture sensitive instruments,<br>like defib paddles etc. The<br>sterilizer process must have<br>maximum material device<br>manufacturers'<br>recommendations for major<br>endoscopic equipment<br>manufacturers/backed by<br>clinical testing. |
|-----|---|---|
| 15. | ULTRASONIC CLEANER (40 L)<br>Ultrasonic cleaner should be either ISO-13485 with<br>declaration of conformity as per BIS/US<br>FDA/European CE certified with four digit notified<br>body number.<br>Request for amendment as :  | Tender terms & conditions<br>prevail  |
|     | Ultrasonic cleaner should be either <b>EN ISO-13485</b><br><b>certified/US FDA</b> /European CE certified with four<br>digit notified body number.  |   |
| 16. | ALL FURNITURE ITEMS         Dimensions of furniture items.         Request for amendment as :         ±10% variation in dimensions of all furniture items   | All CSSD Furniture items should be accepted with ±10% deviations.   |
| 17. | should be given.  | CONTROL & PACKING   |
|     | SHELVES FOR CLEAN AREA<br>The worktop should be made of a robust wood-based<br>core material, surfaced with plastic laminate in a soft<br>beige colour that reduces reflection of light from the<br>surface. All edges should be smooth. The extended   | TABLEWITHTWOSHELVESFORCLEANAREAThe worktop should be madeof a robust wood-based core  |

| -   |  |   |
|-----|--|---|
|     | width of the worktop should be designed to facilitate<br>thorough inspection of instrument trays and allow the | material, surfaced with plastic<br>laminate in a soft beige colour                            |
|     | use of large wrapping material.  | that reduces reflection of light  |
|     | use of large wrapping material.  | from the surface. All edges   |
|     | Request for amendment as :   | should be smooth/The worktop<br>should be made of SS (304) All<br>edges should be smooth. The |
|     | The worktop should be made of SS (304) All edges   | extended width of the worktop   |
|     | should be smooth. The extended width of the worktop  | should be designed to facilitate  |
|     | should be designed to facilitate thorough inspection of  | thorough inspection of  |
|     | instrument trays and allow the use of large wrapping   | instrument trays and allow the  |
|     | material.  | use of large wrapping material.   |
|     |  |   |
| 18. | Should have double workspace. One workplace table  | Should have double work   |
|     | should have 700 mm wide worktop and other workplace should have 1400 mm worktop.                               | space table with one work space of 700 mm wide worktop  |
|     | workprace should have 1400 min worktop.  | and another workspace also of   |
|     |  | 700 mm wide worktop i.e. total  |
|     | Request for amendment as :   | workspace should be 1400 mm   |
|     |  | total workplace should be 1400  |
|     | Should have double work space table with one work  | mm.   |
|     | space of 700 mm wide worktop and another   |   |
|     | workspace also of 700 mm wide worktop i.e. total   |   |
|     | workspace should be 1400 mm total workplace  |   |
|     | should be 1400 mm.   |   |
|     |  |   |

All other terms & conditions remain unchanged.

Sr. Chief General Manager -I, HSCC (I) Ltd.