

Amendment-III**Ref.: IFB No. HSCC/SJH/Medical Equipment/2015/11 Dated 26.12.2015 (Item 1-4).**

Sub.: Procurement of Medical Equipment for New Emergency Block & Super-Specialty Block at Safderjung Hospital, New Delhi.

The amendments are as follows:

Amendments		
	Ten der No. HSCC/SJH/Med.Eqpt./2015/11 dt. 26.12.2015	
Sr.No.	Tender Specification	Amended Specifications
	Haemodialysis Machine with SLED	
	Point No. 1.11. Wide dialysate flow rates options (100-1000 ml/mt with increments of 100 ml/mt	Point No. 1.11. Wide dialysate flow rates options (100-800 ml/mt or more) with increments of 100 ml/mt
	Point no. 3.5: Suitable servo controlled stabilizer/CVT/UPS should be supplied, if required.	Deleted
	<u>Haemodialysis Machine (Regular):</u>	
	Point no. 3.e: Suitable servo controlled stabilizer/CVT/UPS should be supplied, if required.	Deleted
	<u>Reverse Osmosis Water Treatment Plant for Dialysis</u>	
	Point no. 2.9: The bidder must submit at least three performance certificates from government hospitals/institutions where a similar RO Plant has been installed.	Point no. 2.9: The bidder must submit at least three performance certificates from government hospitals/institutions where a similar RO Plant has been installed for dialysis purpose.

<u>Point no. 2.10:</u> Tenders should be quoted with full quality assurance certificate (EC Certificate)	<u>Point no. 2.10:</u> Tenders should be quoted with European CE or USFDA certification
---	--

The bid submission date is now extended from 03.2.2016 to 15.2.2016.

Amendment issued will be uploaded on websites www.tenderwizard.com/HSCC & www.hsccltd.com.

All other tender terms and conditions remain unchanged.

**Medical Superintendent
Safderjung Hospital & VMMC,
New Delhi.**