All Bidders

Amendment -III

Date:31.07.2019

Project: Supply, Installation, Testing & Commissioning of Modular Operation Theatres for Hospital Block at All India Institute of Medical Sciences (AIIMS), Mangalagiri, Distt. Guntur.

IFB No. HSCC/SES/MOT/AIIMS/Guntur/2019 Date: 17.06.2019

This has reference to above IFB.

The following Amendment may be noted which shall be treated as part of the tender document and to be submitted duly signed & stamp along with tender.

Sr. No.	Bidders' Queries	Amended as
1	Wall & Ceiling System (SMS)	
	It should be BIS/European CE certified with notified 4 digit body no./UL Listed	The certificate of the product should be BIS/US FDA
	You are requested to kindly delete European CE Certified with notified 4 digit body no/UL Listed and	European CE Declaration of conformity
	Add European CE Declaration of conformity for Wall and Ceiling System	
2	OT Light-LED	
	It should be compliant with relevant European CE with 4 digit notified body no./USFDA/ETL/UL Listed	European CE/UL Listed /ETL & USFDA Certified.
	Or Declaration of conformity for quoted Model with ISO-13485	
	Kindly amend the same to relevant European CE/UL Listed /ETL & USFDA Certified.	
	Certification should be must for quoted OT Light for safety and quality standard of OT Light.	
3	HD Camera System	
	It should be USFDA/European Certified	
	This clause is only applicable for the OT Lights and not for the Camera installed with it. You are therefore requested to kindly delete this line from tender technical specifications.	Tender terms and conditions prevail

4		
-	WALL PANELING SYSTEM & CEILING PANELING SYSTEM	
	Wall and ceiling system, Sub frame and Sealing Gaskets must be of same manufacturer	
	We request you to include principal company as well and amend as "Wall and ceiling system, Sub frame and Sealing Gaskets must be of same manufacturer/ principal company"	Tender terms and conditions prevail
5	ii) Wall panels:	
	f. Reaction to fire class 1 norm	
	We request you to amend as "Reaction to fire class 1 norm or equivalent" since class 1 norms are specific to certain countries	Tender terms and conditions prevail
6	Amendment II. point 44	
	The certificate of the product should be BIS/European CE Certified with notified 4 digit body no./UL listed	
	Kindly delete this point as this is valid for only medical equipment Wall panels are not classified as medical equipment	The certificate of the product should be BIS/ European CE/US FDA
7	18 HD LED FLAT PANEL MEDICAL GRADE MONITORd)	
	-HD Flat Panel Monitor should be done according to IP 54 regulations.	
	-ESG Safety Glass Cover is must to protect the Monitor Screen from breakage and for additional Safety of Patient.	
	-Video Connectors on the back side of Monitor is to be hide by Cover of Aluminum from the Back side of Monitor.	
	Request for Highlighted points are specific to a particular manufacturer hence request you to delete these points. This is restricting participation of other companies/ suppliers. Moreover, these points have no added advantage over the functionality and the utility of the equipment	Tender terms and conditions prevail
8	18B RECORDING SYSTEM	
	Recording system should be integrated streaming with streaming.	

	Request you to provide more clarity on this point since the sentence is vague.	Tender terms and conditions prevail
	h) ESG Safety Glass Cover is must to protect the Monitor Screen from breakage and for Additional Safety of Patient. n) Recording system should be done according to IP 54 regulations.	
	Request for Highlighted points are specific to a particular manufacturer hence request you to delete these points. This is restricting participation of other companies/ suppliers. Moreover, these points have no added advantage over the functionality and the utility of the equipment	Tender terms and conditions prevail
9	18C MONITOR — DIGITAL DISPLAY (PACS/HMIS)	
	HD Flat Panel Monitor should be done according to IP 54 regulations.	
	ESG Safety Glass Cover shall be used to protect the Monitor Screen from breakage and for additional Safety of Patient.	
	i) Video Connectors on the back side of Monitor is to be hide by Cover of Aluminium from the	
	Request for Highlighted points are specific to a particular manufacturer hence request you to delete these points. This is restricting participation of other companies/ suppliers. Moreover, these points have no added advantage over the functionality and the utility of the equipment	Tender terms and conditions prevail
10	19. PENDANTS FOR ANESTHETIST AND SURGEON m) Shelf of Pendant should be made of High- quality Bio- Clean Material	
	Request you clarify the meaning of Bio- Clean material. This point since it is specific to a particular company. This is restricting participation of other companies/suppliers. Pendants manufactured by most companies are cleanable with hospital disinfectants	Tender terms and conditions prevail
11	n) Pendant should have Anti- microbial Coating.	
	Request you to delete this point since it is specific to certain company. This is restricting participation of other companies/suppliers.	Anti- microbial Coating/Powder coating

	Pendants manufactured by most companies are cleanable with hospital disinfectants	
12	The bid submission date is extended to 31.07.2019	
	Request you to extend the deadline by another week to 5 th August 2019	Extension of date of submission is mentioned below.
13	Request for deletion of Manufacturer Authorisation for Monitor Digital Display (PACS/HMIS)	Deleted
14	Extension of tender submission date from 31.07.2019 to 05.08.2019	Extension of date of submission is mentioned below

The bid submission date is extended from 31.07.2019 to 05.08.2019 and bid security should be valid for 180 days from the date of original bid submission ie. From 17.07.2019.

All other terms & conditions remain unchanged.

Sr. Chief General Manager -I , HSCC (I) Ltd. For & on behalf of Director, AIIMS, Mangalagiri, Guntur