

HSCC (I) Ltd

Amendment –III Dated 31.12.2019

Tender Ref : HSCC/AIIMS-RAEBARELI/Hospital/2/2019 Dated : 28.11.2019

Department: Endocrinology

Item No.1: DEXA Scanner

Spec. Sr. No.	TENDER SPECIFICATION	READ AS
A1	Isocentric or Linear fan beam acquisition	Narrow fan beam / Linear fan beam
2	Energy switching technique - Energy switching should be mentioned with supporting documents	Energy switching technique or equivalent
B 6	Radiation Dose : Scatter dose less 0.1 micro Gy/ hour at 1 meter	scatter dose less than 0.1 micro Gy/hour at 2 meter or less
C5	Greater than 1% in vivo precision	Less or equal to 1% in vivo precision
C23	Context sensitive help software	Deleted
	New point Added :	Should be US-FDA or European CE with 4 Digit notified body number. or BIS

The Bid submission date has been extended for all the items tendered **from 31.12.2019 to 15.01.2020** except for items Sr. No. N1, N3, N4, N5, N6 (A) and N6 (B) as their bid submission date is 07.01.2020

All other terms and conditions of the tender enquiry documents shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/CPMP Website for the Corrigendum/amendments etc. if any, as these will be notified on these portals only. No separate advertisement will be published in the newspaper in this regard.

CGM (Projects)
HSCC (I) Ltd.