

HSCC (I) Ltd

Amendment –III Dated, 13.03.2020

Tender Ref : HSCC/SJH/Medical Equipment /2020/44 Dated : 29.01.2020

Subject: Extension of Bid submission Date.

The reply of the pre bid queries of following item has been received. Details are as follows:

Tender Ref: HSCC/SJH/Med.Eqpt./2020/44 Dated 29.01.2020

Item No. 03, Ultrasound Cum Echo Color Doppler

Sr. No.	Tendered Specification	To be read as
08	System must have fast startup to scanning in less than 30 Seconds from off condition, for use in critical and emergency situations.	System must have startup to scanning in less than 50 Secs from off condition, for use in critical and emergency situations.
09	System should support transducer technologies like phased array , convex linear	System should support Convex Linear and Hockey Stick shape probes.

The bid submission date for Item No.1,3,4, is extended from 13.03.2020 to **20.03.2020**.

All other terms and conditions of the tender enquiry documents including Amendments issued so far shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website and CPPP Website for the Corrigendum/amendments etc. if any, as these will be notified on these portals only. No separate advertisement will published in the newspaper in this regard

**CGM (Proc)
HSCC I LTD**