

Amendment no 03 Dated 17.06.2016
05.05.2016 IFB no. HSCC/KCGMC/Medical Equipment/2016/MC/01 dated 05.05.2016

All bidders are requested note the following:

1.

Item no.	Last date & time on line sale of Tender enquiry document	Last closing date & time for receipt of e-tender	Last Time and date of opening of Techno – Commercial bids.	Earlier date of tender opening
2 to 7	30.06.2016, upto 12.30 hrs IST	30.06.2016, 13.00 hrs IST	30.06.2016, 13.30 hrs IST	17.06.2016

2. The technical specification of **item no. 02 – Micro Tome, rotary** may please read per given below:

1. High precision machine suitable for both delicate as well as hard tissue sectioning.
2. Manual type with specimen clamps, user selectable coarse feed wheel turn direction. retraction on/off function and blade holder for disposable blades.
3. Section thickness setting 0.5-60 μ m with settings in 0.5, 1, 2, 5 increments at different levels.
4. Section increment:
 - from 0 to 2 μ m in 0.5 μ m steps.
 - from 2 to 10 μ m in 1 μ m steps.
 - from 10 to 20 μ m in 2 μ m steps.
 - from 20 to 60 μ m in 5 μ m steps.
5. Specimen advance 28 mm or more.
6. Vertical stroke 60 mm or more.
7. Provision of step trimming.
8. Adjustable Specimen Clamp at least 50 x45 mm with orientation in x y axis.
9. Single disposable blade holder for accommodating both high and low profile blades.
10. Lateral Coarse feed.
11. Integrated removable section waste tray.
12. Spare Low and High Profile Blades in dispenser pack of 50 blades: 6 packets each.
13. The equipment should be USA- FDA/European- CE approved

Accessories:

- Orientable specimen clamp in x, y axis
- Disposable blade holder which accommodates both high and low profile blades Spare Low and High Profile Blades in dispenser pack of 50 blades: 6 packets each

As per SCC clause no. 9, the validity of bid security (EMD) shall be considered from the original date of submission of bid/ tender in case tender opening date extended.

All other terms and conditions of the tender enquiry document shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/ CPP as corrigendum /amendments etc. if any, will be notified on this portal only, no separate advertisement will published in the news papers