

AMENDMENT – III**Dated: 12.01.2016**

**INVITATION FOR BIDS FOR SUPPLY, INSTALLATION, TESTING & COMMISSIONING OF
MEDICAL EQUIPMENTS FOR 200 BEDDED
NEPAL- BHARAT MAITRI EMERGENCY AND TRAUMA CENTRE (NBMETC),
KATHMANDU, NEPAL**

BID REFERENCE: IFB No. HSCC/PUR/NEPAL/2015-16/1 Dated 10.11.15

Bidder is requested to note the following amendment:

Bid sale, submission and opening date for item no. 2,3,5,6,7 of the aforesaid bid is amended as per following:

S. No.	Details	Revised schedule
1	Last date and time for the sale of bid	25.01.2016 by 14.00 hrs.
2	Last date & time for submission of bid	25.01.2016 by 15.00 hrs.
3	Techno-commercial opening of bids	25.01.2016 by 15.30 hrs.
4	Venue of bid opening	HSCC (India) Ltd., E--6A, Block-E, Sector-1, NOIDA

Section – IV Technical Specifications

Item No. 3 Fiber Operative Bronchoscope

Description	Amended as
The working length should be 300 to 550mm	The working length should be 300 to 600mm
Image grabbing and editing software	Different manufacturer would be acceptable for these items
Camera with 3CCD HD monitor, Medical Grade	Camera with Medical Grade HD monitor
Equipment should include facility for recording images and video on DVD recorder and computer and should be equipped with DVD recorder, computer and compatible software to record and edit videos and high resolution still images for patient report.	Different manufacturer would be acceptable for these items
Equipment should have facility for simultaneous recording on DVD and Computer	Different manufacturer would be acceptable for these items
Computer requirement: Core 2 duo, at least 2G b RAM, Minimum 250 GB hard disk, DVD Recorder, at least 17 inch LCD monitor, mouse & Keyboard	Different manufacturer would be acceptable for these items
High resolution color laser printer (at lest 1200 dpi)	Different manufacturer would be acceptable for these items
Digital video and photo capture card (fire-wire port)	Different manufacturer would be acceptable for these items

Item No. 5 – ICU Ventilator

Tender Specification	Amended As
Technical Specification:- 4. Should have inbuilt facility to upgrade with EtcO2	Deleted
Should have following settings (ventilating settings and ranges): 2. Inspiratory peak flow 5-200LPM (Compensated) 14. Inspiratory & expiratory pause, programmable sign	2. Inspiratory peak flow 5-160LPM (Compensated) 14. Inspiratory & expiratory pause
Certification and Accessories 1. Quality Certification: Valid CE/BIS/US FDA	1. Quality Certification: Valid CE(Europe)/US FDA

Item No. -7 DIGITAL MOBILE X-RAY UNITS

Tender Specification	Amended As
Technical Specification:- I. Warranty/ Guarantee: Vendor should offer onsite warranty for 5 years for the whole unit including x-ray tube, detector and all other accessories, batteries and consumables required to run this unit. After expiry of warranty, CMC should be for five years which includes x-ray tube, detector all other accessories, batteries and consumables required to run this unit.	I. Warranty/ Guarantee: Vendor should offer onsite warranty for three years for the whole unit including x-ray tube, detector and all other accessories, batteries and consumables required to run this unit. After expiry of warranty, CMC should be for three years which includes x-ray tube, detector all other accessories, batteries and consumables required to run this unit.

All other terms and conditions of the bid document shall remain unchanged.

S/d
 Joint Secretary –DPA (III)
 Ministry of External Affairs,
 New Delhi