

Amendment no. 04 Dated 01.04.19
HSCC/PUR/AIIA /Blood Bank/B dated 01.02.2019

Procurement of Medical Equipment AIIA Sarita Vihar:-

All bidders are requested to note the following:

Item no 02, REFRIGERATED CENTRIFUGE.

Sr no.	Tender Technical Specification	May please read as
3.	Para no. 22, The firm should have a track record/service/performance record certificate from the well known Govt. Institutions (AIIMS - New Delhi, Safdarjung hospital - New delhi, PGIMER, Chandigarh) of having supplied such equipment within last five years.	The firm should have a track record/service/performance record certificate from the any Govt. Institutions/Hospital / College having supplied such equipment within last five years.
4.	Para no. 10, Stainless steel chamber: Easy to clean and corrosion resistant	Stainless steel chamber: Easy to clean and corrosion resistant with provision of both drain and condensed water collection container.

Item No. 06, Laminar Air Flow Bench (Bio Safety Cabinet)

Sr no.	Tender Technical Specification	May please read as
1	Para no .3, HEPA/ULPA filters on inflow as well as exhaust with an efficacy of 99.999% for equal or more than 0.3u size particles (DOP test certificate to be produced).	HEPA/ULPA filters on inflow as well as exhaust with an efficacy of 99.995% for equal or more than 0.3u size particles (DOP test certificate to be produced).
2.	Para no. 8, Sliding front window, electrically operated	Sliding front window Manually /Electrical operated
3.	Para no.10, Work table should be fitted with white lamp for local illumination, self operating UV germicidal light (15 W) on window opening, static pressure manometer and exhaust system.	Biosafety Cabinet should be fitted with white lamp for local illumination, self operating UV germicidal light (15 W) on window opening, static pressure manometer and exhaust system.
4.	Para no. 12 B (iii), Cabinet temperature in degree Celsius.	Cabinet temperature in degree Celsius (External /inbuilt)
5.	Para no. 23, US FDA / European CE certified	US FDA / European CE / Bureau of Indian Standards

All other terms and conditions of the tender enquiry document shall remain unchanged. Prospective bidders are advised to regularly visit HSCC website/ CPP as corrigendum /amendments etc. if any, will be notified on this portal only, no separate advertisement will published in the news papers.

Sr. CGM-I, HSCC (I) Ltd
For & on behalf of AIIA Sarita Vihar

