

All Bidders**Amendment –IV**

Project: Supply, Installation, Testing & Commissioning of CSSD for Super Speciality Block at Indira Gandhi Medical College, Shimla, H.P.

IFB No. HSCC/SES/CSSD/SSB/Shimla/2019 dated 21.02.2019

This has reference to above IFB.

The following Amendment may be noted which shall be treated as part of the tenderdocument and to be submitted duly signed & stamp along with tender.

Sr. No.	Bidders' Queries	Reply
1	<p><u>Amendment –II</u> Sr. No. 1</p> <p><u>Horizontal Steam Sterilizer Double door 550-600L or more</u> & <u>High Speed Sterilizer- Double door 250L with accessories.</u></p> <p>Control system should have touch sensitive ,5-7 inches or more color display interface at operator loading side while it should have normal interface at unloading side.</p> <p>Request It should be brought to original tender specification as “ Control system should have touch sensitive , 7 inches or more colour display interface at operator loading side while it should have normal interface at unloading side.</p>	Amendment –II shall prevail.
2	<p><u>Amendment -II</u> Sr. No. 2</p> <p>Chamber Capacity: Operational volume should be able to process minimum 12 DIN trays (Approx 480x250x50) in single process. The chamber should be made of SS 316L quality with electro polished washed surfaces.</p> <p><u>Request :</u> We would request that minimum capacity as asked in original tender specification should be kept ie. 300 ltrs or more , it can be 12 DIN trays or 15 DIN Trays , but minimum capacity should be defined.</p> <p>It should be brought to original tender specs with the</p>	Amendment –II shall prevail

	chamber should be made of SS 316 L with electro polished washed surfaces.	
3.	<p>Amendment No. 2 Sr. No. 3 and Sr. No. 44 page no. 8 4 -6 Levels racks Request There is gross difference in 4 levels and 6 levels, one uniformity should be there we would request that it should be 5 to 6 level of racks with 300 ltrs or more operational volume.</p>	4-6 Level racks Amendment –II shall prevail
4.	<p>Amendment No. 2 Sr. No. 4 (ii) 100-120 L Request We would request that it should be mentioned as “usable volume “ of 100 – 120 L, usable volume is missing, whereas it was asked in original tender specs.</p>	100 – 120 L usable volume
5.	<p><u>Amendment No. 2</u> <u>Sr. No. 15 and Sr. No. 34 : RJ 45 as per current standards</u> What is the point of adding RJ 45 , as RJ 45 is for the Ethernet connectivity, we would request that it should be RJ 45 or equivalent port for Ethernet connectivity, if at all it has to be added.</p>	RJ 45 or equivalent port for Ethernet connectivity.
6.	<p><u>Amendment No. 2</u> Sr. No. 44 (d) Chamber Capacity : Operational volume should be able to process minimum 12 DIN trays (Approx. 480 x 250 x 50) in single process. The chamber should be made of SS 316 L quality with electro polished washed surfaces. Request We would request that minimum capacity as asked in original tender specification should be kept ie. 300 ltrs or more , it can be 12 DIN trays or 15 DIN Trays , but minimum capacity should be defined. It should be brought to original tender specs with the chamber should be made of SS 316 L with electro polished washed surfaces.</p>	Amendment No. 2 shall prevail.

The bid submission date is extended from **29.03.2019 to 04.04.2019** and bid security should be valid for 180 days from the date of original bid submission ie. from 19.03.2019.

All other terms & conditions remain unchanged.

Sr. Chief General Manager -I, HSCC (I) Ltd.
For and behalf of Principal, IGMC, Shimla