

HSCC India Ltd.

AMENDMENT – IV

Dated: - 22.05.2017

Subject: Amendment to the tender Enquiry Document.

Tender Ref: HSCC/KCGMC/Medical Equipment/2015-06/13 dated 11.04.17

Bidders are requested to note the technical Amendment in the specification of the following equipment:

Item No. 7 ENT Treatment Unit

Point No.	Specification of Tender	Amended as
10	Ear Irrigation System <ul style="list-style-type: none">• Independently from water grid operable with 4.5 L Canisters	Ear Irrigation System <ul style="list-style-type: none">• Independently from water grid operable with one 4.5 L Stainless Steel Canisters or two 2.5 L Stainless Steel Canisters
11	3CCD Camera with Image memory	The specification are upgraded as 3CCD HD Camera with resolution 1920x1080 p, with image management system for recording of HD video and HD images in same resolution
	Specification Added	UPS backup for the whole unit for at least 2 hours to be supplied by the bidder
	Specification Added	RO system to be provided for the water input to the unit by the bidder. All the plumbing work for the same to be done by the bidder.

Item No. 8 Temporal Bone Lab: No change in the specification

Bid sale, submission and opening date for item no. 4,6,7,8,9,10 has been extended as per details given in Table -1:

Table -1

Sl. No.	Description	Revised Schedule
i.	Sale date of the tender	29.05.17, 3.00 P.M.
ii.	Closing date & time for receipt of tender	29.05.17, 3.30 P.M.
iii.	Time and date of Opening of Tenders	29.05.17, 4.00 P.M.

Bid sale, submission and opening date for item no. 5 has been extended as per details given in Table -2:

Table -2

Sl. No.	Description	Revised Schedule
i.	Sale date of the tender	23.05.17, 3.00 P.M.
ii.	Closing date & time for receipt of tender	23.05.17, 3.30 P.M.
iii.	Time and date of Opening of Tenders	23.05.17, 4.00 P.M.

s/d
CGM, HSCC India Limited
For and on behalf of DGMR, Panchkula

