

HSCC (I) Ltd

HSCC (I) Ltd

Amendment –IV Dated 06.02.2020

Tender Ref : HSCC/AIIMS-RAEBARELI/Hospital/3/2019 Dated : 23.12.2019

Technical Amendments :

Item No. 3 E: ICU Ventilator

Spec. Sr. No.	TENDER SPECIFICATION	READ AS
Point No 21	Oxygen sensor should be paramagnetic /ultrasonic and covered under warranty and cmc and will be supplied free of cost during warranty and cmc period.	Oxygen sensor should be paramagnetic /ultrasonic/galvanic and covered under warranty and cmc and will be supplied free of cost during warranty and cmc period.

The bid submission date extend for all items from 06.02.2020 to 13.02.2020

All other terms and conditions of the tender enquiry documents including Amendments issued so far shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/CPMP Website for the Corrigendum/amendments etc. if any, as these will be notified on these portals only. No separate advertisement will published in the newspaper in this regard.

**CGM (Proc)
HSCC (I) LTD**