

HSCC (I) Ltd

Amendment –IV Dated 07.01.2020

Tender Ref : HSCC/AIIMS-RAEBARELI/Hospital/2/2019 Dated : 28.11.2019

Department: Pediatric Surgery (K)

Item No. 1: Neonatal Open Care System

Specification Sr. No.	Tendered Specification	To be Read As
Point No 2	Warmer Module Swivel 90 degree on either Side.	It should have Reused technology which allows thermoregulation even during the time of X-ray
Point No 5	Heating element: Heater output should less than equal to 600 W, Medical grade with Parabolic reflector and protected by grid, warming system with, microprocessor based control, probes & Alarms	Heating element: Heater output should be less than equal to 600 W, & should be made with Calarod material, Medical grade with Parabolic reflector and protected by grid, warming system with, microprocessor based control, probes & Alarms
Point No 8 (a)	Should Allow tilt for Trendelenburg as well as reverse Trendelenburg position +/- 15 Deg.	Should Allow tilt for Trendelenburg as well as reverse Trendelenburg position +/- 12 Degrees and above
Point No 9	Should have Mattress and it should be sealed in such a way it should not allow ingress of liquid	Should have a pressure diffusing Mattress and it should be sealed in such a way it should not allow ingress of liquid.
Point No 18	Hood Suspended above the basinet integrates heating element and overhead light	It Should have Reused technology which allows the clinician to do the procedure without feeling the heat of temperature.
Additional	Point Added	Should be US-FDA or European CE with 4 Digit notified body number. or BIS

Further, with reference to the above tender, it is informed that specification for all other item no. K2,K3,K4 of Pediatrics Department remain unchanged. Their bid submission date is 15.01.2020.

Tender Ref : HSCC/AIIMS-RAEBARELI/Hospital/2/2019 Dated : 28.11.2019

Department: Dermatology (H)

Item No. 4 Full body UV Phototherapy Device

Specification Sr. No.	Tendered Specification	To be Read As
Name of Item	Full body UV Phototherapy Device	Hand foot UV Phototherapy Device
1	Unit should have 311nm NB UVB lamps, min 4 numbers 36W lamps	Unit should have NB UVB Tubes of 2 Feet Length. Minimum 4 Numbers NB UVB Tubes in 1 Module.
6	Mobile Trolley should have the provision for 4 units. Module, two for hands and two for foot	Mobile Trolley should have the provision for 4 Module, two for hands and two for foot
7	Unit should include UPS and back up batteries	Unit should include UPS and back up batteries
8	Ventilated fans should be included	Ventilated fans optional
9	Additional Point-1 requested in Specification	Integrated Dosimeter
	Point Added	Should be US-FDA or European CE with 4 Digit notified body number. or BIS

Further, with reference to the above tender, it is informed that specification for all other items no. H1, H2, H3 under Dermatology Department remain unchanged. Their bid submission date is 15.01.2020.

It is informed that amendment for all Equipment for Urology Department (C) has been issued on 17.12.2019. Their bid submission date is 15.01.2020.

Amendment for few Equipment (N1,N3,N4, N5,N6A,N6B,) have been issued on 26.12.2019. Their revised bid submission date is from 07.01.2020 to **15.01.2020**.

Amendment for all Equipment for Endocrinology Department (G) has been issued on 31.12.2019. Their bid submission date is 15.01.2020.

For balance items their amendments are to be issued. Meanwhile their bid submission date is extended from 07.01.2020 to 15.01.2020

All other terms and conditions of the tender enquiry documents shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/CPPP Website for the Corrigendum/amendments etc. if any, as these will be notified on these portals only. No separate advertisement will published in the newspaper in this regard.

CGM (Projects)
HSCC (I) Ltd.