

All Bidders

Amendment -V

Subject: Supply , Installation, Testing & Commissioning of CSSD Equipment for PGI Satellite Centre, Sangrur, Punjab.

IFB No. : HSCC/SES/CSSD/PGI/Sangrur/2019 dated 13.12.2019

This has reference to above IFB.

The following Amendment may be noted which shall be treated as part of the tender document and to be submitted duly signed & stamp along with tender.

Sr. No.	Bidders' Queries	Reply
1.	Horizontal Double Door Autoclave 850-1000 Ltrs. or more (12-14 STU) with accessories Page No. 3, Item No.2 The STU mentioned are to favour one company. Kindly amend the same to minimum 15 STU or more.	Horizontal Double Door Autoclave 850-1000 Ltrs. or more (12-15 STU or more) with accessories.
2.	Horizontal Double Door Autoclave 850-1000 Ltrs. or more (12-14 STU) with accessories <u>Amendment Required</u> Construction of material of Door and chamber should be clearly mentioned. Door should be manufactured of AISI 304 Stainless Steel. Chamber should be manufactured of 316 SS. Kindly mention the same.	The chamber and door should be made of solid , high quality 316L Stainless Steel.
3.	Horizontal Double Door Autoclave 850-1000 Ltrs. or more (12-14 STU) with accessories These access levels should be user selectable. The control system should have CPU processor with battery back-up & non volatile memories, Digital input/output controls, analog measuring inputs & COM	These access levels should be user selectable. The control system should have CPU processor with battery back-up & non volatile memories, Digital input/output controls, analog measuring inputs & COM ports for printer & PC connectivity/RS 232 port connectivity.

	<p>ports for printer & PC connectivity.</p> <p><u>Amendment Request</u></p> <p>Please allow RS232 port connectivity.</p>	
4.	<p>Double Door Washer disinfecter 300 Litre (10/15 DIN Trays) with accessories</p> <p>Page No. 6, Item No. 4, Point No. 4 Should supply 15 Nos. of standard DIN trays.</p> <p><u>Request</u></p> <p>In the heading you have mentioned, 10-15 DIN Trays, however in the specifications, 15 Nos. DIN Trays are mentioned.</p> <p>You are requested to correct the same and mention 10 to 15 DIN Trays.</p>	280-300L and (10 -15 DIN Trays)
5.	<p>Page No. 8, Item No. 5 Ultrasonic Cleaner (40L) Point No. 6a</p> <p>The units should be a compact bench model with a built in tank manufactured from high quality (316/304) stainless steel and a solid state generator that sends ultrasonic (approx. 40 KHz)</p> <p><u>Amendment required</u></p> <p>The frequency of 40 KHz may be company specific. Kindly amend the same to 25-40 KHz.</p>	The units should be a compact free standing bench model with a built in tank manufactured from high quality (316/304) stainless steel and a solid state generator that sends ultrasonic (approx. 25-40 KHz)
6.	<p>Ultrasonic Cleaner (40L) Point No. 6e</p> <p>It should have digital read out timer and temperature setting (temperature adjustable from 30 to 69 C or more) monitoring.</p> <p><u>Amendment required</u></p> <p>Kindly amend temperature adjustable setting to 25-75 deg C.</p>	It should have digital read out timer and temperature setting (temperature adjustable from 25-75 deg C or more) monitoring.
7.	<p>Horizontal Sterilizer 550 L -600 L or more with Accessories</p> <p><u>Amendment Required</u></p> <p>Construction of material of Door and chamber should be clearly mentioned.</p> <p>Door should be manufactured of AISI 304 Stainless Steel.</p> <p>Chamber should be manufactured of 316SS. Kindly mention the same.</p>	The chamber and door should be made of solid , high quality 316L Stainless Steel.

8.	<p>Horizontal Sterilizer 550 L -600 L or more with Accessories</p> <p>These access level should be user selectable. The control system should have CPU processor with battery back up & non volatile memories , Digital input/output controls, analog measuring inputs & COM ports for printer & PC connectivity.</p> <p><u>Amendment Request</u></p> <p>Please allow RS232 port connectivity.</p>	<p>These access level should be user selectable. The control system should have CPU processor with battery back up & non volatile memories , Digital input/output controls, analog measuring inputs & COM ports for printer & PC connectivity/RS 232 port connectivity.</p>
9.	<p>We also request you to kindly extend the bid submission end date of subject tender for another 10-15 days to enable us to prepare our bid.</p>	<p>Last date of submission is on 27.01.2020.</p>
10.	<p>The estimated budget of PGIMER, Sangrur is too less for the kind of equipment asked and the amount of turnkey work which is not defined exactly in the tender. Considering the quantity, capacity & the type of equipment. The minimum budget should be around 5 Cr. excluding CMC and also further provision of specified turnkey work budget may please be added accordingly.</p>	<p>Tender terms & conditions prevail.</p>
11.	<p>The foreign currency quote may please be allowed for at least major imported equipments i.e. Steam Sterilizer [Horizontal Autoclaves), Washer Disinfector & Low Temperature Sterilizer, Ultrasonic Cleaner Etc. The BOQ of the same may also be uploaded accordingly.</p>	<p>Tender terms & conditions prevail.</p>
12.	<p>The amended specification of AIIMS, Raebareli Tender may please be consider in toto. However for the different capacity you may amend the same as requested below:</p> <p>Double door washer disinfector 300 liter (10/15 din trays) with accessories 4) Chamber Capacity: Volume should be 300 L. Should supply 15 Nos of standard DIN trays</p> <p>Amendment Request:</p> <p>Double door washer disinfector 300 liter</p>	<p>280-300L and (10 -15 DIN Trays)</p>

	<p>(10/15 din trays) with accessories 4) Chamber Capacity: Volume should be 300 L. Should supply 10 or 15 standard DIN Trays</p> <p>The heading of washer disinfector is correct however the no, of tray please be asked in accordance to be model quoted i.e. 10 DIN or 15 DIN Trays.</p>	
13.	<p>For the low temperature it has been asked ETO Sterilizer of 250 L. The ETO (250L) may please be deleted and to cater approx Low Temperature Sterilization needs, the 2 Nos. of Low Temperature Plasma Sterilizer please be asked of (120L each usable capacity). The Specification of Low Temperature Plasma Sterilizer is attached herewith article was floated by PGIMER, Chandigarh itself for your kind consideration in this tender.</p>	Tender terms & conditions prevails.
14.	<p>Moreover we request you kindly keep the extension after the amendment is published to provide the sufficient time to quote the tender.</p>	Bid submission date is on 27.01.2020.
15.	<p>Please note that in the tender it has been asked to provide /submit exclusive manufacturer authorization for the all items. However we request that exclusive manufacturer authorization should be applicable to major items only like sterilizers (Autoclave) & Washer Disinfectors. As the other items can be third party items & exclusive authorization may not be possible.</p>	Tender terms & conditions prevail.

All other terms & conditions remain unchanged.

Sr. Chief General Manager -I, HSCC (I) Ltd.
For and behalf of Director, PGIMER, Chandigarh