

**All Bidders****Amendment –V**

Project: Supply, Installation, Testing & Commissioning of CSSD for Super Speciality Block at Indira Gandhi Medical College, Shimla, H.P.

IFB No. HSCC/SES/CSSD/SSB/Shimla/2019 dated 21.02.2019

This has reference to above IFB.

The following Amendment may be noted which shall be treated as part of the tender document and to be submitted duly signed & stamp along with tender.

Sr. No.	Bidders' Queries	Reply
1	<p><u>Amendment –II</u> Sr. No. 1</p> <p><u>Horizontal Steam Sterilizer Double door 550-600L or more</u> &amp; <u>High Speed Sterilizer- Double door 250L with accessories.</u></p> <p>Control system should have touch sensitive ,5-7 inches or more color display interface at operator loading side while it should have normal interface at unloading side.</p> <p>Request It should be brought to original tender specification as “ Control system should have touch sensitive , 7 inches or more colour display interface at operator loading side while it should have normal interface at unloading side.</p>	Amendment –II shall prevail.
2	<p><u>Amendment -II</u> Sr. No. 2</p> <p>Chamber Capacity: Operational volume should be able to process minimum 12 DIN trays (Approx 480x250x50) in single process. The chamber should be made of SS 316L quality with electro polished washed surfaces.</p> <p><u>Request</u> : We would request that minimum capacity as asked in original tender specification should be kept ie. 300 ltrs or more , it can be 12 DIN trays or 15 DIN Trays , but minimum capacity should be defined.</p> <p>It should be brought to original tender specs with the</p>	Capacity 300- 350 L 12 Din Trays or more Chamber – SS 316 L with electro polished washed surfaces.

	chamber should be made of SS 316 L with electro polished washed surfaces.	
3.	<p><b>Amendment No. 2</b>  <b>Sl. No. 3 and Sr. No. 44 page no. 8</b>  4 -6 Levels racks  Request  There is gross difference in 4 levels and 6 levels, one uniformity should be there we would request that it should be 5 to 6 level of racks with 300 ltrs or more operational volume.</p>	5-6 Level racks for 300-350 L capacity.
4.	<p><u>Amendment No. 2</u>  <b>Sr. No. 44 (d) Chamber Capacity :</b>  <b>Operational volume should be able to process minimum 12 DIN trays (Approx. 480 x 250 x 50) in single process. The chamber should be made of SS 316 L quality with electro polished washed surfaces.</b>  <b>Request</b>  We would request that minimum capacity as asked in original tender specification should be kept ie. 300 ltrs or more , it can be 12 DIN trays or 15 DIN Trays , but minimum capacity should be defined.  It should be brought to original tender specs with the chamber should be made of SS 316 L with electro polished washed surfaces.</p>	Capacity 300- 350 L 12 Din Trays or more Chamber – SS 316 L with electro polished washed surfaces.
5	<ul style="list-style-type: none"> <li><b>Certificates asked for various furniture items i.e. ISO 18001</b></li> </ul> <p>Objection:  ISO 18001 has been replaced with ISO 45001:2018 (documents attached), so we would request you to kindly amend the same where ever it has been asked in the technical specifications.</p>	Certificates asked for various furniture items i.e. ISO 18001/ISO 45001:2018

The bid submission date is extended from **04.04.2019 to 05.04.2019** and bid security should be valid for 180 days from the date of original bid submission ie. from 19.03.2019.

All other terms & conditions remain unchanged.

Sr. Chief General Manager -I, HSCC (I) Ltd.  
For and behalf of Principal, IGMC, Shimla