

**Amendment no V Dated 26.03.2018**  
**HSCC/PUR/PALI/02 dated 01.02.2018**

**Procurement of Medical Equipment for Rajasthan Pali**

All bidders are requested to note the following:

<b>Item No</b>	<b>NAME OF THE EQUIPMENT</b>	<b>Last date &amp; time sale/downloading of Tender document</b>	<b>Last closing/submission date &amp; time for receipt of tender</b>	<b>Date of opening of Techno – Commercial bids.</b>	<b>Earlier date of opening</b>	<b>Amendment Status for opening date</b>
1	Automatic Tissue Processor	03.04.2018, up to 13.00 hrs IST	03.04.2018, 14.00 hrs IST	03.04.2018, 14.30 hrs IST	26.03.2018	Date Extended with technical Specification updating
2	Automated ELISA Analyzer					
3	5 Part Haematology Analyzer					
5	Non Ref Centrifugal					
9	Paraffin Embedding System					
10	3 Channels Student Digital Physiograph					
11	Pure Tone Audiometer					

<b>Item No.1 Automatic Tissue Processor</b>	
<b>Tender Specifications Item No.1 Automatic Tissue Processor</b>	<b>Tenders Specification may Please read as</b>
The equipment should be carousel type with 12 stations of 4 liter each; 9 reagent stations and 3 wax baths.	Tender Terms prevail.
Metal tissue basket shall have less base diameter compare to upper diameter to avoid sticking of basket and capacity of 180 or more cassettes.	Tender Terms prevail.
5KVA online UPS support with minimum 6 hours power back up should be available.	Tender Terms prevail.
<b>The equipment should be USA-FDA and European-CE approved.</b>	Tender Terms prevail.
<b>Item No.3 5 Part Hematology Analyzer</b>	
The instrument should have random access discrete analysis modes for CBC, CBC DIFFERENTIAL IG.	Tender Terms prevail.
The instrument should have 24 PARAMETERS reported: WBC, RBC, HGB, HCT, MCV, MCH, MCHC, RDW-SD, RDW-CV, PLT, NEUT%, LYMPH%, MONO%, EOS%, BASO%, NEUT#, LYMPH#, MONO#, EOS#, BASO#, PDW, MPV, PCT, P-LCR, IG#, IG%	Tender Terms prevail.

The instrument should have throughput of at least 60 samples per hour in both the discrete analysis modes and auto loader mode.	Tender Terms prevail.
The sample aspiration volume for the complete differential blood count should not be more than ul.	Tender Terms prevail.
The instrument should be equipped with Fluorescence based semiconductor laser fluorescence flow cytometry for Differential channel.	Tender Terms prevail.
Instrument should be able to enumerate immature granulocytes.	Tender Terms prevail.
It should have high linearity of over 4 lacs for WBC's over 40 lacs for Platelets.	Tender Terms prevail.
<b>The system should be USFDA approved.</b>	Tender Terms prevail.
<b>Item No.9 PARAFFIN EMBEDDING SYSTEM.</b>	
1. It should have single module design.	Tender Terms prevail.
4. It should have Low and flat work-surface.	Tender Terms prevail.

All other terms and conditions of the tender enquiry document shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/ CPP as corrigendum /amendments etc. if any, will be notified on this portal only, no separate advertisement will published in the news papers.

s/d  
**Additional, Director**  
**Medical Education Department, Rajasthan**