

**HSCC (I) Ltd**

**Amendment –V Dated 08.01.2020**

**Tender Ref : HSCC/AIIMS-RAEBARELI/Hospital/2/2019 Dated : 28.11.2019**

**Department: Dermatology (H)**

**Item No. 4 Full body UV Phototherapy Device**

<b>Specification Sr. No.</b>	<b>Tendered Specification</b>	<b>To be Read As</b>
<b>Name of Item</b>	Full body UV Phototherapy Device	Hand foot UV Phototherapy Device
1	Unit should have 311nm NB UVB lamps, min 4 numbers 36W lamps.	Unit should have 311nm NB UVB lamps/tubes, min 4 numbers 36W lamps.
6	Mobile Trolley should have the provision for 4 units. Module, two for hands and two for foot	No Change
7	Unit Should include UPS and Back up batteries.	No Change
8	Ventilated fans should be included	Deleted
9	Point Added	Integrated Dosimeter
	Point Added	Should be US-FDA or European CE with 4 Digit notified body number. or BIS

**Department : Others (N)**

**Item No. N1 : Multipara Monitor**

<b>Sr. No.</b>	<b>Present</b>	<b>Amended As</b>
NI	5 no. for General Medicine + 1no. for Psychiary + 1no. for Obs. & Gyn. + 1no. for Dentistry + 2no. for Cardiology + 4no. for CTVS + 2no. for Neurosurgery + 4no. for Gastroenterology + 5no. for Surgical Oncology + 6no. for Urology + 5no. for Nephrology + 5no. for Emengency & Trauma  = 41 nos.	5 no. for General Medicine + 1no. for Psychiary + 1no. for Obs. & Gyn. + 1no. for Dentistry + 2no. for Cardiology + 4no. for CTVS + 2no. for Neurosurgery + 4no. for Gastroenterology + 5no. for Surgical Oncology + 6no. for Urology + 5no. for Nephrology + 5no. for Emengency & Trauma+16 no. for Pediatric & Neonates  = 57 nos.
EMD	Rs. 3,24,000.00	Rs. 4,52,000.00

The bid submission date is 15.01.2020 at 14:30hrs

All other terms and conditions of the tender enquiry documents shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/CPPP Website for the Corrigendum/amendments etc. if any, as these will be notified on these portals only. No separate advertisement will published in the newspaper in this regard.

**CGM (Projects)**  
**HSCC (I) Ltd.**