All Bidders

Amendment -VI

Date:01.03.2019

Project: Supply, Installation, Testing & Commissioning of Medical Gas Manifold System for PGI Satellite Centre, Sangrur, Punjab.

IFBNo: HSCC/SES/MGMS/PGI/SANGRUR/2019 Date: 09.01.2019

This has reference to above IFB.

The following Amendment may be noted which shall be treated as part of the tender document and to be submitted duly signed & stamp along with tender.

Sr. No.	Bidders' Queries	Reply		
1	Tech. SpecPage 20	Ward	vacuum	unit
		regulators	s should	be
	11.2 Ward Vacuum Units	digital.		
	The vacuum regulator will be step-less adjustable and			
	have large vacuum gauge providing digital/analogue			
	indication of the suction supplied by the regulator.			
	We request the ward vacuum unit regulators should be			
	digital instead of digital/analogue. The digital			
	regulators are more accurate and precise in			
	comparison to analogue system. Digital regulators are			
	being used now a days for accurate results.			
2	Tech. SpecPage 20	Theatre		Units
		regulators	s should	be
	11.3 Theatre Vacuum Units	digital.		
	The vacuum regulator will be step-less adjustable and			
	have large vacuum gauge providing Digital/Analogue			
	indication of the suction supplied by the regulator.			
	We request the theatre vacuum unit regulators should			
	be digital instead of digital/analogue. The digital			
	regulators are more accurate and precise in			
	comparison to analogue system. Digital regulators are			
	being used now a days for accurate results.			

All other terms & conditions remain unchanged.