HSCC/SES/MOT/Shimla/2019

All Bidders

Amendment –VI

Project: Supply, Installation, Testing & Commissioning of Modular Operation Theatre and Minor Operation Theatre for Super Speciality Block at Indira Gandhi Medical College, Shimla,H.P.

IFB No. HSCC/SES/MOT/SSB/Shimla/2019 dated 31.01.2019

This has reference to above IFB.

The following Amendment may be noted which shall be treated as part of the tenders document and to be submitted duly signed & stamp along with tender.

Sr. No.	Bidders' Queries	Reply					
1	Para 2, Page 1	Tender prevail.	terms	&	conditions		
	Wall & Ceiling System (SMS)						
	As, Wall & Ceiling System (SMS) is important part of Modular Operation Theatre; for authenticity and genuiness of the product, the certification should be European CE Certified with notified 4 digit body no. / UL Listed.						
	We requested earlier also but no amendment have been issued so far. We request you to kindly						
	incorporate.						
2	Para 1, Page 5	Tender prevail.		&	conditions		
	Touch Screen Control Panel	-					
	(CE or UL Listed)						
	Amendment IV dated 18.03.2019						
	Please refer to the Touch Screen Control Panel; as per Amendment-IV dated 18.03.02019 it is amended to CE or UL Listed. As per BOQ, Touch Screen Control Panel is Imported.						
	This has become contradictory item, as per certification it is indigenous whereas as per BOQ it is imported. The bidder will play with contradiction.						

	Since as per BOQ, Touch Screen Control Panel is Imported, therefore for quality product, item should be called Imported.	
	As per PGIMER Sangrur MOT Tender, the item is Imported.	
3	We request this should be strictly applied. Page 7 & 8	BOQ will prevail.
	10. Hermetically Sealed Doors 2.1 (H) x 1.510B. Double Leaf Door System (2.1 x 1.5)10C. Single Leaf Hinge Door System (2.1 x 1.0)	
	In BOQ, Hermetically Sealed Door 2 sizes are mentioned: 2100 x 1500 mm and 2100 x 1000 mm whereas in Tech. Specs. it is Hermetically Sealed Door 2100 x 1500; Double Leaf Door System 2.1 x 1.5 & Single Leaf Hinge Door 2.1 x 1.0.	
	BOQ is not matching with Technical Specs. Technical Specifications Viz-a-viz BOQ is contradicting. Please confirm.	
4	Ref. Touch Screen Control Panel – As per Amendment No.4, you have amended it to CE or UL Listed.	Tender terms & conditions prevail.
	Please clarify whether you need an indigenous product or imported one.	
5	Ref. Laminar Air Flow System – Please clarify whether this product is of indigenous make or imported.	Tender terms & conditions prevail.
6	Ref. Online Central UPS – We are in receipt of Amendment -4 for online UPS wherein it is mentioned that incase OT's are at different floor, then No. of UPS and standby for each floor will be provided along with UPS wiring –	Tender terms & conditions prevail.
	We therefore request you to kindly provide OT Complex Drawing with OT Room and UPS Room marking in the drawings to enable us to quote as per your requirement. Since online UPS is very costly equipment, hence indicative rates can't be quoted in tender based on assumptions.	
7	Please refer to Item Sr. No.20 of BOQ – Isolation Transformer – As per technical amendment issued, it has been amended to 20 KVA instead of 25 KVA but the same has not been amended in the BOQ.	Revised BOQ is attached.

	You	are	requested	to	kindly	amend	the	same
	acco	rdingl	у.					

All other terms & conditions remain unchanged.

Sr. Chief General Manager -I, HSCC (I) Ltd. For and behalf of Principal, IGMC, Shimla