

**Amendment No.VII Dated 09.02.2018**  
**HSCC/PUR/CNCI/Kolkata/Medical Equipment/03 dt. 15.11.2017**  
**Procurement of Medical Equipment CNCI 2nd Campus**

All bidders are requested to note the following:

<b>Item No.4 Digital Fluro Radiography</b>	
<b>Tender Specification for item No 4. Digital Fluro Radiography</b>	<b>May please read as</b>
Out of X-ray tube, Generator & Flat panel detector at least 2 must be manufactured by the quoting vendor themselves.	Tender terms prevail.
iii) System should have motor driven longitudinal and horizontal table top movements. Please specify the range of movements.	System should have motor driven longitudinal table top/detector (image chain) and horizontal table top movements. Please specify the range of movements.
19.iv. Dual head pressure injector US FDA approved with 2000 syringes	Dual head pressure injector with 1000 syringes.
<b>Item No 5. Digital Mobile X-ray</b>	
Out of X-ray tube, Generator & Flat panel detector at least 2 should be from the same manufacture.	Tender terms prevail.
vi) Exposure time range: 0.004 – 10 s	Exposure time range: Please specify.

All other terms and conditions of the tender enquiry document shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/ CPP as corrigendum /amendments etc. if any, will be notified on this portal only, no separate advertisement will published in the news papers.

**Chief General Manager, HSCC (I) Ltd**  
**For & on behalf of Director CNCI, Kolkata**