

All Bidders

Amendment -VII

Subject: Supply , Installation, Testing & Commissioning of CSSD Equipment for PGI Satellite Centre, Sangrur, Punjab.

IFB No. : HSCC/SES/CSSD/PGI/Sangrur/2019 dated 13.12.2019

This has reference to above IFB.

The following Amendment may be noted which shall be treated as part of the tender document and to be submitted duly signed & stamp along with tender.

Sr. No.	Bidders' Queries	Reply
1.	<p>Item No.2 Horizontal double door autoclave approx. 850 -1000L or more liters (12-14 STU or more) with accessories</p> <p>Amendment required: The STU mentioned are to favour one company. Kindly amend the same to minimum 15 STU or more.</p>	<p>Horizontal double door autoclave approx. 850 -1000L or more liters (15 STU or more) with accessories</p>
2.	<p>These access levels should be user selectable. The control system should have CPU processor with battery back-up &amp; non-volatile memories, Digital input/output controls, analogue measuring inputs &amp; COM ports for printer &amp; PC connectivity</p>	<p>RS232 port connectivity is allowed.</p>
3.	<p>Item No. 4 - DOUBLE DOOR WASHER DISINFECTOR 300 Litre (10/15 DIN Trays) WITH ACCESSORIES. Should supply 15 Nos. of standard DIN trays.</p> <p>In the heading you have mentioned 10-15 DIN Trays, however in the specifications, 15 Nos DIN Trays are mentioned. You are requested to correct the same and mention 10 to 15 DIN Trays</p>	<p>Double Door washer disinfectant 250-300 litre (10-15 Din Trays) with accessories. Should supply 15 nos 10-15 DIN trays.</p>
4.	<p>Item No. 5 – Ultrasonic Cleaner (40 L) Point No.6a</p> <p>The units should be a compact bench top model, with a built-in tank manufactured</p>	<p>The units should be a compact bench top model, with a built-in tank manufactured from high quality AISI(316/304) stainless steel and a solid-state generator</p>

	<p>from high quality AISI(316/304) stainless steel and a solid-state generator that sends Ultrasonic (Approx 40 KHz)</p> <p>The frequency of 40 KHz may be company specific. Kindly amend the same to 25-40 KHz</p>	that sends Ultrasonic (25- 40 KHz)
5.	<p>Item No. 6e</p> <p>It should have digital read out timer and temperature setting (temperature adjustable from 30 -69°C or more) monitoring.</p> <p>Kindly amend temperature adjustable setting to 25 – 75 deg C</p>	It should have digital read out timer and temperature setting (temperature adjustable from 25 - 75°C or more) monitoring.
6.	<p>Item No.-1</p> <p>Horizontal Sterilizer 550 Ltr-600L or more with Accessories</p> <p>Construction of material of Door and Chamber should be clearly mentioned. Door should be manufactured of AISI304 Stainless steel Chamber should be manufactured of 316SS Kindly mention the same</p>	The chamber and doors should be made of solid, high quality AISI 316L Stainless Steel
7.	<p>These access levels should be user selectable. The control system should have CPU processor with battery back-up &amp; non-volatile memories, Digital input/output controls, analogue measuring inputs &amp; COM ports for Printer &amp; PC connectivity.</p> <p>Please allow RS232 port connectivity</p>	RS232 port connectivity is allowed

**The bid submission date is extended from 03.02.2020 to 05.02.2020 and bid security should be valid for 180 days from the date of original bid submission ie. from 06.01.2020.**

All other terms & conditions remain unchanged.

Sr. Chief General Manager -I, HSCC (I) Ltd.  
For and behalf of Director, PGIMER, Chandigarh