

**HSCC (I) Ltd**

**Amendment –VII Dated 04.11.2020**

**Ref. IFB No: HSCC/SJH/Medical Equipment/2020/45 Dated: 21.08.2020**

**Technical Amendments and Revised Price Scheduled / BOQ:**

**Item No. 4.**

**DUAL HEAD GAMMA CAMERA**

| <b>T.S.No</b>  | <b>Existing As</b>  | <b>To be read as</b>   |
|----------------|---|--|
| 1.i            | The 16 Slice SPECT CT model quotes should be top-of-line model available with the manufacturer, meeting the technical specifications. The quoted model should not be older than 5 years. There must be at least 2 installations of same / similar model in India.   | The 16 Slice SPECT CT model quotes should be top-of-line model available with the manufacturer, meeting the technical specifications. The quoted model should not be older than 5 years. There must be at least 2 installations of same / similar model globally (satisfactory functioning and after sales service report must be submitted).  |
| 3. vii         | System planar sensitivity with LEHR/LEHRS collimator should be $\geq 190$ cpm/uCi   | -deleted-  |
| 11.<br>Xxxviii | New Point added   | Lead Glass 1000mm x 1200mm   |
|                | <b>Section – IX<br/>Qualification Criteria</b><br><br>2. (a) The Manufacturer should have supplied and installed in last Five years from the date of Tender Opening, at least 33% of the quoted quantity of the similar equipment which is functioning satisfactorily.<br><br>2 (b). The Tenderer quoting as authorized representative of the manufacturer meeting the above criteria 2 (a) should have executed at least one contract in the last five years from the date of tender opening of similar equipment which is functioning satisfactorily, anywhere in India of the same manufacturer. | For Dual Head Gamma Camera Only.<br><br>2. (a) The Manufacturer should have supplied and installed in last Five years from the date of Tender Opening, at least two installation /contracts of the similar equipment which is functioning satisfactorily anywhere in the world. (Satisfactory functioning and after sales service report must be submitted).<br><br>2 (b). The Tenderer quoting as authorized representative of the manufacturer meeting the above criteria 2 (a). |

**Technical Amendments:**

**Item No. 5 Positron Emission Tomography (PET/ CT )**

| <b>T.S.No</b> | <b>Technical Specifications</b>   | <b>to be read as</b>  |
|---------------|---|---|
| 3.2.11.d      | Bone Mineral Density  | -deleted-   |
| 3.5           | <p>5 independent post processing workstations/5 server based workstations with 5 permanent licenses<br/>Should have display matrix of 1024x1024, with colour monitor<br/>Software for Multiplanar and 3D image reconstruction in Gray scale and colour, CT angiography<br/>Having maximum and minimum intensity projection, volume rendering, multi-modality image fusion, and fly-through image software<br/>Facility for multimodality image fusion including PET, SPECT, CT, IVIRI<br/>Reporting monitor resolution: please specify<br/>All 5 no. client nodes or 5 workstation should be concurrently capable of all post processing function of PET&amp;CT including advanced applications (Neuro, oncology &amp; cardiac).<br/>They should be provided with keyboard, mouse and medical grade flat panel monitor of minimum 19”size</p> | <p>5 independent post processing workstations/5 server based workstations with 5 permanent licenses.<br/>Should have display matrix of 1024x1024, with colour monitor<br/>Software for Multiplanar and 3D image reconstruction in Gray scale and colour, CT angiography. Having maximum and minimum intensity projection, volume rendering, multi-modality image fusion, and fly-through image software. Facility for multimodality image fusion including PET, SPECT, CT, IVIRI.<br/>Reporting monitor resolution: please specify.<br/>All 5 no. client nodes or 5 workstation should be concurrently capable of all post processing function of PET &amp; CT except advanced applications (Neurology &amp; cardiac) mentioned at Tech. Specs. 3.2.11(j), 3.2.11(k), 3.2.11(o) – for these 2 concurrent licenses are required.<br/>They should be provided with keyboard, mouse and medical grade flat panel monitor of minimum 19”size.</p> |
| 6.6           | New point added   | Lead Glass 1000mm x 1200mm  |
| 10.2          | Five new sets of sources for PET QC and calibration shall be provided on demand. Responsibility of disposal of the old sources shall lie with the vendor company  | New sets of sources for PET QC and calibration shall be provided as per equipment requirement. Responsibility of disposal of the old sources shall lie with the vendor company  |

Prospective bidders are advised to participate as per Amendments and Revised Price Scheduled.

All other Specifications shall remain unchanged.

All other terms and conditions of the tender enquiry documents including Amendments issued so far shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website, CPPP Website & HSCC tender Wizard for the Corrigendum/amendments etc. if any, as these will be notified on these portals only. No separate advertisement will published in the newspaper in this regard.

For Medical Superintendent  
VMMC & Safdarjung Hospital